



Designation: D3701 – 17

Standard Test Method for Hydrogen Content of Aviation Turbine Fuels by Low Resolution Nuclear Magnetic Resonance Spectrometry¹

This standard is issued under the fixed designation D3701; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This test method covers the determination of the hydrogen content of aviation turbine fuels.

1.2 Use Test Methods [D4808](#) or [D7171](#) for the determination of hydrogen in other petroleum liquids.

1.3 The values stated in SI units are to be regarded as standard. The preferred units are mass percent hydrogen.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For a specific warning statement, see [6.1](#).

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D1322](#) Test Method for Smoke Point of Kerosine and Aviation Turbine Fuel

¹ This test method is under the jurisdiction of ASTM Committee [D02](#) on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee [D02.03](#) on Elemental Analysis.

Current edition approved June 1, 2017. Published June 2017. Originally approved in 1978. Last previous edition approved in 2012 as D3701 – 01(2012). DOI: 10.1520/D3701-17.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[D1740](#) Test Method for Luminometer Numbers of Aviation Turbine Fuels (Withdrawn 2006)³

[D4057](#) Practice for Manual Sampling of Petroleum and Petroleum Products

[D4808](#) Test Methods for Hydrogen Content of Light Distillates, Middle Distillates, Gas Oils, and Residua by Low-Resolution Nuclear Magnetic Resonance Spectroscopy

[D7171](#) Test Method for Hydrogen Content of Middle Distillate Petroleum Products by Low-Resolution Pulsed Nuclear Magnetic Resonance Spectroscopy

3. Summary of Test Method

3.1 A sample of the material is compared in a continuous wave, low-resolution, nuclear magnetic resonance spectrometer with a reference standard sample of a pure hydrocarbon. The results from the integrator on the instrument are used as a means of comparing the theoretical hydrogen content of the standard with that of the sample, the result being expressed as the hydrogen content (percent mass basis) in the sample.

4. Significance and Use

4.1 The combustion quality of aviation turbine fuel has traditionally been controlled in specifications by such tests as smoke point (see Test Method [D1322](#)), smoke volatility index, aromatic content of luminometer number (see Test Method [D1740](#)). Evidence is accumulating that a better control of the quality may be obtained by limiting the minimum hydrogen content of the fuel.

4.2 Existing methods allow the hydrogen content to be calculated from other parameters or determined by combustion

³ The last approved version of this historical standard is referenced on www.astm.org.

*A Summary of Changes section appears at the end of this standard

techniques. The method specified provides a quick, simple, and more precise alternative to these methods.

5. Apparatus

5.1 *Nuclear Magnetic Resonance Spectrometer*⁴—A low-resolution continuous-wave instrument capable of measuring a nuclear magnetic resonance of hydrogen atoms, and fitted with:

5.1.1 *Excitation and Detection Coil*, of suitable dimensions to contain the test cell.

5.1.2 *Electronic Unit*, to control and monitor the magnet and coil and containing:

5.1.2.1 *Circuits*, to control and adjust the radio frequency level and audio frequency gain.

5.1.2.2 *Integrating Counter*, with variable time period in seconds.

5.2 *Conditioning Block*—A block of aluminum alloy drilled with holes of sufficient size to accommodate the test cells with the mean height of the sample being at least 20 mm below the top of the conditioning block (see Fig. 1).

5.3 *Test Cells*—Nessler-type tubes of approximately 100 mL capacity with a nominal external diameter of 34 mm and a nominal internal diameter of 31 mm marked at a distance of approximately 51 mm above the bottom of the tube by a ring around the circumference. The variation between the internal diameters of the test cells used for the sample and reference material should not be greater than ± 0.5 mm.

NOTE 1—To avoid potential difficulties with tightness when the test cell is introduced into the magnet coil, users are cautioned to avoid test cells that have nominal external diameters that are greater than 34.2 mm.

5.4 *Polytetrafluoroethylene (PTFE) Plugs for Closing Test Cells*—Plugs as shown in Fig. 1 made from pure PTFE and a tight fit in the test cells.

5.5 *Insertion Rod*—A metal rod with a threaded end as shown in Fig. 1 for inserting and removing PTFE plugs from test cells.

5.6 *Analytical Balance*—Top pan type, capable of weighing the test cells in an upright position to an accuracy of ± 0.01 g.


6. Materials

6.1 *Reference Standard*—Dodecane, 99 % by mass minimum purity (**Warning**—Flammable).

⁴ This test method has been written around the Newport Analyzer Mark IIIIF (Oxford Analytical Instruments, Ltd., Oxford, England) and the details of the test method should be read in conjunction with the manufacturer's handbook.

This particular instrument was the only instrument available when the precision program was carried out. Any similar instrument would be accepted into the above method provided the new instrument was adequately correlated and proved to be statistically similar. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend.

The Newport Analyzer Mark IIIIF is no longer in production and was replaced by the manufacturer with the Newport 4000. The Newport 4000 model instrument has been demonstrated to provide equivalent results to those obtained with the Mark III models which were used to generate the precision data. As of the mid-1990's however, the Newport 4000 instrument is no longer being manufactured by the vendor. No newer models are currently being manufactured as replacements for the Newport 4000 instrument.



NOTE 1—All dimensions are nominal values in millimetres, unless the tolerance limit is specified.

FIG. 1 Hydrogen Content of Aviation Turbine Fuels

7. Sampling

7.1 Take a homogenous sample in accordance with Practice D4057.

8. Preparation of Apparatus

8.1 Read the following instructions in conjunction with the manufacturer's handbook. Preparation of the instrument is not critical but take care to prevent rapid temperature fluctuations of the instrument and the conditioning block, for example, avoid them from direct sunlight or from drafts.

8.2 The results obtained during the use of the equipment are susceptible to error arising from changes in the magnetic environment. Exercise care to ensure that there is a minimum of magnetic material in the immediate vicinity of the equipment and that this be kept constant throughout the course of a series of determinations.

8.3 Set the instrument controls to the following conditions:

NOTE 2—On new NMR instruments with variable gates the gate should be set at 1.5 gauss to comply with nonvariable gate instruments.

Radio frequency level	20 μ A
Audio frequency gain	500 on dial
Integration time	128 s

8.4 Switch on the main supply to the spectrometer and allow it to warm up for at least 1 h.

8.5 Place a test cell containing sample in the coil and adjust the tuning of the instrument until the two resonance curves on the oscilloscope are coincident. This setting may need to be readjusted during determinations.