



Designation: **F1538 – 03 (Reapproved 2009) F1538 – 03 (Reapproved 2017)**

Standard Specification for Glass and Glass Ceramic Biomaterials for Implantation¹

This standard is issued under the fixed designation F1538; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers the material requirements and characterization techniques for glass and glass-ceramic biomaterials intended for use as bulk porous or powdered surgical implants, or as coatings on surgical devices, but not including drug delivery systems.

1.2 The biological response to glass and glass-ceramic biomaterials in bone and soft tissue has been demonstrated in clinical use (**1-12**)² and laboratory studies (**13-17**).

1.3 This specification excludes synthetic hydroxylapatite, hydroxylapatite coatings, aluminum oxide ceramics, alpha- and beta-tricalcium phosphate, and whitlockite.

1.4 **Warning**—Mercury has been designated by EPA and many state agencies as a hazardous material that can cause central nervous system, kidney, and liver damage. Mercury, or its vapor, may be hazardous to health and corrosive to materials. Caution should be taken when handling mercury and mercury-containing products. See the applicable product Material Safety Data Sheet (MSDS) for details and EPA's website (<http://www.epa.gov/mercury/faq.htm>) for additional information. Users should be aware that selling mercury or mercury-containing products, or both, in your state may be prohibited by state law.

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:³

C158 Test Methods for Strength of Glass by Flexure (Determination of Modulus of Rupture)

C169 Test Methods for Chemical Analysis of Soda-Lime and Borosilicate Glass

C373 Test Methods for Determination of Water Absorption and Associated Properties by Vacuum Method for Pressed Ceramic Tiles and Glass Tiles and Boil Method for Extruded Ceramic Tiles and Non-tile Fired Ceramic Whiteware Products

C623 Test Method for Young's Modulus, Shear Modulus, and Poisson's Ratio for Glass and Glass-Ceramics by Resonance

C633 Test Method for Adhesion or Cohesion Strength of Thermal Spray Coatings

C693 Test Method for Density of Glass by Buoyancy

C729 Test Method for Density of Glass by the Sink-Float Comparator

C730 Test Method for Knoop Indentation Hardness of Glass

C958 Test Method for Particle Size Distribution of Alumina or Quartz by X-Ray Monitoring of Gravity Sedimentation

C1069 Test Method for Specific Surface Area of Alumina or Quartz by Nitrogen Adsorption

C1070 Test Method for Determining Particle Size Distribution of Alumina or Quartz by Laser Light Scattering

E228 Test Method for Linear Thermal Expansion of Solid Materials With a Push-Rod Dilatometer

F748 Practice for Selecting Generic Biological Test Methods for Materials and Devices

F981 Practice for Assessment of Compatibility of Biomaterials for Surgical Implants with Respect to Effect of Materials on Muscle and Insertion into Bone

¹ This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.13 on Ceramic Materials.

Current edition approved April 1, 2009; May 1, 2017. Published April 2009; June 2017. Originally approved in 1994. Last previous edition approved in 2003/2009 as F1538 – 03 (2009).^{e1} DOI: 10.1520/F1538-03R09-10.1520/F1538-03R17.

² The boldface numbers in parentheses refer to the list of references at the end of this specification.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

2.2 *Code of Federal Regulations*:⁴

Title 21, Part 820

2.3 *United States Pharmacopoeia*:⁵

Lead <252>

Mercury <261>

Arsenic <211>

Heavy Metals <231> Method I

2.4 *U.S. Geological Survey Method*:⁶

Cadmium

3. Terminology

3.1 *Definitions of Terms Specific to This Standard*:

3.1.1 *bioactive glass*—an amorphous silicate-based solid that is not intrinsically adhesive and that is capable of forming a cohesive bond with both hard and soft tissue when implanted, and will develop a hydroxycarbonate apatite layer when exposed to appropriate *in vitro* environments, such as simulated body fluid or tris-hydroxymethylaminomethane buffer.

3.1.2 *bioactive glass-ceramic*—an amorphous-derived crystalline silicate-based solid that is not intrinsically adhesive and that is capable of forming a cohesive bond with bone and soft tissue when implanted, and will develop a hydroxycarbonate apatite layer when exposed to appropriate *in vitro* environments, such as simulated body fluid or tris-hydroxymethylaminomethane buffer.

3.1.3 *bulk material*—intended to describe a unit material used as a load bearing implant.

3.1.4 *coating*—intended to describe a surface layer that is relatively thin compared to the overall dimensions of the prosthetic part that has been coated.

3.1.5 *glass biomaterial*—any one of a number of compositions of amorphous inorganic solids that are used as implant materials for various medical or dental uses, or both.

3.1.6 *glass-ceramic biomaterials*—any one of a number of compositions of an amorphous-derived crystalline solid that is used as an implantable biomaterial for medical or dental use, or both.

3.1.7 *particulate material*—intended to describe several pieces (usually small size) used together within an implant construct.

4. Chemical Requirements

4.1 Bulk compositions shall be tested using Test Method C169.

4.2 The concentration of trace element levels in the bioactive glass and glass-ceramics shall be limited as follows:

Element	ppm, max
Arsenic (As)	3
Cadmium (Cd)	5
Mercury (Hg)	5
Lead (Pb)	30
total heavy metals (as lead)	50

Either inductively-coupled plasma/mass spectroscopy (ICP/MS) (18), atomic absorption (AAS), or the methods listed in 2.3 and 2.4 shall be used.

5. Physical Characterization

5.1 The following physical and mechanical characterizations may be applicable to various bioactive glass and glass-ceramics products and should be used whenever possible to verify the material.

5.1.1 *Density*—The densities of glass and glass ceramic materials are related directly to the processing history and composition of the material. The density of the bulk material shall be measured using Test Methods C373 or C729 and shall be consistent for the specific materials.

NOTE 1—This test should use a non-aqueous liquid for bioactive glass and glass ceramic materials, which are known to react in an aqueous environment and could thereby affect the measurement.

5.1.2 *Flexural Strength*—When used as bulk materials in load bearing applications, the flexural strength of the bulk material shall be measured using Test Methods C158.

5.1.3 *Young's Modulus*—When used as a bulk material, Young's Modulus of glass and glass ceramic biomaterials shall be determined following Test Method C623.

⁴ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.

⁵ Available from U.S. Pharmacopeia (USP), 12601 Twinbrook Pkwy., Rockville, MD 20852-1790, <http://www.usp.org>.

⁶ Crock, J.G., Felichte, F.E., Briggs, P.H., "Determination of Elements in National Bureau of Standards Geological Reference Materials SRM 278 Obsidian and SRM 688 Basalt by Inductively Coupled Plasma-Atomic Emission Spectrometry," *Geostandards Newsletter* Vol 7, 1983pp. 335-340.



5.1.4 *Hardness*—Where applicable, for characterization of the material, the hardness of bulk samples shall be determined using Test Method C730. The Knoop indentation hardness is one of many properties that is used to characterize glasses. Attempts have been made to relate Knoop hardness to tensile strength, but no generally accepted methods are available. Such conversion is limited in scope and should be used with caution, except for special cases in which a reliable basis for conversion has been obtained by conversion tests.

5.1.5 *Surface Area*—The surface area of a particulate may be important in determining the reliability of the bioactivity of the material. Whenever the specific surface area of the material relates to function, the surface area of particulate glass and glass ceramic biomaterials shall be measured using Test Method C1069.

5.1.6 *Bond Strength of Glass or Glass Ceramic Coating*—When used as a coating on a metallic or ceramic substrate, the bond strength of the coating shall be measured following Test Method C633.

5.1.7 *Crystallinity*—For glass-ceramic biomaterials, the percent crystallinity and crystal phases present in glass ceramic biomaterials shall be determined by means of X-ray diffraction analysis. While there is no single standard method for determining the crystallinity and crystal phases of glass ceramic materials, techniques such as those detailed in Refs (19) and (20) should be followed to standardize methods as much as possible.

5.1.8 *Thermal Expansion*—Thermal expansion shall be measured using Test Method E228, when materials are to be used for coatings (raw materials are to be measured), or on finished product as a quality control test.

5.1.9 *Particle Size*—When used as a particulate, the particle size shall be measured in accordance with Test Methods C958 or C1070.

6. Biocompatibility

6.1 Glass and glass-ceramic biomaterials should be evaluated thoroughly for biocompatibility before human use. Bioactive glass and glass-ceramic materials are unique in their mode of action when implanted in the body due to the released ionic species and the mechanisms by which these materials bond with bony tissue. These materials have been found to exhibit an excellent tissue response in laboratory studies (13-17) and clinical usage (1-12). Before any new formulations are used clinically, the tissue response should be characterized by the methods recommended in Practice F748 and F981 as appropriate.

7. Test Specimen Fabrication

7.1 Test specimens should be prepared concurrent with implant devices, as well as from the same batch of material and by the same processes as those used in fabricating the glass and glass-ceramic implant device.

8. Quality Program Requirement

8.1 The manufacturer shall conform to Quality Systems requirements (2.2) or equivalent.

9. Keywords

9.1 bioactive glass; bioactive glass-ceramics; glass biomaterials; glass-ceramic biomaterial; surgical implants

APPENDIXES

(Nonmandatory Information)

X1. RATIONALE

X1.1 A number of glass-ceramic materials are available commercially. Bioactive glass and glass-ceramic materials are available commercially as synthetic graft materials for maintenance of the alveolar ridge; as devices for spinal fusion; as implants for replacement of the vertebral body, iliac crest, and ossicular chain of the middle ear; as bone filler to substitute for bone defects remaining after the excision of bone tumors and extraction of loosened joint prostheses; and as coatings on dental and orthopedic implants. As with any implant material, the bioresponse is critically dependent on the material properties. To achieve reliable biocompatibility, these properties must be known and consistent. This specification provides specifications for biocompatible grades of bioactive glass and glass-ceramics.

X1.2 In order to be called bioactive, the materials must demonstrate that living tissue is bonding to a significantly higher level than non-bonding implant control, as well as demonstrate that ionic species are released from the material into solution in a controlled and reproducible manner.

X1.3 Bioactive glass and glass-ceramic materials are generally silicate-based materials, with additions of oxides of calcium, phosphorous, and various alkalis. They may be phosphate-based materials as well. These materials may also include fluoride and