



SLOVENSKI STANDARD
SIST EN 794-3:2000
01-januar-2000

D`1 b]j Ybh] Urcf]!" "XY. DcgYVbY'nU H]j Y'nUj Ybh] Urcf Y'nUbi 'bc`a YX]Wbg_c
dca c `]b`dfYj cn

Lung ventilators - Part 3: Particular requirements for emergency and transport ventilators

Lungenbeatmungsgeräte - Teil 3: Besondere Anforderungen an Notfall- und Transportbeatmungsgeräte

Ventilateurs pulmonaires - Partie 3: Regles particulieres pour les ventilateurs d'urgence et de transport

iTeh STANDARD PREVIEW
(standards.itteh.ai)

[SIST EN 794-3:2000](https://standards.itteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-59002c073c74/sist-en-794-3-2000)

Ta slovenski standard je istoveten z: EN 794-3:1998

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
11.160	Účel [{ [First aid

SIST EN 794-3:2000

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 794-3:2000

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

ICS 11.040.10; 11.160

Descriptors: electromedical apparatus, artificial breathing apparatus, classifications, safety requirements, detail specifications, accident prevention, protection against electric shocks, earthing, protection against mechanical hazards, radiation protection, fire protection, electromagnetic compatibility, performance evaluation, equipment specifications

English version

Lung ventilators - Part 3: Particular requirements for emergency and transport ventilators

Ventilateurs pulmonaires - Partie 3: Règles particulières pour les ventilateurs d'urgence et de transport

Lungenbeatmungsgeräte - Teil 3: Besondere Anforderungen an Notfall- und Transportbeatmungsgeräte

This European Standard was approved by CEN on 1 July 1998.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 794-3:2000

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Contents	Page
Foreword	6
Introduction	7
Section one : General	8
1 Scope	8
2 Normative references	8
3 Terminology and definitions	11
4 General requirements and general requirements for tests	12
5 Classification	12
6 Identification, marking and documents	12
7 Power input	19
Section two : Environmental conditions	20
8 Basic safety categories	20
9 Removable protective means	20
10 Environmental conditions	20
11 Not used	21
12 Not used	21
Section three : Protection against electric shock hazards	22
13 General	22
14 Requirements related to classification	22
15 Limitation of voltage and/or energy	22
16 Enclosures and protective covers	22
17 Separation	22

ITEH STANDARD PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

Contents (continued)	Page
18 Protective earthing, functional earthing and potential equalization	22
19 Continuous leakage currents and patient auxiliary currents	22
20 Dielectric strength	22
Section four : Protection against mechanical hazards	23
21 Mechanical strength	23
22 Moving parts	23
23 Surfaces, corners and edges	24
24 Stability in normal use	24
25 Expelled parts	24
26 Vibration and noise	24
27 Pneumatic and hydraulic power	24
28 Suspended masses	24
Section five : Protection against hazards from unwanted or excessive radiation	25
29 X-radiation	25
30 Alpha, beta, gamma, neutron radiation and other particle radiation	25
31 Microwave radiation	25
32 Light radiation (including lasers)	25
33 Infra-red radiation	25
34 Ultra-violet radiation	25
35 Acoustical energy (including ultra-sonics)	25
36 Electromagnetic compatibility	25

ITeH STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 794-3:2000](https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-11e3-2000)

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-11e3-2000>

[/sist-en-794-3-2000](https://standards.iteh.ai/catalog/standards/sist-en-794-3-2000)

Contents (continued)	Page
Section six : Protection against hazards of ignition of flammable anaesthetic mixtures	27
37 Locations and basic requirements	27
38 Marking, accompanying documents	27
39 Common requirements for Category AP and Category APG equipment	27
40 Requirements and test for Category AP equipment, parts and components thereof	27
41 Requirements and test for Category APG equipment, parts and components thereof	27
Section seven : Protection against excessive temperatures and other safety hazards	28
42 Excessive temperatures	28
43 Fire prevention	28
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection	28
45 Pressure vessels and parts subject to pressure	29
46 Not used	29
47 Not used	29
48 Biocompatibility	29
49 Interruption of the power supply	29
Section eight : Accuracy of operating data and protection against hazardous output	30
50 Accuracy of operating data	30
51 Protection against hazardous output	30
Section nine : Abnormal operation and fault conditions; environmental tests	36
52 Abnormal operation and fault conditions	36

STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 794-3:2000](#)

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

Contents (continued)	Page
53 Environmental tests	36
Section ten : Constructional requirements	37
54 General	37
55 Enclosures and covers	37
56 Components and general assembly	37
57 Mains parts, components and layout	41
58 Protective earthing - Terminals and connections	41
59 Construction and layout	41
Annexes	
Annexes A to K	42
Annex AA (informative) Rationale	42
Annex BB (normative) Legibility and visibility	48
Annex CC (informative) Bibliography	49
Annex DD (normative) Special national conditions	50
Annex ZA (informative) Clauses of this European Standard addressing essential requirements or other provisions of EU Directives	51

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 794-3:2000](https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000)

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 1999, and conflicting national standards shall be withdrawn at the latest by January 1999.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this standard.

See annex DD for Special National Conditions.

This European Standard applies to lung ventilators and has been prepared in three parts. This Part addresses lung ventilators for emergency and transport use. Parts 1 and 2 address lung ventilators for critical care and lung ventilators for home care respectively.

Annex BB and DD are normative and form part of this Part of this European Standard.

Annexes AA, CC and ZA are for information only.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 794-3:2000

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

Introduction

This European Standard is one of a series based on European Standard EN 60601-1:1990.

In EN 60601-1: 1990 this type of European Standard is referred to as a "Particular Standard". As stated in 1.3 of EN 60601-1: 1990, the requirements of this European Standard take precedence over those of EN 60601-1: 1990.

Clauses and subclauses additional to those in EN 60601-1: 1990 are numbered beginning '101'. Additional annexes are lettered beginning 'AA'. Additional items in lettered lists are lettered beginning 'aa)'. Additional tables and figures are numbered beginning '101'.

Annex AA contains rationale statements for this Part of this European Standard. The clauses and subclauses which have corresponding rationale statements are marked with **R** after their number.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 794-3:2000

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

Section one. General

1 Scope

The scope given in clause 1 of EN 60601-1: 1990 applies with the following addition :

1.101 R) This part of this European Standard specifies requirements for ventilators, driven by a power source and intended for emergency and transport use.

This covers a range of devices, from relatively simple ventilators intended, primarily, for use with a face mask and for limited periods (e.g. gas powered ventilators) through to devices for pre-planned longer term use.

This part does not cover operator-powered ventilators (i.e. manual resuscitators).

Ventilators aboard aircraft are likely to be subject to additional requirements and national/international regulations.

Additional parts, e.g. concerning lung ventilators for critical care (see EN 794-1), home care ventilators (see EN 794-2), operator powered resuscitators and recent developments such as jet and very high frequency ventilation and oscillation are published or under consideration.

2 Normative references

This European Standard incorporates, by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

- | | |
|----------|--|
| EN 475 | <i>Medical devices - Electrically generated alarm signals</i> |
| EN 550 | <i>Sterilization of medical devices - Validation and routine control of ethylene oxide sterilization</i> |
| EN 552 | <i>Sterilization of medical devices - Validation and routine control of sterilization by irradiation</i>
<small>SIST EN 794-3:2000</small> |
| EN 554 | <i>Sterilization of medical devices - Validation and routine control of sterilization by moist heat</i>
<small>https://standards.gob.uk/standards/sist/794-3-2000/975-003/39bb2e673c74/sist-en-794-3-2000</small> |
| EN 556 | <i>Sterilization of medical devices - Requirements for medical devices to be labelled "STERILE"</i> |
| EN 737-1 | <i>Medical gas pipeline systems - Part 1: Terminal units for compressed medical gases and vacuum</i> |

prEN 737-3: 1994	<i>Medical gas pipeline systems - Part 3: Pipelines for compressed medical gases and vacuum</i>
prEN 737-6: 1996	<i>Medical gas pipeline systems - Part 6: Dimensions of probes for terminal units for compressed medical gases and vacuum</i>
EN 738-1	<i>Pressure regulators for use with medical gases - Part 1: Pressure regulators and pressure regulators with flow- metering devices</i>
EN 739	<i>Low-pressure hose assemblies for use with medical gases</i>
EN 980	<i>Graphical symbols for use in the labelling of medical devices</i>
EN 1281-1	<i>Anaesthetic and respiratory equipment - Conical connectors - Part 1: Cones and sockets</i>
EN 1281-2	<i>Anaesthetic and respiratory equipment - Conical connectors - Part 2: Screw-threaded weight-bearing connectors (ISO 5356- 2:1987 modified)</i>
EN 1820	<i>Anaesthetic reservoir bags</i>
EN ISO 4135: 1996	<i>Anaesthesiology - Vocabulary (ISO 4135:1995)</i>
EN ISO 8185	<i>Humidifiers for medical use - General requirements for humidification systems (ISO 8185:1997)</i>
EN 12342	<i>Breathing tubes intended for use with anaesthetic apparatus and ventilators</i>
prEN 12598: 1996	<i>Oxygen monitors for patient breathing mixtures - Particular requirements</i>
EN 60601-1: 1998	<i>Medical electrical equipment - Part 1: General requirements for safety</i>
EN 60601-1-2	<i>SIST EN 794-3:2000 Medical electrical equipment - Part 1: General requirements for safety - Collateral Standard: Electromagnetic compatibility - Requirements and tests</i>
IEC 60068-2-6:	<i>Environmental testing - Tests methods - Test F_c - Vibration (sinusoidal)</i>
IEC 60068-2-29:	<i>Environmental testing procedures - Test - Test E_b and guidance - Bump</i>

IEC 60068-2-32: 1975	<i>Basic environmental testing procedures - Tests methods - Part 2: Tests - Test E_d: Free fall</i>
IEC 60068-2-36	<i>Basic environmental testing procedures - Test methods - Part 2: Tests - Test F_{ab}: Random vibration wide band - Reproducibility medium</i>
IEC 60079-4	<i>Electrical apparatus for explosive gas atmospheres - Part 4: Method of test for ignition temperature</i>
IEC 61000-4-2	<i>Electrostatic discharge immunity test - Basic EMC publication</i>
ISO 32: 1977	<i>Gas cylinders for medical use - Marking for identification of content</i>
ISO 9360: 1992	<i>Anaesthetic and respiratory equipment - Heat and moisture exchangers for use in humidifying respired gases in humans</i>

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 794-3:2000](#)

<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

3 Terminology and definitions

Clause 2 of EN 60601-1: 1990 applies with the following additions, and the definitions given in EN ISO 4135: 1996 apply:

2.1.5 applied part R): Add the following item:

All parts of the ventilator intended to be connected to the patient or to the breathing system.

3.1 clearly legible: Visual attribute of information displayed by the equipment that allows the operator to discern (or identify) qualitative or quantitative values or functions under a specific set of environmental conditions.

3.2 cycling pressure: Pressure in the ventilator breathing system which initiates an inspiratory or expiratory phase.

3.3 emergency and transport ventilator: Portable active medical device for lung ventilation intended for emergency use and/or transportation.

NOTE: Hereinafter called 'ventilator'

3.4 label: Printed or graphic information applied to a medical device or any of its containers or wrappers.

3.5 marking: Inscription in writing or as a symbol applied on a medical device from which the inscription is not dissociable.

3.6 maximum limited pressure ($P_{lim\ max}$): highest pressure, measured at the patient connection port, which can be attained in the ventilator breathing system with a single fault condition of the ventilator.

3.7 operator powered resuscitator: Resuscitation device in which ventilation of the lungs is produced by the operator compressing the compressible unit of the device.

3.8 operator's position: Intended orientation of the operator with respect to the equipment for normal use according to the instructions for use.

3.9 permanent connection: Connection which can be separated only by the use of a tool.

[SIST EN 794-3:2000](https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39b2e672c74/sist-en-794-3-2000)

[https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-](https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39b2e672c74/sist-en-794-3-2000)

3.10 ventilation (\dot{V}): Volume of gas per minute entering or leaving the patient's lungs.

3.11 ventilator breathing system (VBS): Breathing system bounded by the low pressure gas input port(s), the gas intake port(s) and the patient connection port, together with the fresh-gas intake and exhaust port(s), if these are provided .

4 General requirements and general requirements for tests

4.1 Modifications to clause 3 of EN 60601-1:1990

Clause 3 of EN 60601-1:1990 applies with the following additions:

In 3.6 add the following:

aa) Applicable single fault conditions are:

- short and open-circuits of components or wiring which can :
 - cause sparks to occur, or;
 - increase the energy of sparks, or;
 - increase temperature (see section seven).
- incorrect output resulting from software error.

NOTE: See also 54.1.

bb)R) An oxidant leak which is not detected by e.g. an alarm or periodic inspection shall be considered a normal condition and not a single fault condition.

4.2 Clause 4 of EN 60601-1:1990

Clause 4 of EN 60601-1:1990 applies.

5 Classification

Clause 5 of EN 60601-1: 1990 applies.

NOTE: A ventilator can have applied parts of different types.

6 Identification, marking and documents

ITeH STANDARD PREVIEW
(standards.iteh.ai)

Clause 6 of EN 60601-1: 1990 applies with the following additions and modifications:

[SIST EN 794-3:2000](#)

In 6.1 add the following to item e):
<https://standards.iteh.ai/catalog/standards/sist/7b8d32eb-3cda-495b-a003-39bb2e673c74/sist-en-794-3-2000>

If imported from outside the EU, the name and address of the person responsible or of the authorized representative of the manufacturer or the importer established within the EU shall be provided with the label or the accompanying documents.

In 6.1 add the following to item j):

The marking(s) for the rated input requirements of the ventilator required in 6.1j) of

EN 60601-1: 1990 shall be given in amperes.

In 6.1 add the following items :

aa) All operator-interchangeable flow-direction sensitive components shall be permanently marked with a clearly legible arrow indicating the direction of flow.

bb) Any high pressure gas input port shall be marked on or in the vicinity with the name or symbol of the gas as given in EN 739, with the range of supply pressures in kPa and with the maximum flow requirement in l/min (see 6.8.3a), 2nd dash, 6th bullet).

6.1cc) If operator-accessible ports are provided, they shall be marked. The following terms may be used:

- Driving gas input port: "DRIVING GAS INPUT"
- Inflating gas input port: "INFLATING GAS INPUT"
- Fresh gas intake port: "FRESH GAS INTAKE"
- Fresh gas input port: "FRESH GAS"
- Emergency air intake port: "WARNING: EMERGENCY AIR INTAKE - DO NOT OBSTRUCT"
- Manual ventilation port: "BAG"
- Gas output port: "GAS OUTPUT"
- Gas return port: "GAS RETURN"
- Gas exhaust port: "EXHAUST"

Alternatively, other terms, pictograms or symbols may be used, in which case they shall be explained and referred to in the above terms.

dd) Labelling and packaging of the ventilator and accessories (e.g. breathing system attachments).

The labelling and marking of the packages of the devices shall contain the following:

- If the intended purpose of the device is not obvious to the operator, the attachment or its package shall be provided with an instruction leaflet or operating instructions.
- The name or trade name and address of the manufacturer. For attachments imported into the community, 6.1.e) of this European standard applies.
- Device identification and content information.