

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment –  
Part 2-40: Particular requirements for the basic safety and essential performance  
of electromyographs and evoked response equipment**

**Appareils électromédicaux –  
Partie 2-40: Exigences particulières pour la sécurité de base et les performances  
essentielle des électromyographes et des appareils à potentiel évoqué**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

##### [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Medical electrical equipment –  
Part 2-40: Particular requirements for the basic safety and essential performance  
of electromyographs and evoked response equipment**

**Appareils électromédicaux –  
Partie 2-40: Exigences particulières pour la sécurité de base et les performances  
essentiels des électromyographes et des appareils à potentiel évoqué**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.20; 11.040.55; 11.040.99

ISBN 978-2-8327-0033-4

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards .....	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	9
201.4 General requirements .....	10
201.5 General requirements for testing ME EQUIPMENT .....	11
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	11
201.7 ME EQUIPMENT identification, marking and documents.....	11
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	14
201.10 Protection against unwanted and excessive radiation HAZARDS .....	14
201.11 Protection against excessive temperatures and other HAZARDS .....	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	15
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT .....	17
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	17
201.15 Construction of ME EQUIPMENT.....	18
201.16 ME SYSTEMS.....	18
201.17 ELECTROMAGNETIC compatibility of ME EQUIPMENT and ME SYSTEMS .....	18
202 ELECTROMAGNETIC DISTURBANCES – Requirements and tests.....	18
Annexes .....	22
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	22
Annex AA (informative) Particular guidance and rationale.....	23
Bibliography.....	29
Index of defined terms used in this particular standard.....	30
Figure AA.1 – Suggested cable layout for EMISSION and radiated IMMUNITY testing .....	26
Figure AA.2 – Example of test setup for protection against the effects of HF SURGICAL ME EQUIPMENT .....	27
Figure AA.3 – Example of test setup for protection against the effects of HF SURGICAL ME EQUIPMENT .....	28
Table 202.101 – Pass/fail criteria for Table 4 of IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020 .....	19
Table 202.102 – Pass/fail criteria for Table 7 of IEC 60601-1-2:2014.....	20
Table 202.103 – Pass/fail criteria for Table 8 of IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020 .....	21
Table 202.104 – Pass/fail criteria for Table 5, Table 6, Table 9 of IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020.....	21
Table 201.C.101 – Marking on the outside of ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT or its parts.....	22

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-40: Particular requirements for the basic safety and essential performance of electromyographs and evoked response equipment**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-40 has been prepared by subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) added requirements for constant voltage stimulators;
- b) clarified requirements for VISUAL STIMULATORS.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62D/2168/FDIS	62D/2191/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title, or at the beginning of a paragraph or table title, indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](https://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

# Sample Document

get full document from [standards.iteh.ai](https://standards.iteh.ai)

## INTRODUCTION

This document concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT. It amends and supplements IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* (IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020).

The aim of this revision is to bring this document up to date with reference to IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020.

The requirements of this document take priority over those of IEC 60601-1.

A "General guidance and rationale" for the more important requirements of this document is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the document but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this document.

# Sample Document

get full document from [standards.iteh.ai](https://standards.iteh.ai)

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-40: Particular requirements for the basic safety and essential performance of electromyographs and evoked response equipment

#### 201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

##### 201.1.1 Scope

###### *Replacement:*

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT, hereafter referred to as ME EQUIPMENT.

NOTE 1 Myofeedback equipment, where the capturing of muscle contraction is based on electromyography, is within the scope of this document.

NOTE 2 ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT is intended for diagnostic and monitoring applications.

NOTE 3 If the ME EQUIPMENT supports both ELECTROMYOGRAPHY and EVOKED RESPONSE STIMULATION, clauses for electrical, auditory, and visual stimulators are applicable. In case the equipment supports ELECTROMYOGRAPHY, but not EVOKED RESPONSE STIMULATION, clauses concerning solely requirements for stimulators are NOT within the scope of this document.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

The following ME EQUIPMENT are excluded:

- ME EQUIPMENT intended for therapeutic application;
- ME EQUIPMENT intended for transcutaneous electrical nerve stimulators and electrical muscle stimulators (ME EQUIPMENT covered by IEC 60601-2-10).

##### 201.1.2 Object

###### *Replacement:*

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT [as defined in 201.3.201 and 201.3.202.]

##### 201.1.3 Collateral standards

###### *Addition:*

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020 apply as modified in Clause 202. IEC 60601-1-3, IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

#### 201.1.4 Particular standards

##### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

The numbering of clauses and subclauses of this document corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by the text of this document.

"*Addition*" means that the text of this document is additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

"*Amendment*" means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However, due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, for example 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, any applicable collateral standards and this document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

## 201.2 Normative references

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

NOTE Informative references are listed in the bibliography.

*Addition:*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-2:2014/AMD1:2020

IEC 60318 (all parts), *Electroacoustics – Simulators of human head and ear*

ISO 15004-2, *Ophthalmic instruments – Fundamental requirements and test methods – Part 2: Light hazard protection*

## 201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 apply, except as follows:

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

NOTE An index of defined terms is found beginning on page 30.

*Additional terms and definitions:*

### 201.3.201

#### ELECTROMYOGRAPH

ME EQUIPMENT for the detection or recording of biopotentials accompanying nerve and muscle action, either spontaneously, intentionally or evoked by electrical or other stimulation

### 201.3.202

#### EVOKED RESPONSE EQUIPMENT

ME EQUIPMENT for the detection or recording of biopotentials resulting from an evoking stimulus

Note 1 to entry: The stimulus can be electrical, tactile, auditory, visual, olfactory, etc.

**201.3.203****ELECTRICAL STIMULATOR**

part of ME EQUIPMENT for the application of electric currents via ELECTRODES in direct contact with the PATIENT, for the evoking of biopotentials

**201.3.204****PULSE DURATION**

duration of the electrical stimulus pulse WAVEFORM at 50 % of the peak amplitude

**201.3.205****WAVEFORM**

variations in magnitude of an electrical stimulus output (either voltage or current) as a function of time appearing in the APPLIED PART(S) of the ELECTRICAL STIMULATOR or the collected biopotentials by the BIOPOTENTIALS INPUT PART

**201.3.206****AUDITORY STIMULATOR**

part of ME EQUIPMENT for the application of sound pressure from a transducer (headphone, bone conductor or free-field) to the ear(s) of the PATIENT, for the evoking of biopotentials

**201.3.207****VISUAL STIMULATOR**

part of ME EQUIPMENT for the application of pulsed electromagnetic radiation in the visible spectrum from a transducer to the eyes of the PATIENT, for the evoking of biopotentials

**201.3.208****BIOPOTENTIAL INPUT PART**

APPLIED PART(S) of ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT for the collection of biopotentials

**201.3.209****ELECTRODE**

conductive portion that is applied to the PATIENT to detect electrical activity or to apply the stimulus from the ELECTRICAL STIMULATOR to the PATIENT

**201.3.210****PATIENT LEAD**

cable connected between an ELECTRODE and either a PATIENT CABLE or the ME EQUIPMENT

**201.3.211****PATIENT CABLE**

multiwire cable used to connect PATIENT LEADS to ME EQUIPMENT

**201.4 General requirements**

Clause 4 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

**201.4.2 \*RISK MANAGEMENT PROCESS for ME EQUIPMENT or ME SYSTEMS**

*Addition:*

MANUFACTURERS shall include, within their RISK MANAGEMENT FILE, the RISK associated with the potential use of their STIMULATORS and accessories to deliver current exceeding 10 mA RMS or current densities for any ELECTRODE exceeding 2 mA/cm<sup>2</sup>.

### 201.4.3 ESSENTIAL PERFORMANCE

*Addition:*

Because of the variety of clinical applications for ELECTROMYOGRAPHS and EVOKED RESPONSE, no additional ESSENTIAL PERFORMANCE is specified in this document. However, ESSENTIAL PERFORMANCE is determined by the manufacturer in accordance with the requirements of 4.3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

#### 201.4.11 Power input

*Replacement:*

The power input is measured with a load resistance of the lowest value specified in the technical description (see 201.7.9.3.101 a)), and with any output controls set to result in maximum power input.

### 201.5 General requirements for testing ME EQUIPMENT

Clause 5 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows.

#### 201.5.4 Other conditions

*Addition:*

Where values of voltage and current are used in this document, they mean the RMS values of an alternating, direct or composite voltage or current averaged over 1 s unless stated otherwise.

### 201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

#### 201.6.2 \* Protection against electric shock

*Amendment:*

Delete TYPE B APPLIED PART.

### 201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

#### 201.7.2 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts

##### 201.7.2.3 \* Consult ACCOMPANYING DOCUMENTS

*Replacement:*

Safety sign ISO 7010-M002 shall be used (see Table D.2, safety sign 10 in Annex D of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020).

### 201.7.2.7 Electrical input power from the SUPPLY MAINS

*Replacement:*

The RATED power input of MAINS operated ME EQUIPMENT shall be the maximum power input averaged over any period of 5 s under the specified operating conditions set out by the manufacturer.

### 201.7.2.8 Output connectors

#### 201.7.2.8.2 Other power sources

*Addition:*

See also 201.12.4.102.

#### 201.7.2.13 \* Physiological effects (SAFETY SIGNS and warning statements)

*Addition:*

ME EQUIPMENT capable of delivering electrical stimulus outputs into a load resistance of 1 000  $\Omega$  in excess of 10 mA RMS or 10 V RMS averaged over any period of 5 s shall be marked near the ELECTRODE connections with the safety sign ISO 7010-M002 (see safety sign 10 in Table D.2 of Annex D of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020).

### 201.7.4 Marking of controls and instruments

#### 201.7.4.2 Control devices

*Replacement:* get full document from [standards.iteh.ai](https://standards.iteh.ai)

The following information shall be provided in the instructions for use:

- a) electrical stimulator output range: specify minimum and maximum values;
- b) type of control used: specify if the output control is continuous or in discrete increments;
- c) if discrete increments are used, the minimum increment shall be specified;
- d) load impedance range: specify minimum and maximum load.

*Compliance is checked by inspection of ACCOMPANYING DOCUMENTS.*

### 201.7.9 ACCOMPANYING DOCUMENTS

#### 201.7.9.2 Instructions for use

*Additional subclause:*

##### 201.7.9.2.101 Additional information in instructions for use

The instructions for use shall contain additionally:

- a) \* Information on the output WAVEFORM(s), including any DC component, PULSE DURATIONS, pulse repetition frequencies, maximum amplitude of output voltage or current, and the effect of load impedance on the demanded parameters.
- b) \* Advice on the size of ELECTRODES to be used and the method of application for each particular type of examination for which the ELECTRICAL STIMULATOR is intended.
- c) Advice on any necessary precautions to be taken when the output contains a DC component larger than 10  $\mu$ A when averaged over 1 s.

- d) \* Advice that a PATIENT with an implanted electronic device (for example a cardiac pacemaker) should not be subjected to electrical stimulation unless specialist medical opinion has first been obtained.
- e) Advice to avoid trans-thoracic stimulation.
- f) A warning on the following potential HAZARDS:
  - Connection of a PATIENT to a high frequency (HF) surgical equipment and to an ELECTROMYOGRAPH or EVOKED RESPONSE EQUIPMENT simultaneously can result in burns at the site of the ELECTRODES and possible damage to the APPLIED PARTS;
  - Operation in close proximity to a shortwave or microwave therapy equipment can produce instability in the APPLIED PARTS.
- g) \* For ME EQUIPMENT capable of delivering output values in excess of 10 mA RMS or 10 V RMS into the specified load impedance (see 201.7.9.3.101a)), averaged over 1 s, or having an energy greater than 10 mJ per pulse into the specified load impedance:
  - a list of recommended ELECTRODES that can be used with the ME EQUIPMENT.
- h) \* Advice to avoid accidental contact between connected but unapplied APPLIED PARTS and other conductive parts including those connected to protective earth.
- i) \* Any known susceptibilities to electromagnetic phenomena.

### 201.7.9.3 Technical description

*Additional subclause:*

#### 201.7.9.3.101 Additional information in the technical description

The technical description shall additionally contain the following:

- a) The technical description shall specify the parameters mentioned in 201.7.9.2.101 along with the range of load impedances for which these parameters are valid.

### 201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

#### 201.8.3 \* Classification of APPLIED PARTS

*Replacement:*

The APPLIED PARTS of ELECTRICAL STIMULATORS, VISUAL STIMULATORS, AUDITORY STIMULATORS and BIOPOTENTIAL INPUT PARTS shall be TYPE BF APPLIED PARTS or TYPE CF APPLIED PARTS.

#### 201.8.5.2.3 PATIENT LEADS or PATIENT CABLES

*Addition:*

PATIENT LEADS of ELECTROMYOGRAPHS are usually kept short (30 cm or less) and tied together; therefore, any PATIENT LEAD which falls off will stay in the vicinity of the PATIENT and thus there are no additional requirements for the ELECTRODE.

Where the PATIENT LEADS are long (longer than 30 cm) or not tied together, compliance is verified by inspection of the RISK MANAGEMENT FILE.