

INTERNATIONAL STANDARD

CONSOLIDATED VERSION

**Radiation protection instrumentation - Measurement of personal dose
equivalents for X, gamma, neutron and beta radiations - Active personal
dosemeters**

get full document from standards.iteh.ai



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2026 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search -

webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	6
INTRODUCTION	8
1 Scope	9
2 Normative references	9
3 Terms and definitions	12
4 Units and symbols	19
4.1 Units	19
4.2 Symbols	19
5 Mechanical characteristics	21
5.1 Case	21
5.2 Switches	22
6 General characteristics	22
6.1 Storage of dose information	22
6.2 Indication	22
6.3 Dosimeter markings	22
6.4 Retention of radioactive contamination	22
6.5 Ranges for dose and dose rate	22
6.6 Rated range of an influence quantity	22
6.7 Effect of radiation not intended to be measured and use of more than one dosimeter	23
6.8 Intrinsic and environmental background of the instrument	23
6.9 Dose or dose rate alarms	23
6.9.1 General	23
6.9.2 Dose alarms	23
6.9.3 Dose rate alarms	24
6.9.4 Alarm output	24
6.10 Indication of malfunction	24
6.11 Assignment of the dose value to the dosimeter	24
6.12 Algorithm to evaluate the indicated value	24
7 General test procedures	25
7.1 Nature of tests	25
7.2 Reference conditions and standard test conditions	25
7.3 Tests for influence quantities of type F	26
7.4 Tests for influence quantities of type S	26
7.5 Phantom for testing	26
7.6 Position of detector assembly for the purpose of testing	26
7.7 Position of dosimeter during use	27
7.8 Minimum rated range of influence quantity	27
7.9 Low dose rates	27
7.10 Statistical fluctuations	27
7.11 Production of reference radiation	27
8 Additivity of indicated value	27
8.1 Requirements	27
8.2 Method of test	28
8.2.1 General	28
8.2.2 Preparation of the test	28

8.2.3	Practical test	29
8.3	Interpretation of the results	29
9	Radiation performance requirements and tests	29
9.1	General.....	29
9.2	Consideration of the uncertainty of the conventional quantity value.....	30
9.3	Constancy of the dose response, dose rate dependence and statistical fluctuations	30
9.3.1	General	30
9.3.2	Requirements	30
9.3.3	Method of test using sources	35
9.3.4	Interpretation of the results of the test using sources.....	36
9.3.5	Method of test for photon dosimeters using natural radiation	37
9.3.6	Interpretation of the results of the test using natural radiation.....	37
9.4	Variation of the response due to photon radiation energy and angle of incidence	38
9.4.1	Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$	38
9.4.2	Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$	38
9.4.3	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	39
9.5	Variation of the response due to neutron radiation energy and angle of incidence	40
9.5.1	General	40
9.5.2	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	40
9.6	Variation of the response due to beta radiation energy and angle of incidence	41
9.6.1	General	41
9.6.2	Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$	41
9.6.3	Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$	42
9.6.4	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	43
9.7	Over indication due to radiation incident from the side of an $H_p(10)$, $H_p(3)$ or $H_p(0,07)$ dosimeter	43
9.7.1	Requirements	43
9.7.2	Method of test	43
9.7.3	Interpretation of the results.....	44
9.8	Retention of dose reading	44
9.8.1	General	44
9.8.2	Requirements	44
9.8.3	Method of test and interpretation of the results	44
9.9	Overload characteristics	44
9.9.1	General	44
9.9.2	Requirements	45
9.9.3	Method of test and interpretation of the results	45
9.10	Alarm.....	46
9.10.1	General	46
9.10.2	Response time for dose rate indication and alarm.....	46
9.10.3	Accuracy of dose alarm	47
9.10.4	Accuracy of dose rate alarm	47
9.11	Model function	48

10	Electrical and environmental performance requirements and tests	48
10.1	General.....	48
10.2	Power supplies	48
10.2.1	General requirements	48
10.2.2	Specific primary batteries requirements	49
10.2.3	Specific secondary batteries requirements.....	49
10.2.4	Method of test and interpretation of the results (primary and secondary batteries).....	49
10.3	Ambient temperature	50
10.3.1	Requirements	50
10.3.2	Method of test and interpretation of the results	51
10.4	Relative humidity	52
10.4.1	Requirements	52
10.4.2	Method of test and interpretation of the results	52
10.5	Atmospheric pressure	52
10.6	Sealing	52
10.7	Storage.....	52
10.8	Light exposure	52
10.8.1	General	52
10.8.2	Requirements	53
10.8.3	Interpretation of the results.....	53
10.9	Dose build-up, fading and self-irradiation (hybrid dosimeters).....	53
11	Electromagnetic performance requirements and tests.....	54
11.1	General.....	54
11.2	Electrostatic discharge.....	55
11.2.1	Requirements	55
11.2.2	Method of test and interpretation of the results	55
11.3	Radiated electromagnetic fields	56
11.3.1	Requirements	56
11.3.2	Method of test and interpretation of the results	56
11.4	Conducted disturbances induced by fast transients or bursts	56
11.4.1	Requirements	56
11.4.2	Method of test and interpretation of the results	56
11.5	Conducted disturbances induced by surges	57
11.5.1	Requirements	57
11.5.2	Method of test and interpretation of the results	57
11.6	Conducted disturbances induced by radio-frequencies.....	57
11.6.1	Requirements	57
11.6.2	Method of test and interpretation of the results	57
11.7	50 Hz/60 Hz magnetic field	57
11.7.1	Requirements	57
11.7.2	Method of test and interpretation of the results	57
11.8	Voltage dips and short interruptions.....	57
11.8.1	Requirements	57
11.8.2	Method of test and interpretation of the results	58
12	Mechanical performance, requirements and tests	58
12.1	General.....	58
12.2	Drop test.....	58

12.2.1	Requirements	58
12.2.2	Method of test and interpretation of the results	58
12.3	Vibration test	59
12.3.1	Requirements	59
12.3.2	Method of test and interpretation of the results	59
12.4	Microphonics test.....	59
12.4.1	Requirements	59
12.4.2	Method of test and interpretation of the results	59
13	Software.....	60
13.1	General.....	60
13.2	Design and structure of the software.....	61
13.2.1	Requirements	61
13.2.2	Method of test	61
13.3	Identification of the software	61
13.3.1	Requirements	61
13.3.2	Method of test	61
13.4	Authenticity of the software and the presentation of results.....	62
13.4.1	Requirements	62
13.4.2	Method of test	62
13.5	Alarm and stop of system operation under abnormal operating conditions.....	62
13.5.1	Requirements	62
13.5.2	Method of test	62
13.6	Control of instrument parameters.....	63
13.6.1	Requirements	63
13.6.2	Method of test	63
13.7	Storage of measurement data	63
13.7.1	Requirements	63
13.7.2	Method of test	63
13.8	Transmission of data.....	64
13.8.1	Requirements	64
13.8.2	Method of test	64
13.9	Hardware interfaces and software interfaces.....	64
13.9.1	Requirements	64
13.9.2	Method of test	65
13.10	Documentation for the software test.....	65
13.10.1	Requirements	65
13.10.2	Method of test	66
14	Uncertainty.....	66
15	Documentation	66
15.1	Type test report	66
15.2	Certificate	66
16	Operation and maintenance manual	67
Annex A (normative)	Statistical fluctuations	68
Annex B (normative)	Coverage intervals.....	68
Annex C (informative)	Causal connection between readout signals, indicated value and measured value	73
Annex D (informative)	Procedure to determine the variation of the relative response due to radiation energy and angle of radiation incidence.....	74

Annex E (informative) Computational method of test for mixed irradiations	76
Bibliography.....	78
Figure B.1 – Test for coverage interval	70
Figure C.1 – Data evaluation in dosimetry systems	73
Figure E.1 – Flow chart of a computer program to perform tests according to 8.2	77
Table 1 – Symbols and abbreviated terms	19
Table 2 – Reference conditions and standard test conditions	25
Table 3 – Radiation characteristics of $H_p(0,07)$ dosimeters for X, gamma and beta radiation	31
Table 4 – Radiation characteristics of $H_p(3)$ dosimeters for X, gamma and beta radiation	32
Table 5 – Radiation characteristics of $H_p(10)$ dosimeters for X and gamma radiation	33
Table 6 – Radiation characteristics of $H_p(10)$ dosimeters for neutron radiation	34
Table 7 – Values of c_1 and c_2 for w different dose values and n indications for each dose value	37
Table 8 – Electrical and environmental characteristics of dosimeters	53
Table 9 – Electromagnetic disturbance characteristics of dosimeters	55
Table 10 – Mechanical disturbances characteristics of dosimeters.....	58
Table A.1 – Number of instrument readings required to detect true differences (95 % level of confidence) between two sets of instrument readings on the same instrument.....	69
Table B.1 – Student's t -value for a double sided 95 % coverage interval.....	71
Table E.1 – Example of dosimeter response table and range limits.....	76

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**Radiation protection instrumentation -
Measurement of personal dose equivalents for
X, gamma, neutron and beta radiations -
Active personal dosimeters**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 61526 edition 4.1 contains the fourth edition (2024-03) [documents 45B/1047/FDIS and 45B/1049/RVD] and its amendment 1 (2026-06) [documents 45B/1113/FDIS and 45B/1121/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC 61526 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2010. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) Modification of the title;
- b) Inclusion of the measurement quantity for the dose in the lens of the eye, $H_p(3)$;
- c) Inclusion of measurement quantity for dose in the skin and extremities, $H_p(0,07)$;
- d) Inclusion of dosimeters between active and passive: "hybrid dosimeters";
- e) Inclusion of software requirements;
- f) Harmonization of requirements for linearity to IEC 62387;
- g) Revised neutron energy response requirements.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1047/FDIS	45B/1049/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

INTRODUCTION

This document applies to active, (direct reading or hybrid) personal dosimeters and monitors used for measuring personal dose equivalents $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, for X, gamma, neutron, and beta radiations.

For personal dose equivalent $H_p(10)$ and for X and gamma radiation, two minimum rated ranges for the photon energy are given. The first from 20 keV to 150 keV is for workplaces where low energy X-rays are used, e.g., in diagnostic medicine, the second from 80 keV to 1,25 MeV is for workplaces where high energy X-rays and/or gamma sources are used, e.g., in industry. For neutron radiation the minimum rated range of neutron energy is from 0,025 eV (thermal neutrons) to 10 MeV. The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

For personal dose equivalent $H_p(3)$ and for X and gamma radiation, a minimum rated range for photon energy from 30 keV to 250 keV is given. For personal dose equivalent $H_p(0,07)$ a range of 30 keV to 1250 keV or, for workplaces where low energy X-rays are used, 20 keV to 150 keV, is given. For beta radiation for both quantities, the minimal rated range is from 0,24 MeV to 0,8 MeV (mean beta particle energy). The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

In some applications, for example, at a nuclear reactor installation where 6 MeV photon radiation is present, measurement of personal dose equivalent (rate) $H_p(10)$ for photon energies up to 10 MeV should be required. In some other applications, measurement of $H_p(10)$ down to 10 keV should be required.

For personal dosimeters, requirements for measuring the dose quantities $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, and for monitoring of the respective dose rate quantities are given. The measurement of these dose rate quantities is an option for personal dosimeters.

Establishments in some countries may be permitted to use this type of personal dosimeter as the dosimeter to provide the dose of record by an approved dosimetry service.

1 Scope

This document applies to personal dosimeters with the following characteristics:

- a) They are worn on the trunk, close to the eye, or on the extremities.
- b) They measure the personal dose equivalents $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, from external X and gamma, neutron (not for $H_p(3)$), and beta radiations, and may measure the respective personal dose equivalent rates for the same radiations (for alarming purposes).
- c) They have a digital indication. This indication may or may not be attached.
- d) They have alarm functions for the personal dose equivalents or personal dose equivalent rates except for hybrid dosimeters. For hybrid dosimeters an alarm function for the personal dose equivalents shall be implemented in the associated readout system.

NOTE 1 When reference is made in this document to "dose", this is meant to indicate personal dose equivalent, unless otherwise stated.

NOTE 2 When reference is made in this document to "dosimeter", this is meant to include all personal dosimeters, unless otherwise stated.

NOTE 3 This document does not cover neutron dosimeters for the $H_p(3)$ measurements up to now.

This document specifies requirements for the dosimeter and, if supplied, for its associated readout system.

Usually, a dosimeter is not able to measure all quantities given above. Thus, the dosimeter is only tested with regard to those quantities and types of radiation it is intended to be used for. Other types of radiation are considered as influence quantities which also may have requirements.

This document specifies, for the dosimeters described above, general characteristics, general test procedures, radiation characteristics as well as electrical, mechanical, safety and environmental characteristics. The only requirements specified for associated readout systems are those which affect its accuracy of readout of the personal dose equivalent and alarm settings, and those which concern the influence of the reader on the dosimeter.

This document does not cover special requirements for accident or emergency dosimetry, although the dosimeters may be used for this purpose.

This document does not apply to dosimeters used for measurement of pulsed radiation, such as radiation emanating from many medical diagnostic X-ray facilities, linear accelerators or similar equipment.

NOTE 4 Requirements and testing procedure for dosimeters used in pulsed field of ionizing radiation can be found in IEC TS 63050 or IEC TS 62743.

This document does not apply for dosimeters to measure ambient or directional dose equivalent.

NOTE 5 Requirements for ambient or directional dose equivalent meters can be found e.g. in: IEC 60846-1; IEC 62387, IEC 61017 or IEC 60532.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395:2014, *International Electrotechnical Vocabulary (IEV) – Part 395: Nuclear instrumentation – Physical phenomena, basic concepts, instruments, systems, equipment and detectors*

IEC 60050-395:2014/AMD1:2016

IEC 60050-395:2014/AMD2:2020

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60086-1:2021, *Primary batteries – Part 1: General*

IEC 60086-2:2021, *Primary batteries – Part 2: Physical and electrical specifications*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*

IEC 60529:1989/AMD1:1999

IEC 60529:1989/AMD2:2013

IEC 60904-3, *Photovoltaic devices – Part 3: Measurement principles for terrestrial photovoltaic (PV) solar devices with reference spectral irradiance data*

IEC 61000-4-2, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

IEC 61000-4-5, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-6, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8, *Electromagnetic compatibility (EMC) – Part 4-8: Testing and measurement techniques – Power frequency magnetic field immunity test*

IEC 61000-4-11, *Electromagnetic compatibility (EMC) – Part 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase*

IEC 61000-6-2, *Electromagnetic compatibility (EMC) – Part 6-2: Generic standards – Immunity for industrial environments*

IEC 61187:1993, *Electrical and electronic measuring equipment – Documentation*

IEC 62387:2020, *Radiation protection instrumentation – Dosimetry systems with integrating passive detectors for individual, workplace and environmental monitoring of photon and beta radiation*

IEC TR 62461:2015, *Radiation protection instrumentation – Determination of uncertainty in measurement*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 98-3:2008/Suppl.1:2008, *Propagation of distributions using a Monte Carlo method and Corr.1 (2009)*

ISO 4037-1:2019, *Radiological protection – X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy – Part 1: Radiation characteristics and production methods*

ISO 4037-2:2019, *Radiological protection – X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy – Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3:2019, *Radiological protection – X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy – Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO 4037-4:2019, *Radiological protection – X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy – Part 4: Calibration of area and personal dosimeters in low energy X reference radiation fields*

ISO 6980-1:2023, *Nuclear energy – Reference beta-particle radiation – Part 1: Methods of production*

ISO 6980-2:2023, *Nuclear energy – Reference beta-particle radiation – Part 2: Calibration fundamentals related to basic quantities characterizing the radiation field*

ISO 6980-3:2023, *Nuclear energy – Reference beta-particle radiation – Part 3: Calibration of area and personal dosimeters and the determination of their response as a function of beta radiation energy and angle of incidence*

ISO 8529-1:2021, *Reference neutron radiations – Part 1: Characteristics and methods of production*

ISO 8529-2:2000, *Reference neutron radiations – Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 8529-3:2023, *Neutron reference radiation fields – Part 3: Calibration of area and personal dosimeters and determination of their response as a function of neutron energy and angle of incidence*

ISO 12789-1, *Reference radiation fields – Simulated workplace neutron fields – Part 1: Characteristics and methods of production*

ISO 12789-2, *Reference radiation fields – Simulated workplace neutron fields – Part 2: Calibration fundamentals related to the basic quantities*

ISO 21909-1:2021, *Passive neutron dosimetry systems – Part 1: Performance and test requirements for personal dosimetry*

ISO 80000-10:2019, *Quantities and units – Part 10: Atomic and nuclear physics*

ICRU Report 51:1993, *Quantities and units in radiation protection dosimetry*

ANSI N42.17A, *American National Standard For Performance Specifications For Health Physics Instrumentation – Portable Instrumentation For Use In Normal Environmental Conditions*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60050-395, IEC 60529 and ICRU Report 51, as well as the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.1

acceptance test

contractual test to prove to the customer that the device meets certain conditions of its specification

[SOURCE: IEC 60050-151:2001, 151-16-23]

3.2

active personal dosimeter

dosimeter measuring personal dose equivalent using electronic components (e.g. either a direct reading dosimeter or a hybrid personal dosimeter including the associated readout system)

3.3

calibration

quantitative determination of the reference calibration coefficient, N_0 , and the correction for non-constant response, r_n , under a controlled set of standard test conditions for which all the m relative response values, r_q , are unity and all the l deviations, D_p , are zero

Note 1 to entry: Methods of calibration are given in ISO 4037 for photons, ISO 6980 for beta and ISO 8529 for neutron radiations. Determinations concerning the build-up and phantom as given in these standards should be taken into account during the calibration of the dosimeter, for the use of phantoms see 7.5.

3.4

calibration coefficient

N

quotient of the conventional quantity value of a quantity, H_r , and the indicated value, G_r , at the point of test for a specified reference radiation under specified reference conditions. It is expressed as

$$N = \frac{H_r}{G_r}$$

Note 1 to entry: The calibration coefficient N is dimensionless when the instrument indicates the quantity to be measured, then N is equal to the calibration factor. A dosimeter indicating the conventional quantity value correctly has the calibration coefficient of one.

[SOURCE: ISO 29661:2012, 3.1.5, modified: replaced notes by new one.]

3.5

coefficient of variation

ratio of the standard deviation s to the arithmetic mean \bar{x} of a set of n measurements x_i given by the following formula:

$$v = \frac{s}{\bar{x}} = \frac{1}{\bar{x}} \sqrt{\frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2}$$

3.6 combined standard measurement uncertainty combined standard uncertainty

u_c

standard measurement uncertainty that is obtained using the individual standard measurement uncertainties associated with the input quantities in a measurement model

Note 1 to entry: In case of correlations of input quantities in a measurement model, covariances must also be taken into account when calculating the combined standard measurement uncertainty; see also ISO/IEC Guide 98-3:2008, 2.3.4.

[SOURCE: ISO/IEC Guide 98-3:2008, 2.31]

3.7 conventional quantity value conventional value of a quantity conventional value

quantity value attributed by agreement to a quantity for a given purpose

Note 1 to entry: The term "conventional true quantity value" is sometimes used for this concept, but its use is discouraged.

Note 2 to entry: Sometimes a conventional quantity value is an estimate of a true quantity value.

Note 3 to entry: A conventional quantity value is generally accepted as being associated with a suitably small measurement uncertainty, which might be zero.

[SOURCE: ISO/IEC Guide 99:2007, 2.12, modified: skipped examples.]

3.8 correction for non-constant response

r_n

quotient of the response, R , under specified conditions where only the quantity to be measured is varied, and the reference response, R_0 . It is expressed as:

$$r_n = \frac{R}{R_0}$$

Note 1 to entry: For an instrument with constant response, r_n is equal to one.

3.9 detector assembly

assembly of a radiation detector and the associated components needed for converting signal from the accumulated dose or prevailing dose rate integrated over time to dose measurand

Note 1 to entry: The detector assembly contains both mechanical parts, e.g. energy filtering and electrical parts e.g. which convert the active detector signal to detectable pulse indication (of active detector) or passive detector signal to monotonically changing voltage or current signal.

EXAMPLE A personal dosimeter is to be calibrated using a phantom. The combination of personal dosimeter and phantom and possibly further reading instruments and cables comprise one detector assembly.

3.10 deviation

D

difference between the indicated values for the same value of the measurand of a dose(rate) meter, when an influence quantity assumes, successively, two different values