

INTERNATIONAL ELECTROTECHNICAL COMMISSION

IEC 62127-1
Edition 2.0 2022-03

Ultrasonics - Hydrophones -
Part 1: Measurement and characterization of medical ultrasonic fields

INTERPRETATION SHEET 1

This interpretation sheet has been prepared by IEC technical committee 87: Ultrasonics.

The text of this interpretation sheet is based on the following documents:

DISH	Report on voting
87/941/DISH	87/946/RVDISH

Full information on the voting for the approval of this interpretation sheet can be found in the report on voting indicated in the above table.

get full document from standards.iteh.ai

Subclause 5.1.7.1, last paragraph

This is clarified by the following:

Medical ultrasound devices often require additional acoustic field measurements during version upgrades to incorporate newly developed probes, modes, or transmission technologies into the system. In such cases, even within the same device, 5.1.7.1 permits data from a second edition conforming measurement system and data from earlier measurement systems to coexist depending on the testing period.