



IEC 62127-3

Edition 2.1 2026-02

# INTERNATIONAL STANDARD

CONSOLIDATED VERSION

Ultrasonics - Hydrophones -  
Part 3: Properties of hydrophones for ultrasonic fields

Sample Document

get full document from [standards.iteh.ai](https://standards.iteh.ai)



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2026 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD .....	3
INTRODUCTION .....	5
INTRODUCTION to Amendment 1 .....	5
1 Scope .....	6
2 Normative references .....	6
3 Terms, definitions and symbols .....	6
4 List of symbols .....	9
5 Hydrophone characteristics .....	10
5.1 General .....	10
5.2 Basic information .....	10
5.3 Sensitivity .....	11
5.4 Frequency response .....	11
5.4.1 Stated frequency band .....	11
5.4.2 Frequency dependence .....	12
5.5 Directional response .....	12
5.5.1 General .....	12
5.5.2 Determination of the directional response .....	12
5.5.3 Symmetry of directional response .....	13
5.6 Effective hydrophone size .....	13
5.6.1 General .....	13
5.6.2 Model of directional response .....	14
5.6.3 Fitting of experimental response to theoretical predictions .....	15
5.7 Dynamic range, linearity and electromagnetic interference .....	16
5.8 Electric output characteristics .....	16
5.8.1 Hydrophone without pre-amplifier .....	16
5.8.2 Hydrophone assembly .....	17
5.8.3 Output lead configuration .....	17
5.9 Environmental aspects .....	17
5.9.1 Temperature range .....	17
5.9.2 Water tightness .....	18
5.9.3 Water properties and incompatible materials .....	18
5.9.4 Exposed material .....	18
5.10 Guidance manual .....	18
5.11 List of hydrophone characteristics .....	18
Annex A (informative) Examples of information on hydrophone properties .....	19
A.1 General .....	19
A.2 Basic information .....	19
A.3 Sensitivity and frequency response .....	19
A.4 Directional response .....	23
A.5 Effective dimension .....	24
A.6 Dynamic range, linearity and electromagnetic interference .....	25
A.6.1 Lower dynamic limit .....	25
A.6.2 Upper dynamic limit .....	25
A.7 Electric output characteristics .....	26
A.8 Environmental aspects .....	26
Annex B (informative) Rationale .....	27

B.1	General.....	27
B.2	Changes to the determination of directional response .....	27
B.3	Changes to the determination of effective radius.....	27
Annex C (informative)	Membrane hydrophone directivity model .....	29
C.1	General.....	29
C.2	Details of model.....	29
Bibliography	.....	30
Figure A.1	– Frequency response of 0,2 mm needle hydrophone in the range 1 MHz to 40 MHz .....	21
Figure A.2	– Frequency response of 0,2 mm needle hydrophone in the range 100 kHz to 1 MHz .....	22
Figure A.3	– Directional response of 0,2 mm needle hydrophone .....	24
Figure A.4	– Effective radius of 0,2 mm needle hydrophone.....	24
Figure A.5	– Comparison of modelled and experimentally derived directional response .....	25
Table A.1	– Example of basic information for 0,2 mm needle hydrophone assembly .....	19
Table A.2	– Frequency response of 0,2 mm needle hydrophone in the range 1 MHz to 40 MHz .....	20
Table A.3	– Frequency response of 0,2 mm needle hydrophone in the range 100 kHz to 1 MHz .....	21

Sample Document

get full document from [standards.iteh.ai](https://standards.iteh.ai)

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**Ultrasonics - Hydrophones -  
Part 3: Properties of hydrophones for ultrasonic fields**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 62127-3 edition 2.1 contains the second edition (2022-12) [documents 87/818/FDIS and 87/824/RVD] and its amendment 1 (2026-02) [documents 87/897/CDV and 87/914/RVC].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

IEC 62127-3 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures to determine the effective hydrophone size have been changed according to the rationale outlined in Annex B.
- d) Requirements on the frequencies for which the effective hydrophone size shall be provided have been changed to achieve practicality for increased frequency bands.
- e) The new Annex B and Annex C have been added.
- f) Annex A has been updated to reflect the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/818/FDIS	87/824/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

## INTRODUCTION

The spatial and temporal distribution of acoustic pressure in an ultrasonic field in a liquid medium is commonly determined using miniature ultrasonic **hydrophones**. The properties of these **hydrophones** have been dealt with in a number of IEC standards in various aspects. The purpose of this document is to bring together all these specifications and to establish a common standard on the properties of ultrasonic **hydrophones**. The main **hydrophone** application in this context is the measurement of ultrasonic fields emitted by medical diagnostic equipment in water. Other medical applications are field measurements for therapy equipment such as that used in lithotripsy, high-intensity focused ultrasound (HIFU) and physiotherapy. **Hydrophones** are also used extensively in non-medical applications for both product development and quality control including:

- mapping of the ultrasound field within ultrasonic cleaning baths;
- characterization of acoustic fields used in transmission measurement systems (e.g. ultrasonic spectrometers, ultrasonic attenuation meters and velocimeters);
- characterization of acoustic fields used in reflection measurement systems (e.g. Doppler flowmeters).

While the term **hydrophone** can be used in a wider sense, it is understood here as referring to miniature piezoelectric **hydrophones**. It is this instrument type that is used today in various areas of ultrasonics and, in particular, to quantitatively characterize the field structure of medical diagnostic instruments. With regard to other pressure sensor types, such as those based on fibre optics, some of the requirements of this document are applicable to these as well but others are not. If in the future these other **hydrophone** types gain more importance in field measurement practice, their properties will have to be dealt with in a revised version of this document or in a separate one.

Underwater **hydrophones** as covered by IEC 60500, IEC 60565-1, and IEC 60565-2 are not included in this document, although there is an overlap in the frequency ranges. Underwater **hydrophones** are used in natural waters, even in the ocean, and this leads to different technical concepts and requirements. In addition, the main direction of acoustic incidence in underwater applications is at various angles and often at right angles to the **hydrophone axis**, whereas in this document it is assumed that the main direction of acoustic incidence is in the direction of the **hydrophone axis**.

Historically, ultrasonic **hydrophones** were used almost exclusively as amplitude sensors. However, the complex-valued nature of a **hydrophone's** system response function is well understood and IEC 62127-1:2022 makes use of this within the deconvolution procedures it contains. In this document, requirements are specified for the amplitude aspect of the **hydrophone** sensitivity and recommendations are provided for the phase aspect which can be derived either via calibration, or via calculation methods that are discussed in IEC 62127-1:2022.

### INTRODUCTION to Amendment 1

The second edition of IEC 62127-3 was published in 2022. The application of the document has revealed improvement potential with respect to the measurement requirements for **directional responses** and regarding data evaluation requirements for **effective hydrophone size** determination. This amendment addresses these items to ensure the practicality of the required measurements as well as the generation of a common data base to support **hydrophone** spatial averaging correction methods and the simplifying assumptions therein.

## 1 Scope

This part of IEC 62127 specifies relevant **hydrophone** characteristics.

This document is applicable to:

- **hydrophones** employing piezoelectric sensor elements, designed to measure the pulsed and continuous wave ultrasonic fields generated by ultrasonic equipment;
- **hydrophones** used for measurements made in water;
- **hydrophones** with or without an associated pre-amplifier.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 62127-1, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields* ~~up to 40 MHz~~

## 3 Terms, definitions and symbols

For the purposes of this document, the terms and definitions given in IEC 62127-1, IEC 62127-2 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

### 3.1

#### acoustic pulse waveform

temporal waveform of the **instantaneous acoustic pressure** at a specified position in an acoustic field and displayed over a period sufficiently long to include all significant acoustic information in a single pulse or tone-burst, or one or more cycles in a continuous wave

Note 1 to entry: Temporal waveform is a representation (e.g. oscilloscope presentation or formula) of the **instantaneous acoustic pressure**.

[SOURCE: IEC 62127-1:2022, 3.1]

### 3.2

#### directional response

description of the response of a **hydrophone**, as a function of direction of propagation of the incident plane sound wave, in a specified plane through the **reference centre** and at a specified frequency

Note 1 to entry: Although **directional response** is a complex-valued function, it is generally the modulus of **directional response** that is of most interest and this is commonly presented graphically.

[SOURCE: IEC 60565:2006, 3.5, modified – In the definition, ", generally presented graphically," has been deleted; "electro-acoustic transducer" has been replaced by

"hydrophone"; and "radiated or incident sound" has been replaced by "incident plane sound wave,".]

### 3.3 effective hydrophone size

$a_h$

size of a theoretical receiver **hydrophone** that has a predicted **directional response** function with an angular width equal to the observed angular width

Note 1 to entry: The size is usually a function of frequency. For representative experimental data, see [1].

Note 2 to entry: The **effective hydrophone size** is expressed in metres (m).

Note 3 to entry: For hydrophones with a circular geometry, the **effective hydrophone size** is a radius.

Note 4 to entry: For hydrophones with a rectangular geometry, the **effective hydrophone size** is the half of the largest value of the length or width.

[SOURCE: IEC 62127-1:2022, 3.20]

### 3.4 electric load impedance

$Z_L$

complex electric input impedance (consisting of a real and an imaginary part) to which the **hydrophone** unit output cable is connected or is intended to be connected

Note 1 to entry: The **electric load impedance** is expressed in ohms ( $\Omega$ ).

[SOURCE: IEC 62127-1:2022, 3.22]

### 3.5 end-of-cable

<**hydrophone** or **hydrophone assembly** with integral output cable> relating to the end of the integral output cable

### 3.6 end-of-cable

<**hydrophone** or **hydrophone assembly** without integral output cable> relating to the output connector firmly connected with the **hydrophone** or **hydrophone assembly** and not to an extra cable

### 3.7 end-of-cable loaded sensitivity

$\underline{M}_L(f)$

<of a **hydrophone** or **hydrophone assembly**> quotient of the Fourier transformed **hydrophone** voltage-time signal  $\Phi(u_L(t))$  at the end of any integral cable or output connector of a **hydrophone** or **hydrophone assembly**, when connected to a specified **electric load impedance**, to the Fourier transformed **acoustic pulse waveform**  $\Phi(p(t))$  in the undisturbed free field of a plane wave in the position of the **reference centre** of the **hydrophone** if the **hydrophone** were removed

$$\underline{M}_L(f) = \frac{F(u_L(t))}{F(p(t))} \quad (1)$$

Note 1 to entry: The **end-of-cable loaded sensitivity** is a complex-valued parameter. Its modulus is expressed in units of volt per pascal (V/Pa), its phase angle is expressed in degrees, and represents the phase difference between the electrical voltage and the sound pressure.

### 3.8 end-of-cable open-circuit sensitivity

$\underline{M}_c(f)$

<of a **hydrophone**> quotient of the Fourier transformed **hydrophone** open-circuit voltage-time signal  $\Phi(u_c(t))$  at the end of any integral cable or output connector of a **hydrophone** to the Fourier transformed **acoustic pulse waveform**  $\Phi(p(t))$  in the undisturbed free field of a plane wave in the position of the **reference centre** of the **hydrophone** if the **hydrophone** were removed

$$\underline{M}_c(f) = \frac{F(u_c(t))}{F(p(t))} \quad (2)$$

Note 1 to entry: The **end-of-cable open-circuit sensitivity** is a complex-valued parameter. Its modulus is expressed in units of volt per pascal (V/Pa), its phase angle is expressed in degrees, and represents the phase difference between the electrical voltage and the sound pressure.

### 3.9 end-of-cable loaded sensitivity level

$L_{M_L}(f)$

<of a **hydrophone** or **hydrophone assembly**> twenty times the logarithm to the base 10 of the quotient of the modulus of the **end-of-cable loaded sensitivity**  $|\underline{M}_L(f)|$  to a reference sensitivity  $M_{ref}$

$$L_{M_L}(f) = 20 \log_{10} \frac{|\underline{M}_L(f)|}{M_{ref}} \text{ dB} \quad (3)$$

Note 1 to entry: Commonly used values of the reference sensitivity  $M_{ref}$  are 1 V/μPa or 1 V/Pa.

Note 2 to entry: The **end-of-cable loaded sensitivity level** is expressed in decibels (dB).

### 3.10 free field

sound field in a homogeneous and isotropic medium in which the effects of boundaries are negligible

[SOURCE: IEC 60565:2006, 3.13]

### 3.11 hydrophone geometrical size geometrical size of a hydrophone active element

$a_g$

size defined by the lateral extents of the active element of a **hydrophone**

Note 1 to entry: The **hydrophone geometrical size** is expressed in metres (m).

Note 2 to entry: For **hydrophones** with a circular geometry, the **hydrophone geometrical size** is a radius.

Note 3 to entry: For **hydrophones** with a rectangular geometry, the **hydrophone geometrical size** is a half of the largest value of the length or width.

### 3.12 hydrophone

transducer that produces electric signals in response to pressure fluctuations in water

[SOURCE: IEC 60050-801:2021, 801-32-26]

### 3.13

#### **hydrophone assembly**

combination of **hydrophone** and **hydrophone pre-amplifier**

### 3.14

#### **hydrophone axis**

nominal symmetry axis of the **hydrophone** active element

Note 1 to entry: Unless stated otherwise (explicitly and quantitatively) by the manufacturer, it is understood for the purposes of this document that this is given by the apparent geometrical symmetry axis of the **hydrophone**.

### 3.15

#### **hydrophone pre-amplifier**

active electronic device connected to, or to be connected to, a particular **hydrophone** and reducing its output impedance

Note 1 to entry: A **hydrophone pre-amplifier** requires a supply voltage (or supply voltages).

Note 2 to entry: The **hydrophone pre-amplifier** may have a forward voltage transmission factor of less than one, i.e. it need not necessarily be a voltage amplifier in the strict sense.

### 3.16

#### **reference centre**

point on or near a **hydrophone** about which its acoustic receiving sensitivity is defined

Note 1 to entry: Unless stated otherwise (explicitly and quantitatively) by the manufacturer, it is understood for the purposes of this document that this is given by the geometrical centre of the front surface of the **hydrophone** active element.

[SOURCE: IEC 60565:2006, 3.25, modified – In the definition, "transducer" has been replaced by "hydrophone" and "transmitting responses" has been omitted. The note has been replaced.]

### 3.17

#### **uncertainty**

parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand

Note 1 to entry: See ISO/IEC Guide 98-3:2008 [2], 2.2.3.

[SOURCE: IEC 62127-1:2022, 3.92]

## 4 List of symbols

$a_g$	<b>hydrophone geometrical size</b>
$a_h$	<b>effective hydrophone size</b>
$c$	speed of sound in a medium
$D_H$	model function used during the determination of <b>effective hydrophone size</b> from a measured <b>directional response</b> of a <b>hydrophone</b>
$f$	frequency
$k$	wavenumber, equal to $2\pi/\lambda$
$L_{M_L}(f)$	<b>end-of-cable loaded sensitivity level</b>
$\underline{M}$	general symbol for the complex <b>hydrophone</b> sensitivity, $M =  \underline{M} $ being its modulus and $\arg(\underline{M})$ being its argument (equal to phase angle)
$\underline{M}_c$	<b>end-of-cable open-circuit sensitivity</b>

$M_L$	<b>end-of-cable loaded sensitivity</b>
$Z_h$	complex electric output impedance of a <b>hydrophone</b> or <b>hydrophone assembly</b>
$Z_L$	<b>electric load impedance</b>
$\theta$	angle of incidence of an ultrasonic wave with respect to the <b>hydrophone axis</b>

## 5 Hydrophone characteristics

### 5.1 General

For a full characterization of the hydrophone performance in the frequency range of this document, the following information is required. Examples of information on hydrophone characteristics are provided in Annex A.

NOTE Determination methods are covered in IEC 62127-2.

### 5.2 Basic information

The following shall be briefly stated:

- the basic physical principles of the transduction process, the type of sensor material involved, the form and geometrical dimensions (diameter, thickness) of the **hydrophone** active element and the needle diameter in case of a needle **hydrophone**;
- the configuration and design of the **hydrophone**;
- whether or not a pre-amplifier is associated with the **hydrophone**; if the pre-amplifier can be disconnected from the **hydrophone**, clear information shall be given as to which pre-amplifier type belongs to which **hydrophone** type;
- the nominal direction of ultrasonic incidence in relation to the **hydrophone**.

NOTE The last point is important, as it has been found in the literature [3]<sup>1</sup> that even with membrane **hydrophones**, the response can change upon reversal of the ultrasonic propagation direction in relation to the **hydrophone**.

The following should be briefly stated:

- the lateral and thickness dimensions of the **hydrophone** active element;
- the frequency of the fundamental thickness resonance of the **hydrophone** active element;
- the size and mass of the **hydrophone**;
- the recommended **directional response** model (see 5.6.2) appropriate for the **hydrophone**;
- in the case of a membrane **hydrophone**, the acoustic reflection and transmission factor (preferably as a function of frequency);
- information on preamplifier roll-off at low frequencies outside the hydrophone calibration range, if applicable, to support appropriate calibration data extrapolation in accordance with IEC 62127-1:2022.

General note relating to 5.3 and 5.4: if phase information is available, the phase angle (which equals the argument of the complex **hydrophone** sensitivity) should be stated in addition to the sensitivity (which equals the modulus of the complex **hydrophone** sensitivity), as well as the frequency dependence of the phase angle in addition to the frequency dependence of the sensitivity.

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

### 5.3 Sensitivity

The modulus of the **end-of-cable** sensitivity of the **hydrophone** or **hydrophone assembly** shall be stated in units of volt per pascal (V/Pa) or in decimal submultiples, or as a logarithmic level in decibels (dB) with reference to a stated sensitivity value.

NOTE 1 Refer to 3.9 regarding the definition of the **end-of-cable loaded sensitivity level**.

If a pre-amplifier contributes to the sensitivity value given, this shall be stated.

It shall be stated whether the sensitivity value given is understood as the **end-of-cable open-circuit sensitivity** or as the **end-of-cable loaded sensitivity**. In the latter case, the relevant electric loading conditions shall be stated, i.e. the **electric load impedance**, in order to obtain the stated sensitivity.

The **uncertainty** of the stated sensitivity shall be given.

The frequency interval over which the sensitivity is given and over which the **uncertainty** applies shall be stated. For the purposes of this document, sensitivity and **uncertainty** values may be given separately for several frequency intervals.

The methods by which the sensitivity and its **uncertainty** have been obtained shall be described.

~~The particular water temperature to which the stated sensitivity relates shall be given. The temperature dependence of the sensitivity should shall be described given. The hydrophone sensitivity should shall be stated as a function of the water temperature, at least over the temperature range 19 °C to 25 °C, or the particular water temperature to which the stated sensitivity relates should shall be stated together with the temperature coefficient of the sensitivity.~~

NOTE 4 The former requirement to state the temperature dependence has been modified to become a recommendation for practical reasons. Though being of importance for some applications, the former general requirement here posed disproportional measurement burden to conform with IEC 62127-3. As long as water temperatures at calibration and measurement application differ by one or two degrees only, consideration of the effect within an uncertainty component is reasonable [15],[16],[17].

A recommended calibration period shall be provided in the instructions for use. This recommendation applies unless otherwise stated by specific device application standards.

NOTE 2 A calibration period of one year will be appropriate in most cases.

The **reference centre** shall be stated if the sensitivity does not relate to the geometrical centre of the front surface of the **hydrophone** active element.

NOTE 3 This is particularly important for any phase considerations.

The direction of acoustic incidence shall be stated if the sensitivity does not relate to an incidence in the direction of the **hydrophone axis**.

### 5.4 Frequency response

#### 5.4.1 Stated frequency band

The frequency band claimed for the **hydrophone** or **hydrophone assembly** shall be stated by giving the lower frequency limit and the upper frequency limit. The modulus of the **end-of-cable** sensitivity of the **hydrophone** or **hydrophone assembly** shall be constant over the stated frequency band with a tolerance which shall also be stated.

## 5.4.2 Frequency dependence

The modulus of the **end-of-cable** sensitivity or sensitivity level of the **hydrophone** or **hydrophone assembly** as a function of frequency shall be stated ~~either graphically or~~ as a list of values and over a frequency range containing at least the frequency band claimed under 5.4.1. ~~If it is given as a list of values or as discrete points in a graph,~~ The frequency distance between adjacent points should be low enough so that all important details of the frequency dependence are shown and the sensitivity level does not vary by more than  $\pm 1$  dB between adjacent points. ~~In addition to the list of values, a graphical representation of the data may be provided.~~

The frequency response may be given in terms of absolute sensitivity values or in a relative representation, relative with reference to the absolute sensitivity of the **hydrophone** or **hydrophone assembly** at a certain frequency. In the case of the relative representation, the reference sensitivity and the frequency to which it applies shall be stated.

The phase of the **end-of-cable** sensitivity may be given as a function of frequency ~~either graphically or~~ as a list of values. ~~In addition, a graphical representation of the data may be provided.~~

The statement of the frequency response shall refer to the same conditions (i.e. loaded or open-circuit output of the **hydrophone** or **hydrophone assembly**) as the sensitivity statement in accordance with 5.3.

If the **uncertainty** of the sensitivity values in the frequency response representation differs from the **uncertainty** assessment of 5.3, this shall be clearly stated and the new or additional **uncertainty** shall be given.

If the frequency response is given as a list of absolute sensitivity values (**end-of-cable**, loaded or open-circuit), the sensitivity statement in accordance with 5.3 may be omitted.

NOTE 1 The frequency response can depend on the electric load conditions.

NOTE 2 If, in a practical application, the **hydrophone** or **hydrophone assembly** is used with subsequent electronic components such as an amplifier, oscilloscope, etc., the frequency response of the whole system will also be, of course, influenced by the frequency response of these additional components.

## 5.5 Directional response

### 5.5.1 General

Directional response, and quantities derived from it, are subject to numerous changes relative to the first edition of IEC 62127-3. Details for the rationale behind these changes can be found in Annex B.

### 5.5.2 Determination of the directional response

The **directional response** of the **hydrophone** shall be stated at a minimum of three frequencies where the lowest and highest frequencies are as widely separated as possible.

- The lowest frequency at which the **directional response** of the **hydrophone** has been determined should be the lower limit of the frequency band claimed under 5.4.1. If not, the reason(s) for variation shall be stated.
- The highest frequency at which the **directional response** of the **hydrophone** has been determined should be the upper limit of the frequency band claimed under 5.4.1. If not, the reason(s) for variation shall be stated.

NOTE Experimental determination of **directional response** becomes increasingly challenging as frequency increases. Consequently, the highest frequency at which it is practical to determine the **directional response** of the **hydrophone** can be lower than the upper frequency band claimed under 5.4.1. However, **effective hydrophone size** becomes asymptotic as frequency increases. Therefore, it is likely that there is negligible

variation in **effective hydrophone size** at frequencies above the highest frequency at which **directional response** has been evaluated.

- The midpoint frequency that is within  $\pm 15\%$  of the geometric mean of the lowest and highest frequencies at which **directional response** of the **hydrophone** has been determined is the third frequency at which the directional response shall be stated. If not, the reason(s) for variation shall be stated.
- If the **hydrophone** has a fundamental thickness resonance within range of the lowest and highest frequencies in this subclause 5.5.2, the **directional response** at this frequency shall also be stated. If not, the reason(s) for variation shall be stated.

The **directional response** should be measured by rotating the **hydrophone** about an axis, which passes through the **reference centre** and which is perpendicular to the **hydrophone axis**, at least from  $-35^\circ$  up to  $+35^\circ$  (with the **hydrophone axis** as reference), or at least from the first left-hand minimum to the first right-hand minimum, whichever of the angular spans is the smaller. If this method is used, this shall be done twice, namely about two rotational axes perpendicular to each other. If, in the plane perpendicular to its axis, a **hydrophone** has a certain distinct direction (for example that of the electric leads in the case of a membrane **hydrophone**), the rotational axes should be in this direction and perpendicular to it. If the active element is non-circular, one of the rotational axes shall be in the direction of the largest dimension. The directions of the rotational axes shall be identified on the **hydrophone** using a mark or in the accompanying literature.

The angular increment shall be chosen to ensure that the variation between adjacent points in the **directional response** is not more than 1 dB for frequencies up to 20 MHz. At higher frequencies larger variation between adjacent points may be accepted for practicability.

The measurement of the **directional response** shall be carried out in a quasi-planar wave ultrasonic field.

If the active element is irregular in shape, or has more than two symmetry axes, the **directional response** should be measured around additional axes.

Each of the resulting **directional responses** obtained from the measurements shall be stated.

### 5.5.3 Symmetry of directional response

If, in any of the **directional response** results obtained, the angle between the direction of maximum response and the **hydrophone axis** is greater than one tenth of the angular difference between the left-hand  $-6$  dB direction and the right-hand  $-6$  dB direction, this shall be stated and the deviation-of-axis angle shall be given. The sensitivity level in the direction of the **hydrophone axis** shall be not lower than the maximum in any other direction minus 2 dB.

The symmetry of any **directional response** should be such that if a normalized sensitivity level of  $-6$  dB occurs for some particular direction subtending a certain angle to the direction of maximum sensitivity (0 dB), then the sensitivity level measured on the opposite side subtending the same angle to the direction of maximum sensitivity shall be  $-6$  dB  $\pm$  3 dB.

NOTE Problems in field measurement practice will arise if the direction of maximum **hydrophone** response varies significantly with frequency.

## 5.6 Effective hydrophone size

### 5.6.1 General

From the **directional response** results obtained in accordance with 5.5, a value for the **effective hydrophone size** shall be derived and stated as follows.