



IEC 80601-2-89

Edition 1.0 2025-12

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment -

**Part 2-89: Particular requirements for the basic safety and essential performance
of medical beds for children**

Appareils électromédicaux -

**Partie 2-89: Exigences particulières pour la sécurité de base et les performances
essentielles des lits médicaux pour enfants**

IEC 80601-2-89:2025

<https://standards.iteh.ai/catalog/standards/iso/6f1a17d9-9364-475b-9f53-9c366ad46c8d/iec-80601-2-89-2025>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2025 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search -

webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications, previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Warning! Make sure that you obtained this publication from an authorized distributor.

Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

| | |
|--|-----|
| FOREWORD | 4 |
| INTRODUCTION | 7 |
| 201.1 Scope, object and related standards | 8 |
| 201.2 Normative references | 10 |
| 201.3 Terms and definitions | 11 |
| 201.4 General requirements | 17 |
| 201.5 *General requirements for testing of ME EQUIPMENT | 18 |
| 201.6 Classification of ME EQUIPMENT and ME SYSTEMS | 23 |
| 201.7 ME EQUIPMENT identification, marking and documents | 23 |
| 201.8 Protection against electrical HAZARDS from ME EQUIPMENT | 31 |
| 201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS | 32 |
| 201.10 Protection against unwanted and excessive radiation HAZARDS | 64 |
| 201.11 Protection against excessive temperatures and other HAZARDS | 64 |
| 201.12 Accuracy of controls and instruments and protection against hazardous outputs | 68 |
| 201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT | 68 |
| 201.14 Programmable electrical medical systems (PEMS) | 69 |
| 201.15 Construction of ME EQUIPMENT | 69 |
| 201.16 ME SYSTEMS | 73 |
| 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS | 73 |
| Annexes | 74 |
| Annex AA (informative) Particular guidance and rationale | 75 |
| Annex BB (normative) Additional design requirements and recommendations for MEDICAL BEDS | 90 |
| Annex CC (informative) Particular guidance for assessing RISK of entrapment in V-shaped openings | 94 |
| Annex DD (informative) Guidance and recommendations for periodic inspection | 100 |
| Bibliography | 102 |
| Index of defined terms | 104 |

| | |
|--|----|
| Figure 201.101 – COT, general arrangement (example, schematic presentation only) | 13 |
| Figure 201.102 – CRIB, general arrangement (example, schematic presentation only) | 13 |
| Figure 201.103 – MEDICAL BED, general arrangement (example, schematic presentation only) | 15 |
| Figure 201.104 – HEAD DOWN TILT, Example | 16 |
| Figure 201.105 – FOOT DOWN TILT, Example | 17 |
| Figure 201.106 – Small finger probe \varnothing 5,6 | 18 |
| Figure 201.107 – Entrapment test TOOLS | 20 |
| Figure 201.108 – Loading pad | 21 |
| Figure 201.109 – Impactor | 22 |
| Figure 201.110 – Side impactor TOOL | 23 |
| Figure 201.111 – Graphic symbol for maximum PATIENT weight and SAFE WORKING LOAD | 24 |
| Figure 201.112 – Graphic symbol for mass; weight | 24 |

| | |
|--|----|
| Figure 201.113 – Graphic symbol for machine washable MEDICAL BED | 24 |
| Figure 201.114 – Graphic Symbol for jet stream washable MEDICAL BEDS..... | 25 |
| Figure 201.115 – Graphic symbol for manual cleaning only..... | 25 |
| Figure 201.116 – Description of allowed length of PATIENT | 26 |
| Figure 201.117 – Example of marking for positioning PATIENT in MEDICAL BED..... | 27 |
| Figure 201.118 – MEDICAL BED function controls and actuators: guidelines for creating graphic symbols | 28 |
| Figure 201.119 – Examples of marking on the MEDICAL BED of storage location for wired and wireless PENDANT CONTROLS | 29 |
| Figure 201.120 – Example of MEDICAL BED with segmented or split SIDE RAIL | 33 |
| Figure 201.121 – Example of MEDICAL BED with single piece SIDE RAIL and PROTECTION PANEL | 34 |
| Figure 201.122 – Dimension of handle for LIFTING POLE | 39 |
| Figure 201.123 – Allowable spacing for fingers in areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM | 41 |
| Figure 201.124 – Example using barriers for clearance measurement around the perimeter of the MATTRESS SUPPORT PLATFORM to mitigate PATIENT-finger entrapment..... | 42 |
| Figure 201.125 – Clearance areas | 43 |
| Figure 201.126 – Required minimum radii of edges and corners..... | 44 |
| Figure 201.127 – Retention of loop and mass..... | 45 |
| Figure 201.128 – Lateral stability test along the side of the MEDICAL BED | 48 |
| Figure 201.129 – Longitudinal stability test with removable FOOT BOARD | 48 |
| Figure 201.130 – Longitudinal stability test with fixed HEAD BOARDS and FOOT BOARDS..... | 49 |
| Figure 201.131 – Distribution of SAFE WORKING LOAD per maximum PATIENT weight for tests..... | 53 |
| Figure 201.132 – Position of loading pad and impactor (see Figure 201.108 for loading pad and Figure 201.109 for impactor)..... | 56 |
| Figure 201.133 – Impact to slats and solid elements of MEDICAL BEDS | 58 |
| Figure 201.134 – Application of forces for test of SIDE RAIL..... | 60 |
| Figure 201.135 – Height of PROTECTION PANELS, SIDE RAIL and PROTECTION PERIMETER..... | 63 |
| Figure 201.136 – Direction of movement for rough handling test..... | 70 |
| Figure 201.137 – Configurations of the MATTRESS SUPPORT PLATFORM | 71 |
| Figure 201.138 – Small parts cylinder..... | 73 |
| Figure AA.1 – Example of marking for compatible mattresses specified by the MANUFACTURER | 77 |
| Figure AA.2 – Example of marking for detachable SIDE RAILS specified by the MANUFACTURER | 77 |
| Figure AA.3 – Resultant forces without mattress | 80 |
| Figure AA.4 – Resultant forces with mattress | 80 |
| Figure AA.5 – 40 mm gap measurement of B..... | 80 |
| Figure AA.6 – Angle measurement example of B | 80 |
| Figure AA.7 – Placement of a measurement TOOL for the measurement of D_X | 81 |
| Figure AA.8 – Example of area D_X measurement that passes | 82 |
| Figure AA.9 – Example of area D_X measurement that fails (on limit)..... | 82 |
| Figure AA.10 – Example of area D_X measurement that fails..... | 82 |

| | |
|---|----|
| Figure AA.11 – Example of potential PATIENT entrapment in area A within the SIDE RAIL..... | 83 |
| Figure AA.12 – Example of potential PATIENT entrapment in area A below the SIDE RAIL | 83 |
| Figure AA.13 – Example of potential PATIENT entrapment in area B..... | 83 |
| Figure AA.14 – Example of potential PATIENT entrapment in area C between segmented or split SIDE RAIL | 83 |
| Figure AA.15 – Example of potential PATIENT entrapment in area C between SIDE RAIL and HEAD BOARD | 83 |
| Figure AA.16 – Example of potential PATIENT entrapment in area D..... | 83 |
| Figure AA.17 – Example of potential PATIENT entrapment in area A below a single piece SIDE RAIL | 84 |
| Figure BB.1 – Schematic presentation of under MEDICAL BED clearance | 92 |
| Figure BB.2 – Angles for different sections of the MATTRESS SUPPORT PLATFORM..... | 93 |
| Figure CC.1 – Wedge TOOL..... | 95 |
| Figure CC.2 – V-shaped opening in relation to B..... | 96 |
| Figure CC.3 – Acceptance criteria in relation to area B..... | 96 |
| Figure CC.4 – Positioning of wedge TOOL..... | 97 |
| Figure CC.5 – Acceptance criteria in relation to area C between HEAD BOARD and FOOT BOARD | 98 |
| Figure CC.6 – Acceptance criteria in relation to area C between segmented or split SIDE RAILS | 99 |
| Table 201.101 – Protection against PATIENT entrapment | 35 |
| Table 201.102 – Minimum SAFE WORKING LOADS | 52 |
| Table 201.103 – Protection against inadvertent PATIENT falls and climbing out of MEDICAL BED | 63 |
| Table 201.104 – Allowable maximum temperatures for skin contact with MEDICAL BED APPLIED PARTS | 64 |
| Table 201.105 – Machine washable compliance PROCEDURE | 66 |
| Table AA.1 – Protection against PATIENT entrapment in non-moving parts..... | 79 |
| Table AA.2 – Height of protective barriers according to age | 88 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**Medical electrical equipment -
Part 2-89: Particular requirements for the basic safety
and essential performance of medical beds for children**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 80601-2-89 has been prepared by a Joint Working Group of IEC subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems, and ISO technical committee 173: Assistive products. It is an International Standard.

This publication is published as a double logo standard.

The text of this International Standard is based on the following documents of IEC:

| Draft | Report on voting |
|---------------|------------------|
| 62D/2239/FDIS | 62D/2272/RVD |

Full information on the voting for its approval can be found in the report on voting indicated in the above table. In ISO, the document was approved by XXX P members out of YYY having cast a vote.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- terms defined in clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

IEC 80601-2-89:2025

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.