

---

---

**Dentistry — Powered polymerization  
activators**

*Médecine bucco-dentaire — Activeurs électriques de polymérisation*

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 10650:2018](https://standards.iteh.ai/catalog/standards/iso/c7027c46-90a2-48bb-b9ff-2b60494c0b1f/iso-10650-2018)

<https://standards.iteh.ai/catalog/standards/iso/c7027c46-90a2-48bb-b9ff-2b60494c0b1f/iso-10650-2018>



**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[ISO 10650:2018](https://standards.iteh.ai/catalog/standards/iso/c7027c46-90a2-48bb-b9ff-2b60494c0b1f/iso-10650-2018)

<https://standards.iteh.ai/catalog/standards/iso/c7027c46-90a2-48bb-b9ff-2b60494c0b1f/iso-10650-2018>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword .....	iv
Introduction .....	v
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Classification</b> .....	<b>2</b>
<b>5 Requirements</b> .....	<b>2</b>
5.1 General .....	2
5.1.1 Design .....	2
5.1.2 Connection .....	2
5.1.3 Operating controls .....	3
5.1.4 Cleaning, disinfection and sterilization .....	3
5.1.5 Excessive temperatures .....	3
5.2 Radiant exitance .....	3
5.2.1 Radiant exitance in the 380 nm to 515 nm wavelength region .....	3
5.2.2 Radiant exitance in the wavelength region below 380 nm .....	3
5.2.3 Radiant exitance in the wavelength region above 515 nm .....	3
5.3 Electrical safety requirements .....	3
5.4 Physical and mechanical safety .....	4
5.5 Usability .....	4
5.6 Instructions for use manual .....	4
5.7 Technical Description .....	4
5.8 Marking .....	4
5.9 Packaging .....	4
<b>6 Sampling</b> .....	<b>4</b>
<b>7 Measurement and test methods</b> .....	<b>4</b>
7.1 General .....	4
7.1.1 General provisions for tests .....	4
7.1.2 Atmospheric conditions .....	5
7.2 Inspections .....	5
7.2.1 Visual inspection of powered polymerization activator .....	5
7.2.2 Manual inspection of powered polymerization activator .....	5
7.2.3 Visual inspection of manufacturer's document .....	5
7.3 Measurement of the optical cross-sectional area of the light guide/light-emitting tip .....	5
7.3.1 Apparatus .....	5
7.3.2 Procedure .....	5
7.4 Measurement of radiant exitance .....	5
7.4.1 Method A using spectrometer .....	5
7.4.2 Method B using filters .....	8
<b>8 Information (instructions for use) to be supplied by the manufacturer</b> .....	<b>12</b>
<b>9 Marking</b> .....	<b>13</b>
<b>10 Packaging</b> .....	<b>14</b>
<b>Bibliography</b> .....	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

This second edition cancels and replaces the first edition (ISO 10650:2015), which has been technically revised.

The main changes compared to the previous edition are as follows:

- a test procedure using spectrometer (Method A, [7.4.1](#)) was included;
- a test procedure using filters (Method B, [7.4.2](#)) was modified;
- an upper limit to the radiant exitance for the 380 nm to 515 nm wavelength region was added.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).