

Second edition
2019-12

Corrected version
2020-05

**Implants for surgery — Active
implantable medical devices —**

**Part 7:
Particular requirements for cochlear
and auditory brainstem implant
systems**

*Implants chirurgicaux — Dispositifs médicaux implantables actifs —
Partie 7: Exigences particulières pour les systèmes d'implant
cochléaire et d'implant auditif du tronc cérébral*

[ISO 14708-7:2019](https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019)

<https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019>



Reference number
ISO 14708-7:2019(E)

© ISO 2019

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 14708-7:2019](https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019)

<https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Symbols and abbreviations.....	3
5 General requirements for non-implantable parts.....	3
5.1 General requirements for non-implantable parts.....	3
5.2 General requirements for software.....	3
5.3 Usability of non-implantable parts.....	3
5.4 Data security and protection from harm caused by unauthorized information tampering.....	3
5.5 General requirements for risk management.....	3
5.6 Misconnection of parts of the active implantable medical device.....	3
5.7 Protection against external electrical hazards for fully implantable systems.....	3
6 Inspection and measurement.....	4
6.1 General.....	4
6.2 Measurement of output signal characteristics.....	4
6.3 Measurement of the output signal amplitude and pulse width.....	4
6.4 Impedance measurement accuracy.....	4
6.5 Inductive link characterization.....	4
6.6 Sound processor battery testing.....	4
7 General arrangement of the packaging.....	4
8 General markings for active implantable medical devices.....	4
9 Markings on the sales packaging.....	4
10 Construction of the sales packaging.....	5
11 Markings on the sterile pack.....	5
12 Construction of the non-reusable pack.....	5
13 Markings on the active implantable medical device.....	5
14 Protection from unintentional biological effects being caused by the active implantable medical device.....	6
15 Protection from harm to the patient or user caused by external physical features of the active implantable medical device.....	6
16 Protection from harm to the patient caused by electricity.....	6
17 Protection from harm to the patient caused by heat.....	7
18 Protection from ionizing radiation released or emitted from the active implantable medical device.....	8
19 Protection from unintended effects caused by the device.....	8
20 Protection of the device from damage caused by external defibrillators.....	9
21 Protection of the device from changes caused by high power electrical fields applied directly to the patient.....	9
22 Protection of the active implantable medical device from changes caused by miscellaneous medical treatments.....	10
23 Protection of the active implantable medical device from mechanical forces.....	18

24	Protection of the active implantable medical device from damage caused by electrostatic discharge	22
25	Protection of the active implantable medical device from damage caused by atmospheric pressure changes	22
26	Protection of the active implantable medical device from damage caused by temperature changes	23
27	Protection of the active implantable medical device from electromagnetic non-ionising radiation	23
27.1	Protection from static magnetic fields.....	23
27.2	Radiated magnetic field test for frequencies 16,6 Hz to 27 MHz.....	23
27.3	Radiated electric field test for frequencies 10 MHz to 2,7 GHz.....	25
27.4	General test configuration and setup.....	25
27.4.1	Test configuration and setup.....	25
27.4.2	Operating functions, modes and settings.....	26
27.4.3	Patient physiological simulation.....	26
27.5	Acceptance Criteria.....	26
28	Accompanying documentation	27
	Annex A (informative) General guidance and rationale	31
	Annex B (informative) Relationship between the fundamental principles in ISO/TR 14283 and the clauses of this document	42
	Annex C (informative) Notes on EN 45502-2-3 (basis for this document)	61
	Annex D (informative) Notes on EMI measurements to demonstrate compliance with Clause 27	62
	Bibliography	66

iteh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 14708-7:2019](#)

<https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 6, *Active implants*.

This second edition cancels and replaces the first edition (ISO 14708-7:2013), which has been technically revised. The main changes compared to the previous edition are as follows:

- alignment to the revised ISO 14708-1:2014;
- significant changes to [Clauses 17, 22](#) and [27](#);
- many clauses have been replaced by references to ANSI/AAMI CI86:2017.

A list of all part in the ISO 14708 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 7063:2018 incorporates the following correction: in [16.2](#), the word "direct" was added in the following sentence: "The maximum direct current density at the *electrode contact* opening shall be no more than 0,75 $\mu\text{A}/\text{mm}^2$ ".

Introduction

This document specifies particular requirements for active implantable medical devices used to treat hearing impairment via electrical stimulation (for example, *cochlear implant systems* or *auditory brainstem implant systems*), to provide basic assurance of safety for both patients and users.

A *cochlear implant system* or *auditory brainstem implant system* is an active implantable medical device comprising implantable and *non-implantable parts* (external parts). The power source can be externally derived or from an internal battery. The *implant system* is designed to restore hearing via electrical stimulation of the auditory pathways. Externally or internally processed acoustic information is converted to electrical stimulation signals which are delivered via one or more electrodes. The working parameters of the device may be adjusted via a non-implantable accessory.

This document is relevant to all parts of *implant systems*, including accessories.

The requirements of this document supplement or modify those of ISO 14708-1:2014.

In this document, terms printed in italic letters are used as defined in [Clause 3](#). Where a defined term is used as a qualifier in another term, it is not printed in italic letters unless the concept thus qualified is also defined.

Information is also provided in [Annex B](#) that explains the relationship between ISO/TR 14283, ISO 14708-1:2014 and this document.

Notes on EN 45502-2-3 (basis for this document) is provided in [Annex C](#) for information.

ITeH Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 14708-7:2019](#)

<https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019>

Implants for surgery — Active implantable medical devices —

Part 7: Particular requirements for cochlear and auditory brainstem implant systems

1 Scope

This document specifies requirements that are applicable to those active implantable medical devices that are intended to treat hearing impairment via electrical stimulation of the auditory pathways. Devices which treat hearing impairment via means other than electrical stimulation are not covered by this document.

The tests that are specified in this document are type tests and are to be carried out on samples of a device to show compliance.

This document is also applicable to *non-implantable parts* and accessories of the devices (see NOTE).

The electrical characteristics of the implantable part are determined by either the appropriate method detailed in this document or by any other method demonstrated to have an accuracy equal to, or better than, the method specified. In the case of dispute, the method detailed in this document applies.

NOTE A device that is commonly referred to as an active implantable medical device can in fact be a single device, a combination of devices, or a combination of a device or devices and one or more accessories. Not all of these parts are required to be either partially or totally implantable, this document specifies those requirements of *non-implantable parts* and accessories which could affect the safety or performance of the implantable part.

<https://standards.iteh.ai/catalog/standards/iso/f7ce3b9d-2b2f-4e30-986d-ab0b0070d463/iso-14708-7-2019>

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/TS 10974, *Assessment of the safety of magnetic resonance imaging for patients with an active implantable medical device*

ISO 14708-1:2014, *Implants for surgery — Active implantable medical devices — Part 1: General requirements for safety, marking and for information to be provided by the manufacturer*

IEC 60068-2-31, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60601-1-2, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests*

IEC 61000-4-2, *Electromagnetic compatibility (EMC) — Part 4-2: Testing and measurement techniques — Electrostatic discharge immunity test*

EN 1593, *Non-destructive testing — Leak testing — Bubble emission techniques*

EN 13185, *Non-destructive testing — Leak testing — Tracer gas method*