
Cardiovascular implants and artificial organs — Requirements for single-use tubing packs for cardiopulmonary bypass and extracorporeal membrane oxygenation (ECMO)

Implants cardiovasculaires et organes artificiels — Exigences pour les paquets de tubes à usage unique pour pontage cardiopulmonaire et oxygénation des membranes extracorporelles

Sample Document

get full document from standards.iteh.ai



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	3
4.1 Biological characteristics	3
4.1.1 Sterility and non-pyrogenicity	3
4.1.2 Biocompatibility	3
4.2 Physical characteristics	3
4.2.1 General	3
4.2.2 Dimensions	3
4.2.3 Material properties	3
4.3 Performance characteristics	3
4.3.1 Priming volume	3
4.3.2 Life to failure testing	4
4.3.3 Spallation	4
4.3.4 Shelf life	4
5 Tests and measurements	4
5.1 General	4
5.2 Biological characteristics	4
5.2.1 Sterility and non-pyrogenicity	4
5.2.2 Biocompatibility	4
5.3 Physical characteristics	5
5.3.1 Blood pathway integrity	5
5.3.2 Connections	5
5.3.3 Tubing material property testing	5
5.4 Performance characteristics	5
5.4.1 Tubing life	5
5.4.2 Spallation in tubing used in roller pumps	5
5.4.3 Shelf life	6
6 Information supplied by the manufacturer	6
6.1 Information on the tubing pack	6
6.1.1 Information on the unit container	6
6.1.2 Information on the shipping container	6
6.2 Information on the accompanying documents	7
6.3 Information in the accompanying documents in a prominent form	7
6.4 Information to be provided by manufacturer upon request	7
7 Packaging	7
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This second edition cancels and replaces the first edition (ISO 15676:2005), which has been technically revised.

Introduction

The intent of this document is to ensure that medical grade tubing in single-use tubing packs for the transfer of blood and fluid during the period of cardiopulmonary bypass (CPB) and extracorporeal membrane oxygenation (ECMO) is adequately tested for both safety and function. The user commonly provides the specifications for the tubing pack. Furthermore, the purpose of this document is to ensure that the tubing pack characteristics be appropriately disclosed in the labelling and manufacturing information package. Tubing performance characteristics are specifically addressed within the context of this document as a component part of a single-use tubing pack.

This document therefore contains recommended procedures to evaluate such medical grade tubing intended for use during CPB procedures and ECMO. Test procedures to determine the material characteristics, the useful life of the tubing when used in a roller pump, and cleanliness are described. The limits for these characteristics are not specified.

This document also includes minimum reporting requirements. Ready identification of the performance characteristics should assist the user in the selection of such medical grade tubing for the procedure appropriate to the patient and procedure. This information may be useful in a clinic's quality control process that aims to improve the safety of CPB and ECMO procedures.

This document makes reference to other International Standards, which references methods for the determination of characteristics common to medical devices.

Requirements for animal and clinical studies are not included in this document. Such studies, however, may be part of a manufacturer's quality system.

This document contains only those requirements that are specific to such medical grade tubing for use during CPB and ECMO. Non-specific requirements are covered by reference to other International Standards listed in the Normative References section.

get full document from standards.iteh.ai