

---

---

**Medical devices — Connectors  
for reservoir delivery systems for  
healthcare applications —**

**Part 7:  
Connectors for intravascular infusion**

*Dispositifs médicaux — Connecteurs pour systèmes de livraison de  
réservoir pour des applications de soins de santé —*

*Partie 7: Connecteurs pour perfusion intravasculaire*

Document Preview

ISO 18250-7:2018

<https://standards.iteh.ai/catalog/standards/iso/16676ca9-fc4c-4294-837a-e77dfa4602df/iso-18250-7-2018>



**iTeh Standards**  
**(<https://standards.itih.ai>)**  
**Document Preview**

ISO 18250-7:2018

<https://standards.itih.ai/catalog/standards/iso/16676ca9-fc4c-4294-837a-e77dfa4602df/iso-18250-7-2018>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

|                                                                                                                                                                      |           |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|
| Foreword .....                                                                                                                                                       | iv        |
| Introduction .....                                                                                                                                                   | vi        |
| <b>1 *Scope .....</b>                                                                                                                                                | <b>1</b>  |
| <b>2 Normative references .....</b>                                                                                                                                  | <b>1</b>  |
| <b>3 Terms and definitions .....</b>                                                                                                                                 | <b>2</b>  |
| <b>4 General requirements .....</b>                                                                                                                                  | <b>2</b>  |
| 4.1 Non-interconnectability through conformance to this document .....                                                                                               | 2         |
| 4.2 * Intravascular RESERVOIR CONNECTOR types .....                                                                                                                  | 2         |
| <b>5 Material requirements .....</b>                                                                                                                                 | <b>3</b>  |
| 5.1 CONNECTOR Type 1: spikes and administration ports .....                                                                                                          | 3         |
| 5.2 CONNECTOR Type 2: LUERS .....                                                                                                                                    | 3         |
| <b>6 Dimensional requirements .....</b>                                                                                                                              | <b>3</b>  |
| 6.1 CONNECTOR Type 1: Spikes and Administration Ports .....                                                                                                          | 3         |
| 6.2 CONNECTOR Type 2: LUERS .....                                                                                                                                    | 3         |
| <b>7 Performance requirements .....</b>                                                                                                                              | <b>3</b>  |
| 7.1 * CONNECTOR Type 1: Spikes and Administration Ports .....                                                                                                        | 3         |
| 7.2 CONNECTOR Type 2: LUERS .....                                                                                                                                    | 3         |
| <b>Annex A (informative) Rationale and guidance .....</b>                                                                                                            | <b>4</b>  |
| <b>Annex B (normative) Design of CONNECTORS* .....</b>                                                                                                               | <b>13</b> |
| <b>Annex C (informative) Assessment of RESERVOIR CONNECTORS and their attributes with CONNECTIONS to medical devices within this intravascular APPLICATION .....</b> | <b>15</b> |
| <b>Annex D (informative) Summary of usability and requirements for CONNECTORS for intravascular RESERVOIRS .....</b>                                                 | <b>16</b> |
| <b>Annex E (informative) Summary of RESERVOIR CONNECTOR design requirements for INTRAVASCULAR INFUSION .....</b>                                                     | <b>19</b> |
| <b>Annex F (informative) Summary of engineering analysis and residual MISCONNECTIONS for CONNECTORS for intravascular RESERVOIRS .....</b>                           | <b>24</b> |
| <b>Annex G (informative) Alternative spikes .....</b>                                                                                                                | <b>30</b> |
| <b>Bibliography .....</b>                                                                                                                                            | <b>34</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing international standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. [www.iso.org/directives](http://www.iso.org/directives)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. [www.iso.org/patents](http://www.iso.org/patents)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](https://standards.iteh.ai/Foreword-Supplementary-information)

This document was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*.

A list of all the parts in the ISO 18250 series can be found on the ISO website.

In this document, the following print types are used:

- Requirements and definitions: roman type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in smaller type.
- TERMS DEFINED IN STANDARD OR AS NOTED: SMALL CAPITALS.
- *Compliance requirements: italic type.*

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document,
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document, and
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in [Annex A](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

**iTeh Standards**  
**(<https://standards.itih.ai>)**  
**Document Preview**

ISO 18250-7:2018

<https://standards.itih.ai/catalog/standards/iso/16676ca9-fc4c-4294-837a-e77dfa4602df/iso-18250-7-2018>

## Introduction

During the development of the standard for enteral small-bore CONNECTORS (ISO 80369-3), it became clear that the risk of MISCONNECTIONS was not limited to the PATIENT access CONNECTORS and that the whole enteral system needed to be considered. The possible MISCONNECTION between enteral feed RESERVOIR access CONNECTORS and the spikes typically used on intravenous administration sets was also reviewed. However, as feed RESERVOIR CONNECTORS are not exactly within the definition of small-bore CONNECTORS, it was decided to develop a separate standard (ISO 18250-3) for these CONNECTORS, taking into account the risks of MISCONNECTIONS with other devices such as intravenous (IV) bags.

During the development of ISO 18250-3, it became apparent that the ubiquitous intravenous spike was specified in various ISO medical device standards but that the geometry and materials requirements for the female port on the intravenous RESERVOIR were not defined. However, the performance of this female port was defined in various ISO standards.

This document, therefore, specifies the design, dimensions and materials for the female port. It makes reference to existing performance standards. It also includes tests to validate that the spike and the female port do not interconnect with the other RESERVOIR CONNECTORS of the ISO 18250 series.

This document also includes analysis sufficient to include traditional LUER CONNECTORS of ISO 80369-7 as permissible RESERVOIR CONNECTORS for intravascular APPLICATIONS.

This document is not a device standard as it specifies only the requirements of the interfaces for CONNECTORS used in intravascular RESERVOIRS and INTRAVASCULAR INFUSION SETS.

**ITeH Standards**  
(<https://standards.iteh.ai>)  
**Document Preview**

[ISO 18250-7:2018](https://standards.iteh.ai/catalog/standards/iso/16676ca9-fc4c-4294-837a-e77dfa4602df/iso-18250-7-2018)

<https://standards.iteh.ai/catalog/standards/iso/16676ca9-fc4c-4294-837a-e77dfa4602df/iso-18250-7-2018>