
**Water quality — Determination
of short-chain polychlorinated
alkanes (SCCPs) in sediment, sewage
sludge and suspended (particulate)
matter — Method using gas
chromatography-mass spectrometry
(GC-MS) and electron capture
negative ionization (ECNI)**

*Qualité de l'eau — Détermination des alcanes polychlorés à chaîne
courte dans les sédiments et matières en suspension (particules) —
Méthode par chromatographie en phase gazeuse-spectrométrie de
masse (CPG-SM) et ionisation chimique négative (ICN)*



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Principle.....	1
4 Interferences.....	2
5 Reagents.....	3
6 Apparatus.....	6
7 Sampling and sample pretreatment.....	7
8 Procedure.....	7
8.1 Extraction.....	7
8.2 Extract clean-up.....	7
8.3 Measurement and integration of the chromatogram.....	8
8.4 Calibration.....	8
8.4.1 General.....	8
8.4.2 Basic calibration.....	9
8.4.3 Identification and quantification with mass fragment combinations.....	10
8.4.4 Calculation of the results.....	10
8.4.5 Quality checks for internal standardization.....	11
9 Expression of results.....	11
10 Test report.....	11
Annex A (normative) Programme for pressurized liquid extraction.....	12
Annex B (informative) Explanation of the calibration of the sum of SCCPs with multiple linear regression.....	13
Annex C (informative) Examples of typical GC-MS conditions.....	18
Annex D (normative) Alternative clean-up with column chromatography.....	22
Annex E (informative) Alternative clean-up with gel chromatography.....	24
Annex F (informative) Typical chromatograms of standard solutions 1 µg/ml.....	25
Annex G (informative) Presentation of goodness of fit of the calibration.....	26
Annex H (informative) Example for recoveries of quality assurance solutions.....	27
Annex I (informative) Chromatograms of real samples.....	29
Annex J (informative) Performance data.....	31
Annex K (informative) Quality control check solutions.....	32
Bibliography.....	33

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

get full document from standards.iteh.ai

Introduction

The user should be aware that particular problems might require the specifications of additional marginal conditions.

Sample Document

get full document from standards.iteh.ai

Sample Document

get full document from standards.iteh.ai

Water quality — Determination of short-chain polychlorinated alkanes (SCCPs) in sediment, sewage sludge and suspended (particulate) matter — Method using gas chromatography-mass spectrometry (GC-MS) and electron capture negative ionization (ECNI)

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

1 Scope

This International Standard specifies a method for the quantitative determination of the sum of short-chain polychlorinated *n*-alkanes also known as short-chain polychlorinated paraffins (SCCPs) in the carbon bond range, *n*-C₁₀ to *n*-C₁₃, inclusive in mixtures with chlorine mass fractions (“contents”) between 50 % and 67 %, including approximately 6 000 of approximately 8 000 congeners.

This method is applicable to the determination of the sum of SCCPs in sediment and suspended (particulate) matter, sewage sludge, and soil using gas chromatography-mass spectrometry with electron capture negative ionization (GC-ECNI-MS).

Depending on matrix and the detection capabilities of the GC-ECNI-MS, the method can be applied to samples containing, e.g. 0,03 µg/g to 3 µg/g sum of SCCPs.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-12, *Water quality — Sampling — Part 12: Guidance on sampling of bottom sediments*

ISO 5667-13, *Water quality — Sampling — Part 13: Guidance on sampling of sludges*

ISO 5667-17, *Water quality — Sampling — Part 17: Guidance on sampling of bulk suspended solids*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

ISO 12010, *Water quality — Determination of short-chain polychlorinated alkanes (SCCPs) in water — Method using gas chromatography-mass spectrometry (GC-MS) and negative-ion chemical ionization (NCI)*

ISO/TS 13530, *Water quality — Guidance on analytical quality control for chemical and physicochemical water analysis*

3 Principle

Determination of the sum of SCCPs in the carbon bond range, *n*-C₁₀ to *n*-C₁₃, inclusive in technical and environmental transposed mixtures with chlorine mass fractions (“contents”) between 50 % and 67 %

(e.g. approximately 3 to 10 chlorine atoms per molecule) and independent of the C-number distribution pattern of the congeners. No recognition of the chlorine content is necessary.

The samples are fortified with an internal standard and extracted using pressurized liquid extraction with an organic solvent. The sample extraction procedure is followed by a clean-up procedure by column chromatography and gel permeation chromatography to eliminate interfering compounds. Gas chromatography (GC) is undertaken using a short capillary column within a short retention time range. The detection of selected mass fragments is carried out using mass spectrometry (MS) in selected ion-monitoring mode using electron capture negative ionization mode (ECNI). The mass fragments and the compositions of the calibration solutions used in this International Standard are essential for the analysis of the sum of SCCPs.

The selected ion chromatogram is integrated over the full retention time range of the SCCPs. The quantification of the sum of SCCPs is carried out after establishing a calibration by a multiple linear regression measuring solutions of different SCCP mixtures fortified with an internal standard.

The analysed sum of SCCPs includes the variety of SCCPs with their differing chlorine content and C-number distribution patterns as found in technical mixtures, as well as compositions in the environment. The calibration requires at least three differently composed standard mixtures.

These standard mixtures mimic different mixtures found in the environment. Only the multiple linear regression quantification with these specific mixtures enables the quantification of the variety of observed mixtures of SCCP in the environment and in technical compositions described in the scope and in Reference [6]. It is not possible to use only one standard mixture for that complex task.

4 Interferences

Non-specific matrix interferences, as well as interferences from other environmental situations, are dealt with using the given clean-up procedure. A further reduction of matrix effects can be achieved by reducing the mass spectrometric resolution to, e.g., 0,4 amu, which is often possible with a quadrupole mass spectrometer. The exact masses are 374,9588; 410,9169; and 422,9355 (see Reference [6]).

Applying the entire procedure, a selection of chlorinated pollutants has been tested and found not to cause interferences below the concentrations given in [Table 1](#).

Table 1 — Highest concentration level which causes no interferences higher than the limit of quantification of 0,03 µg/g

Potential interfering compounds	Highest concentration level which causes no interferences higher than the limit of quantification of 0,03 µg/g
Aroclor 1262 ^a	0,25 µg/g
Aroclor 1242 ^a	2 µg/g
Aroclor 1221 ^a	2 µg/g
Campheclor (toxaphene)	0,35 µg/g
Halowax 1014 ^a	2 µg/g
Halowax 1051 ^a	0,08 µg/g
MCCP (medium-chain chlorinated <i>n</i> -alkanes) 42 %	2 µg/g
MCCP (medium-chain chlorinated <i>n</i> -alkanes) 52 %	1,2 µg/g
MCCP (medium-chain chlorinated <i>n</i> -alkanes) 57 %	2 µg/g

^a Aroclor 1262, Aroclor 1242, Aroclor 1221, Halowax 1014, and Halowax 1051 are products commercially available. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of these products.

5 Reagents

Use solvents and reagents of sufficient purity, i.e. with negligibly low concentrations of SCCPs, e.g. lower than the limit of detection of the method.

Check blanks regularly over the entire procedure to establish proper analytical quality control.

5.1 Solvents for extraction, column chromatography, and preparation of stock solutions.

The solvent for extraction is *n*-heptane. Other non-polar solvents, e.g. *n*-hexane (C₆H₁₄), cyclohexane (C₆H₁₂) can be used if the extraction efficiency is comparable with those of *n*-heptane.

For conditioning of the clean-up columns, use mixtures of *n*-heptane and propanone (acetone) as described below.

5.2 Reference SCCP stock solutions.

Use commercially available solutions, e.g. in cyclohexane or *n*-hexane, of the single mixtures of SCCP congeners with defined carbon chain length and with different defined chlorine contents (see [Table 2](#), first two columns). Alternatively, use commercially available ready mixed solutions with the same composition.

Mixtures of synthetic solutions are used to simulate environmentally occurring SCCPs or technical products of SCCPs. For example, the synthetic mixed calibration stock solution "Lake Ontario water" is mixed to resemble a Lake Ontario water as reported in Reference [5]. Its characteristic is a relatively high content of C₁₀ to C₁₂, especially C₁₂ and a low chlorine content as partly reported in water samples too. The synthetic mixed calibration stock solution "Perch" simulates a C-number distribution found in a perch (see Reference [6]). The standard mixture "Sediment Drevnice" simulates a natural mixture reported about a sediment of the river Drevnice (see Reference [7]) with a high content of C₁₃ and a higher chlorine content.

The compositions of the calibration mixtures, as well as of the independent quality assurance solutions, are mandatory to achieve the quantification of the variety of SCCP-mixtures. An example for recoveries of quality assurance solutions is given in [Annex H](#).

Prepare the solutions "Lake Ontario water", "Perch", and "Sediment Drevnice" according to [Table 2](#).

Table 2 — Reference substances stock solutions

Standard solutions, e.g. in <i>n</i> -heptane			Synthetic mixed standard solutions which resemble environmental mixtures		
<i>n</i> -alkane chain length	Chlorine content (%) of the individual C-number mixtures	Mean number of chlorines in the molecules (calculated)	"Lake Ontario water"	"Perch"	"Sediment Drevnice"
Chlorine content calculated (%)			50,2	60,6	65,0
Composition, ng/ml					
C ₁₀	44,82	3,22			
C ₁₀	50,18	3,97	1 000		
C ₁₀	55,00	4,79	1 000		
C ₁₀	60,09	5,86		500	
C ₁₀	65,02	7,16		1 100	280
C ₁₁	45,50	3,63	1 000		
C ₁₁	50,21	4,37	1 000		

Table 2 (continued)

Standard solutions, e.g. in <i>n</i> -heptane			Synthetic mixed standard solutions which resemble environmental mixtures		
<i>n</i> -alkane chain length	Chlorine content (%) of the individual C-number mixtures	Mean number of chlorines in the molecules (calculated)	"Lake Ontario water"	"Perch"	"Sediment Drevnice"
Chlorine content calculated (%)			50,2	60,6	65,0
Composition, ng/ml					
C ₁₁	55,20	5,31		600	
C ₁₁	60,53	6,55		1 000	500
C ₁₁	65,25	7,94		3 000	660
C ₁₂	45,32	3,93	2 000		
C ₁₂	50,18	4,76	2 000	800	
C ₁₂	55,00	5,74	2 000	2 000	
C ₁₂	65,08	8,59		900	1 000
C ₁₂	69,98	10,62			830
C ₁₃	44,90	4,19			
C ₁₃	50,23	5,16			
C ₁₃	55,03	6,22			
C ₁₃	59,98	7,56		100	730
C ₁₃	65,18	9,34			6 000
Sum of SCCP (ng/ml)			10 000	10 000	10 000

The chlorine content (third column) of the mixtures is calculated as the weighted mean.

Store the prepared solutions in a refrigerator at 2 °C to 6 °C.

5.3 Internal standard stock solutions from individual congeners.

Use commercially available individual congener standard solutions and prepare a stock solution in propanone (acetone) (5.1) at a concentration of, for example, 1 µg/ml.

- Individual SCCP congeners with chlorine contents of between 50 % and 67 % are suitable as internal standards if the mass trace in mass spectrometric detection is not affected by matrix components, e.g. 1,1,1,3,11,13,13,13-octachlorotridecane at e.g. 0,1 µg/ml.

NOTE 1 The different individual SCCP congeners used as internal standard substances probably contribute in environmental samples to the sum of SCCPs. Nevertheless, the contribution is approximately <1 % which means that the enhancement of the measurement uncertainty is negligible.

NOTE 2 Different individual SCCP congeners can produce different response factors, hence, it can be necessary for different concentrations to be used.

The solutions should be stored in a refrigerator at 2 °C to 6 °C.

5.4 Calibration solutions.

Use the standard mixtures according to Table 2. Prepare a minimum of nine calibration solutions (see Table 3) with concentrations according to the detection capability of the mass spectrometer. Combine and dilute the solutions (5.2) and the internal standard stock solution (5.3) with *n*-heptane to produce solutions for the calibration range, e.g. as shown in Table 3.

Table 3 — Calibration solutions

Mixture	"Lake Ontario" water	"Perch"	"Sediment Drevnice"	Internal standard
				1,1,1,3,11,13,13,13-octachlorotridecane
	µg/ml	µg/ml	µg/ml	µg/ml
Sum of SCCPs, µg/ml				
0,15	0,15			0,1
0,15		0,15		0,1
0,15			0,15	0,1
0,6	0,6			0,1
0,6		0,6		0,1
0,6			0,6	0,1
1	1			0,1
1		1		0,1
1			1	0,1
2	2			0,1
2		2		0,1
2			2	0,1
3	3			0,1
3		3		0,1
3			3	0,1

The solutions can be stored in a refrigerator, for at least four weeks. Check the concentration of the calibration solutions against an independently prepared standard prior to use.

Quality control check solutions can be prepared to check the calibration independently. For that, use the calibration mixtures as given in ISO 12010 (see [Annex K](#)). These mixtures are commercially available (e.g. in cyclohexane or *n*-hexane).

5.5 Extraction auxiliary and clean-up materials.

5.5.1 High purity silica sand, grain size 600 µm to 850 µm, blank free from contamination.

5.5.2 Copper powder, grain size <63 µm.

5.5.3 2 mol/l hydrochloric acid.

5.5.4 Al₂O₃, neutral, high activity, (10 % water).

5.5.5 Glass wool.

5.6 Operating gases, for GC-MS of high purity and in accordance with manufacturer's specifications.

5.7 Nitrogen, N₂, purity ≥99,996 % volume fraction for concentrating the solutions.

5.8 Sodium sulfate, anhydrous, Na₂SO₄, powdered.

6 Apparatus

Clean all glassware by rinsing with acetone (propanone) ([5.1](#)).

6.1 Wide-necked bottle, 1 000 ml up to 5 000 ml capacity for wet sediment or sludge.

6.2 Freeze drying apparatus.

6.3 Deep freezer.

6.4 Mortar and pestle or a grinding mill.

6.5 Drying ovens, capable of maintaining temperatures in the ranges of 100 °C to 400 °C for baking and storage of clean-up materials, for baking of glassware, and for dry residue determination of samples.

6.6 Sieve shaker with appropriate sieve meshes (aperture size), e.g. 2 mm.

6.7 Desiccator.

6.8 Pressurized liquid extractor (PLE) and filter suited for the device.

6.9 Evaporation device.

For example, rotary evaporator, turboevaporator, or vacuum concentration device.

6.10 Glass columns for chromatographic clean-up.

6.11 GPC clean-up system (with modular design).

6.11.1 Pump, sampling injector, sample rack, fraction collector.

6.11.2 Column, Shodex CLNpakPAE 800 AC^{®1)} 8,0 mm × 300 mm, or equivalent.

6.12 Volumetric cylinders, 250 ml and 500 ml.

6.13 Volumetric flasks, 1 ml, 2 ml, 10 ml, and 25 ml.

6.14 Pasteur pipettes, e.g. 2 ml.

6.15 Syringes, 2 µl, 5 µl, 10 µl, and 50 µl, volume precision ±2 %.

6.16 Sample vials.

Amber glass with fluoropolymer-lined screw-cap is most suitable.

6.17 Gas chromatograph, with a splitless injection port coupled to a mass spectrometer (GC-MS) with chemical ionization (ECNI) and appropriate reactant gas (e.g. CH₄).

1) Shodex CLNpakPAE 800 AC is a product commercially available. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product.