



International
Standard

ISO 18862

**Coffee and coffee products —
Determination of acrylamide —
Methods using high-performance
liquid chromatography with
tandem mass spectrometric
detection (HPLC-MS/MS) and
gas chromatography with mass
spectrometric detection (GC-MS)
after derivatization**

**Second edition
2025-11**

<https://standards.iteh.ai/catalog/standards/iso/accc401e-3089-472f-b49f-bc806a1bb11e/iso-18862-2025>

*Café et dérivés du café — Dosage de l'acrylamide — Méthodes
par chromatographie liquide à haute performance avec
détection par spectrométrie de masse en tandem (CLHP-SM/
SM) et chromatographie en phase gazeuse avec détection par
spectrométrie de masse (CG-SM) après dérivation*

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

ISO 18862:2025

<https://standards.itih.ai/catalog/standards/iso/accc401e-3089-472f-b49f-bc806a1bb11e/iso-18862-2025>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	2
6 Apparatus	3
7 Sampling	4
8 Procedure	4
8.1 General	4
8.2 Preparation of the sample extract	4
8.3 Clean-up of the extracts	5
8.3.1 Carrez precipitation	5
8.3.2 Solid-phase extraction	5
8.4 HPLC-MS/MS measurement	5
8.4.1 High-performance liquid chromatography (HPLC)	5
8.4.2 Identification and quantification by mass spectrometry (HPLC-MS/MS)	6
8.5 Measurement with GC-MS	6
8.5.1 Derivatization and sample preparation for gas chromatography	6
8.5.2 Gas chromatography	7
8.5.3 Identification and quantification by mass spectrometry	7
9 Calibration	7
9.1 General advice	7
9.2 Determination of linearity and definition of the working range	7
9.3 Calibration with internal standard solution	7
9.4 Determination of the laboratory specific recovery	8
10 Evaluation	8
10.1 Criteria for identification	8
10.2 Calculation and final results	8
11 Precision data	9
11.1 General	9
11.2 Repeatability	9
11.3 Reproducibility	9
11.4 Recovery	9
12 Measurement uncertainty	9
13 Test report	9
Annex A (informative) Performance characteristics	11
Annex B (informative) Examples of absorber materials	12
Annex C (informative) Examples of columns and analysis conditions	13
Annex D (informative) Examples for sample preparation and chromatographic conditions using HPLC-MS/MS	19
Bibliography	25

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 15, *Coffee*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 275, *Food analysis - Horizontal methods*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 18862:2016), which has been technically revised.

The main changes are as follows:

- [Annex D](#) with examples for sample preparation and chromatographic conditions using HPLC-MS/MS has been added.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Coffee and coffee products — Determination of acrylamide — Methods using high-performance liquid chromatography with tandem mass spectrometric detection (HPLC-MS/MS) and gas chromatography with mass spectrometric detection (GC-MS) after derivatization

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to take appropriate measures for ensuring the safety and health of the personnel prior to application of this document and to fulfil statutory requirements for this purpose.

1 Scope

This document specifies methods for the determination of acrylamide in coffee and coffee products by extraction with water, clean-up by solid-phase extraction (SPE) and determination by high-performance liquid chromatography with tandem mass spectrometric detection (HPLC-MS/MS) and gas chromatography with mass spectrometric detection (GC-MS) after derivatization. The methods were validated in a validation study for roasted coffee, soluble coffee, coffee substitutes and coffee products with ranges from 53 µg/kg to 612,1 µg/kg.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

<https://standards.iteh.ai/catalog/standards/iso/accc401e-3089-472f-b49f-bc806a1bb11e/iso-18862-2025>
ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Principle

The coffee sample is extracted with water or, in the case of soluble products, dissolved in water. A clean-up by SPE is employed to remove interfering matrix compounds. Two alternative methods can be used for the determination: HPLC-MS/MS or, after a bromination of the acrylamide, GC-MS. In both cases, solutions of isotopic labelled internal standard are used.