
**Microscopes — Definition and
measurement of illumination
properties —**

**Part 1:
Image brightness and uniformity in
bright field microscopy**

*Microscopes — Définition et mesure des propriétés d'éclairage —
Partie 1: Luminosité et uniformité de l'image en microscopie à champ
clair*

get full document from standards.iteh.ai



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
1 Scope	1
2 Measurands	1
2.1 General.....	1
2.2 Brightness.....	1
2.3 Uniformity.....	2
3 Measurement procedure	2
3.1 General.....	2
3.2 Diaphragm settings.....	2
3.3 Brightness (illuminance and irradiance).....	2
3.4 Uniformity.....	3
3.5 Spectral information.....	3
4 Information provided to the user	3
Bibliography	5

Sample Document

get full document from standards.iteh.ai

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 5, *Microscopes and endoscopes*.

ISO 19056 consists of the following parts, under the general title *Microscopes — Definition and measurement of illumination properties*:

— *Part 1: Image brightness and uniformity in bright field microscopy*

Microscopes — Definition and measurement of illumination properties —

Part 1:

Image brightness and uniformity in bright field microscopy

1 Scope

This part of ISO 19056 specifies procedures for the measurement of image brightness and uniformity for bright field light microscopy. These measurements for image brightness and uniformity are defined in image planes or intermediate image planes only, when these planes are suitable for detection by electronic imaging devices.

This part of ISO 19056 defines how image brightness and uniformity are measured and how this information is provided to the user.

NOTE The scope is intentionally limited to electronic imaging devices and (intermediate) image planes. The visual observation by means of eyepieces would require a different measurement procedure and hence result in unambiguities in the description of measurement procedures. Nevertheless, this part of ISO 19056 will give useful estimates for the image brightness with visual observation as in this case, an eyepiece is used to observe an intermediate image plane (which is under the scope of this part of ISO 19056).

2 Measurands

2.1 General

For obtaining the desired image quality when using a light microscope, the brightness and uniformity of the image play an important role. This holds true for different applications and various types of instruments.

Intentionally, this part of ISO 19056 is limited to bright field microscopy and detection by means of electronic imaging devices but within this limitation, this part of ISO 19056 does apply to transmitted or reflected light microscopy and conventional or laser illumination.

2.2 Brightness

The image shall be bright enough to allow the detection of the details of the object under investigation.

NOTE This is especially true for microscopic contrasting techniques like phase contrast, differential interference contrast, dark field microscopy, or fluorescence microscopy. Although these contrasting techniques are beyond the scope of this part of ISO 19056, the given concepts can be easily extended.

As this part of ISO 19056 is based on measurement procedures, the image brightness shall be expressed in the corresponding SI units. The photometric unit Illuminance (measured in lm/m^2) shall be used whenever the illumination source covers the visible spectral range (380 nm to 780 nm) continuously, which is usually the case for filament and halogen lamps as well as for phosphor-type white LEDs. The radiometric unit irradiance (measured in W/m^2) shall be used for all other illumination sources (like discharge arc lamps, lasers, and single colour LEDs) and are more generally applicable, e.g. to the ultraviolet and infrared spectral range.