

---

---

**Tissue-engineered medical  
products — Bioactive ceramics —  
Method to measure cell migration in  
porous materials**

*Produits médicaux issus de l'ingénierie tissulaire — Céramiques  
bioactives — Méthode de mesure de la migration cellulaire dans les  
matériaux poreux*

ITeH Standards  
(<https://standards.iteh.ai>)  
Document Preview

ISO 19090:2018

<https://standards.iteh.ai/catalog/standards/iso/c186d014-0cb6-485c-8036-2109c17421ea/iso-19090-2018>



**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

ISO 19090:2018

<https://standards.iteh.ai/catalog/standards/iso/c186d014-0cb6-485c-8036-2109c17421ea/iso-19090-2018>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword.....	iv
Introduction.....	v
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>2</b>
<b>5 Test specimen</b> .....	<b>3</b>
<b>6 Procedures</b> .....	<b>3</b>
<b>7 Calculation</b> .....	<b>11</b>
7.1 Mean and standard deviation.....	11
7.2 Relative mean and standard deviation by normalization.....	12
<b>8 Test report</b> .....	<b>13</b>
<b>Annex A (informative) Results of tests for determination of test conditions</b> .....	<b>14</b>
<b>Annex B (informative) Results of international round robin test</b> .....	<b>26</b>
<b>Bibliography</b> .....	<b>33</b>

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 19090:2018](#)

<https://standards.iteh.ai/catalog/standards/iso/c186d014-0cb6-485c-8036-2109c17421ea/iso-19090-2018>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

ISO 19090 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 7, *Tissue-engineered medical products*.

ISO 19090:2018

<https://standards.iteh.ai/catalog/standards/iso/c186d014-0cb6-485c-8036-2109c17421ea/iso-19090-2018>

## Introduction

“Bioactive ceramics” are widely used in orthopaedic and dental fields due to their bioactivities and bioaffinities. Porous bioactive ceramics are designed as bone void fillers, and cell migration from tissue into their pores is an expectation for effective repair of bone defects; thus, they are one of the promising candidates for cell scaffolds for bone tissue engineering medical products.

To clarify the clinical safety and usefulness of these bioactive ceramics, physical, chemical and biological properties must be examined. In the methods used, animal tests are the ultimate and essential methods to examine biological properties of bioactive ceramics; however, numbers of both animals and animal tests must be reduced under the concept of 3R (Replacement, Reduction and Refinement)[3].

The first and most important property for porous biomaterials including bioactive ceramics is cell migration capability, because cell proliferation, differentiation, tissue formation and tissue maturation in and surroundings of porous biomaterials do not occur without cell migration.

Currently, two different cell-seeding methods are used for estimating “cell migration” property: One is dropping a cell suspension on the top surface of a porous material. This method tests the penetration ability of the “cell suspension” under gravity and estimates the number of cells that migrate into and are held within the porous material. The other method is shaking a porous material in the cell suspension. This method also tests the penetration ability of the “cell suspension” like the above method but uses shaking to drive the cells into the porous scaffolds. Both methods test the abilities of cell penetration and retention only, and do not test the intrinsic ability of the cell to migrate simulating what happens *in vivo*. Body fluid itself can sufficiently carry cells across a minor gap between the implanted material and the host bone. Accordingly, no cell migration test methods have been reported that mimic cell behaviour *in vivo*.

When porous bioceramics are implanted into bone defect, cells migrate into the pore to form new bone. In this process, migration of osteoblasts mainly plays important roles for osteoconduction. That is to say, no osteoconduction nor bone formation can occur without osteoblast migration.

Therefore, it is imperative to establish a quantifiable method to measure cell migration potential of porous bioactive ceramics in a manner similar to how cells behave *in vivo*, in order to evaluate their potential appropriately as materials for tissue-engineered medical products.

