
**Laboratory glass and plastics ware —
Principles of design and construction
of volumetric instruments**

*Matériel de laboratoire en verre ou en plastique — Principes de
conception et de construction d'instruments volumétriques*

Sample Document

get full document from standards.iteh.ai



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Unit of volume and reference temperature	1
4.1 Unit of volume	1
4.2 Reference temperature	1
5 Volumetric accuracy	1
6 Methods of calibration and use	2
7 Construction	3
7.1 Material	3
7.2 Wall thickness	3
7.3 Shape	3
7.4 Capacity	3
7.5 Stability	3
7.6 Delivery jets	3
7.7 Stoppers	4
7.8 Stopcocks or similar devices	4
8 Linear dimensions	4
9 Graduation lines	5
10 Scales	6
10.1 Spacing of graduation lines	6
10.2 Length of graduation lines (see Figure 2)	6
10.2.1 General	6
10.2.2 Graduation pattern I	6
10.2.3 Graduation pattern II	6
10.2.4 Graduation pattern III	6
10.2.5 Special cases	7
10.3 Sequence of graduation lines (see Figure 1)	7
10.4 Position of graduation lines (see Figure 2)	8
11 Figuring of graduation lines	9
12 Marking	10
13 Visibility of graduation lines, figures and inscriptions	11
Annex A (normative) Maximum permissible error in relation to the inner diameter at the meniscus	12
Bibliography	15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 48, *Laboratory equipment*.

This second edition cancels and replaces the first edition (ISO 384:1978), which has been technically revised to incorporate the following modifications.

- a) Volumetric instruments made from plastics have been added to the scope.
- b) Volumetric instruments of class AS have been added.
- c) The thickness of graduation lines has been modified.
- d) The basic principles for construction have been modified such that they comply with the product standards ISO 1042, ISO 648, ISO 835, ISO 385, ISO 4788 and ISO 4787.
- e) The relation between maximum permissible error and the inner diameter has been specified by an equation.
- f) [Annex A](#), explaining that relation, has been reworded.

Laboratory glass and plastics ware — Principles of design and construction of volumetric instruments

1 Scope

This International Standard sets out principles for the design of volumetric instruments manufactured from glass or from plastics in order to facilitate the most reliable and convenient use to the intended degree of accuracy.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 383, *Laboratory glassware — Interchangeable conical ground joints*

ISO 4787, *Laboratory glassware — Volumetric instruments — Methods for testing of capacity and for use*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99 apply.

4 Unit of volume and reference temperature

4.1 Unit of volume

The unit of volume shall be the millilitre (ml), which is equivalent to one cubic centimetre (cm³).

4.2 Reference temperature

The standard reference temperature, i.e. the temperature at which the volumetric instrument is intended to contain or deliver its volume (capacity), shall be 20 °C.

When the volumetric instrument is required for use in a country which has adopted a standard reference temperature of 27 °C, this figure shall be substituted for 20 °C.

NOTE The capacity of volumetric instruments varies with change of temperature. A volumetric instrument which was adjusted at 20 °C, but used at 27 °C or vice versa, would show an extra error of only 0,007 % if it is made of borosilicate glass having a coefficient of cubic thermal expansion of $9,9 \times 10^{-6} \text{ °C}^{-1}$ and of 0,02 % if it is made of soda-lime glass having a coefficient of cubic thermal expansion of $27 \times 10^{-6} \text{ °C}^{-1}$. These errors are smaller than the limits of error for most volumetric instruments. It follows, therefore, that the reference temperature is of minor importance in practical use when dealing with glassware. However, when performing calibrations, it is important to refer to the reference temperature, especially when considering volumetric plastic ware.

5 Volumetric accuracy

5.1 There are two classes of accuracy:

- the higher grade shall be designated “class A” or “class AS”;
- the lower grade shall be designated “class B”.

5.2 The maximum permissible error shall be specified for each type of volumetric instrument in regard to the method and purpose of use and the class of accuracy.

5.3 The numerical values of maximum permissible error for volumetric instruments for general purposes shall be preferably chosen from the series 10 – 12 – 15 – 20 – 25 – 30 – 40 – 50 – 60 – 80, or a suitable decimal multiple thereof.

NOTE This series of preferred numbers has been adopted because decimal sub-multiples of some of the unrounded numbers, for example 31,5, would appear to imply a degree of precision which is not intended and which could not be measured in practice.

5.4 The maximum permissible error specified for a series of sizes of a volumetric instrument should provide a reasonably uniform progression in relation to capacity.

5.5 The maximum permissible error permitted for class B should, in general, be approximately twice as permitted for class A or AS.

5.6 For volumetric instruments having a scale, the maximum permissible error for either class of accuracy shall not exceed the volume equivalent (see [Annex A](#)) of the smallest scale division.

5.7 The maximum permissible error MPE for class A or AS depends on the internal diameter D (in millimetres) at the related graduation line and shall not be smaller than derived by Formula (1):

$$\text{MPE} \geq \frac{\pi}{4} D^2 (0,4 + 0,01 D) \quad (1)$$

The corresponding class B limit shall be derived in accordance with [5.5](#).

NOTE The above formula applies for the most common volumetric instruments which have a circular cross-section, but may be transferred to non-circular cross-sections as well. See [Annex A](#).

5.8 In addition to [5.7](#), the maximum permissible error specified for any volumetric instrument designed to deliver shall also be not less than four times the standard deviation determined experimentally by an experienced operator from a series of at least 10 consecutive determinations of delivered capacity on the same volumetric instrument, carried out strictly in accordance with the method specified for this volumetric instrument in ISO 4787.

6 Methods of calibration and use

The method of calibration and use for each type of volumetric instrument is extensively described in ISO 4787.

The general procedure is based upon a gravimetric determination of the volume of water, either contained in or delivered by the volumetric instrument under test. This volume of water is calculated from its mass under consideration of air buoyancy and water density.

Volumetric instruments manufactured from plastics should be considered to be calibrated more often than glass instruments, because of the lower long-term stability of plastic instruments.

7 Construction

7.1 Material

Volumetric instruments shall be constructed of glass or plastic of suitable chemical and thermal properties. They shall be as free as possible from visible defects and shall be reasonably free from internal stress.

7.2 Wall thickness

The volumetric instruments shall be sufficiently robust in construction to withstand usual laboratory usage and the wall thickness shall show no gross departure from uniformity.

7.3 Shape

All volumetric instruments shall be of a shape which will facilitate the intended use, and should preferably be of circular cross-section.

7.4 Capacity

7.4.1 The numerical values of capacity of volumetric instruments for general purposes should preferably be chosen from the series 10 – 20 – 25 – 50, or a decimal multiple or sub-multiple thereof.

The capacity of volumetric instruments for special applications may have differing values; there are e.g. pipettes with capacities of 3 ml to 9 ml.

7.4.2 The numerical value of the volume equivalents of the smallest division on volumetric instruments having a scale shall be chosen from the series 1 – 2 – 5, or a decimal multiple or sub-multiple thereof.

7.4.3 In the case of a special purpose volumetric instrument which is to be graduated for direct reading of capacity when used with a specific liquid other than water, the specification should preferably indicate the corresponding capacity when used with pure water, so that the latter can be used for calibration.

7.5 Stability

Volumetric instruments provided with a flat base shall stand vertically without rocking or spinning when placed on a level surface and, unless specified otherwise, the axis of the graduated portion of the vessel should be vertical.

7.6 Delivery jets

7.6.1 Delivery jets at the lower end of volumetric instruments should be strongly constructed either with a smooth and gradual taper or a capillary end, both without sudden constriction at the orifice which could give rise to turbulent outflow.

7.6.2 The end of the jet shall be finished by one of the methods listed below in order of preference:

- a) hot finished square with the axis, slightly bevelled on the outside and polished;
- b) smoothly ground square with the axis and optionally fire-polished;
- c) cut square with the axis and polished.

A fire-polished finish of glass jets reduces the danger of chipping in use, but should not result in sudden constriction or in undue stress.