
Hypodermic needles for single use — Colour coding for identification

*Aiguilles hypodermiques non réutilisables — Code de couleurs pour
l'identification*

Sample Document

get full document from standards.iteh.ai



Reference number
ISO 6009:2016(E)

© ISO 2016

Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Terms and definitions	1
3 Colour code	1
Annex A (informative) Reference colour samples	3
Bibliography	6

Sample Document

get full document from standards.iteh.ai

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 84, *Devices for administration of medicinal products and catheters*.

This fourth edition cancels and replaces the third edition (ISO 6009:1992), which has been technically revised. It also incorporates a Technical Corrigendum ISO 6009:1992/Cor.1:2008.

The main changes to the previous edition of ISO 6009 introduced by this revision are the following:

- a) broadening of the range of needles to designated needle size down to 0,18 mm;
- b) review of the use of instrumentally determined colour zones (chromaticity and luminance index) as used in previous editions to specify opaque colours and has decided that instrumental measurement is not practicable;
- c) revision of [Annex A](#);
- d) deletion of Annex B.

Introduction

The intention of this International Standard is to specify colours to enable rapid visual identification of the outer diameter of single-use hypodermic needles. The presence of colour coding on a needle or package does not absolve the user of the responsibility to check the marked size of the needle. This revision defines, in addition, colours for more fine needles to follow the trend in the market.

The colours used to code needles may be applied in either opaque or transparent form, and the colour code is equally applicable to regular walled, thin-walled, extra-thin and ultra-thin walled needles. The nominal outer diameters of needles listed in this International Standard for which colours are given are those specified in ISO 9626.

This International Standard establishes a colour code but does not specify that needles are to be colour-coded or to what portion of the needle and/or packaging the colour is to be applied. Such requirements may be given in the relevant product standards such as ISO 7864.

The measurement of the colour zone of an opaque colour, especially of an item of the size and shape of the hub of a needle, is a complex procedure requiring apparatus and expertise that is to be found in relatively few laboratories and test houses. It may therefore be inconvenient, difficult or impossible for a manufacturer or purchaser routinely to assess compliance of a product with colour zone values. Such difficulties are compounded in the case of translucent colours, which are being used increasingly by needle manufacturers to allow air bubbles inside the hub to be seen and eliminated before injection.

As a consequence, the colours in this International Standard are only referenced by a colour reference system (RAL) or by Pantone Matching System accepting that this inevitably introduces a certain amount of subjectivity in the assessment of compliance.

Guidance on transition periods for implementing the requirements of this International Standard is given in ISO/TR 19244.

get full document from standards.iteh.ai