
**Microbiology of the food chain —
Preparation of test samples, initial
suspension and decimal dilutions for
microbiological examination —**

Part 1:

**General rules for the preparation of
the initial suspension and decimal
dilutions**

*Microbiologie de la chaîne alimentaire — Préparation des
échantillons, de la suspension mère et des dilutions décimales en vue
de l'examen microbiologique —*

*Partie 1: Règles générales pour la préparation de la suspension mère
et des dilutions décimales*



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Diluents	3
5.1 Basic materials	3
5.2 Diluents for general use	3
5.2.1 Peptone salt solution	3
5.2.2 Buffered peptone water	4
5.2.3 Double-strength buffered peptone water	4
5.3 Diluents for special purposes	4
5.4 Distribution and sterilization of the diluent	4
5.5 Performance testing for diluents	5
6 Apparatus	5
7 Sampling	6
8 Preparation of samples	6
8.1 General	6
8.2 Frozen products	7
8.2.1 General	7
8.2.2 Small samples defrosted before testing	7
8.2.3 Large pieces or blocks sampled while frozen	7
8.3 Hard and dry products	8
8.4 Dehydrated and other low-moisture products	8
8.5 Liquid and non-viscous products	8
8.6 Acidic products	8
8.7 High-fat (over 20 %) foods	9
8.8 Multi-component products	9
8.9 Packaged products	9
8.10 Surface samples (swabs and other devices)	10
9 Specific procedures	10
9.1 Test portion and initial suspension (primary dilution)	10
9.2 Duration of the procedure	11
9.3 Pooling and compositing procedures for qualitative tests	11
10 Further dilutions	11
10.1 Decimal dilution series	11
10.2 Other dilution series	12
Annex A (informative) Illustrations of pooling and compositing procedures	13
Annex B (informative) Method for sampling frozen test pieces or blocks	18
Annex C (informative) Data showing reliability of test results according to size of test portions	20
Annex D (informative) Verification protocol for pooling samples for qualitative tests	23
Bibliography	26

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 9, *Microbiology*.

This second edition cancels and replaces the first edition (ISO 6887-1:1999), which has been technically revised.

A list of parts in the ISO 6887 series can be found on the ISO website.

Introduction

Because of the large variety of food and animal feed products, this horizontal method might not be appropriate in every detail for certain products. In this case, different methods which are specific to these products can be used if absolutely necessary for justified technical reasons.

When this document is next reviewed, account will be taken of all information then available regarding the extent to which this horizontal method has been followed and the reasons for deviations from this method in the case of particular products.

The harmonization of test methods cannot be immediate and for certain groups of products, International Standards and/or national standards may already exist that do not comply with this horizontal method. It is hoped that when such standards are reviewed, they will be changed to comply with this document so that eventually, the only remaining departures from this horizontal method will be those necessary for well-established technical reasons.

This document defines the general rules for the preparation of samples, initial suspensions and subsequent dilutions for microbiological examination. The remaining parts of ISO 6887 give specific rules for the preparation of samples and initial suspensions, each covering the variety of food and feed products and environmental samples to which ISO 6887 applies.

For a number of products, it is necessary to take special precautions, especially when preparing the initial suspension, because of the physical state of the product (such as dry products, highly viscous products) or the presence of inhibitory substances (such as spices, high salt content) or the acidity, etc. These are covered in general terms in this document.

Any special diluents or practices required for particular products or microorganisms in specific standard methods take priority over the general rules listed in the ISO 6887 series. These can include the following:

- specific rehydration procedures for foods of low water activity to minimize osmotic shock;
- the use of adequate temperatures to aid suspension of cocoa, gelatine, milk powder, etc.;
- resuscitation procedures for the improved recovery of stressed microorganisms resulting from food processing and storage;
- homogenization procedures and duration specific to certain products (e.g. cereals) and/or to certain determinations (e.g. yeasts and moulds).

Sample Document

get full document from standards.iteh.ai

Microbiology of the food chain — Preparation of test samples, initial suspension and decimal dilutions for microbiological examination —

Part 1:

General rules for the preparation of the initial suspension and decimal dilutions

WARNING — The use of this document may involve hazardous materials, operations and equipment. It is the responsibility of the user of this document to establish appropriate safety and health practices and to determine the applicability of regulatory limitations before use.

1 Scope

This document defines general rules for the aerobic preparation of the initial suspension and of dilutions for microbiological examinations of products intended for human or animal consumption.

This document is applicable to the general case and other parts apply to specific groups of products as mentioned in the foreword. Some aspects might also be applicable to molecular methods where matrices can be associated with inhibition of the PCR steps and consequently affect the test result.

This document excludes preparation of samples for both enumeration and detection test methods where preparation instructions are detailed in specific International Standards.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7218, *Microbiology of food and animal feeding stuffs — General requirements and guidance for microbiological examinations*

ISO 11133, *Microbiology of food, animal feed and water — Preparation, production, storage and performance testing of culture media*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

laboratory sample

sample prepared for sending to the laboratory and intended for inspection or testing

[SOURCE: ISO 7002:1986, A.19]

3.2

composite sample

mixed sample of a number of the same items of food, animal feed, animals or environment, from which a test portion is taken for examination in the laboratory

Note 1 to entry: See illustration of a composite sample in [Annex A](#).

3.3

pooled sample

mixed sample of a number of the same items of food, animal feed, animals or environment, where the complete mixture is the test portion and is taken as a whole for examination in the laboratory

Note 1 to entry: See illustration of a pooled sample in [Annex A](#).

3.4

test sample

sample prepared from the *laboratory sample* ([3.1](#)) according to the procedure specified in the method of test and from which *test portions* ([3.5](#)) are taken

Note 1 to entry: Preparation of the laboratory sample before the test portion is taken is infrequently used in microbiological examinations.

[SOURCE: ISO 7002:1986, A.47]

3.5

test portion

measured (volume or mass) representative sample taken from the *laboratory sample* ([3.1](#)) for use in the preparation of the *initial suspension* ([3.6](#))

Note 1 to entry: Sometimes preparation of the *laboratory sample* ([3.1](#)) is required before the test portion is taken, but this is infrequently used in microbiological examinations.

3.6

initial suspension

primary dilution

suspension, solution or emulsion obtained after a weighed or measured quantity of the product under examination (or of a test sample prepared from the product) has been mixed with, normally, a nine-fold quantity of diluent, allowing large particles, if present, to settle

Note 1 to entry: Nine-fold quantities of diluent are normally used to produce a decimal dilution series, but other ratios may be required for specific purposes, such as to enumerate low numbers.

3.7

further dilution

suspension or solution obtained by mixing a measured volume of the *initial suspension* ([3.6](#)) with an *x*-fold volume of diluent and by repeating this operation with further dilutions until a dilution series, suitable for the inoculation of culture media, is obtained

Note 1 to entry: Ten-fold dilutions are normally used to produce a decimal dilution series, but other ratios may be required for specific purposes.

3.8

pooled test portions

mixture of test portions from a number of the same items of food, animal feed, animals or environment, where the complete mixture is the test portion examined

Note 1 to entry: See illustration of pooled test portions in [Annex A](#).

3.9**pooled (pre-)enriched test portions**

individually (pre-)enriched test portions from a number of the same items of food, animal feed, animals or environment, from which specified volumes are combined for further examination

Note 1 to entry: See illustration of pooled (pre-)enriched test portions in [Annex A](#).

3.10**specific standard**

International Standard or guidance document describing the examination of a specific product (or group of products) for the detection or enumeration of a specific microorganism (or group of microorganisms)

4 Principle

Preparation of the initial suspension ([3.6](#)) in such a way as to obtain as uniform a distribution as possible of the microorganisms contained in the test portion ([3.5](#)).

Preparation, if necessary, of further dilutions ([3.7](#)) in order to reduce the number of microorganisms per unit volume to allow, after incubation, observation of their growth or not (in the case of tubes or bottles) or colony counting (in the case of plates), as stated in each specific standard.

NOTE In order to restrict the range of enumeration to a given optimum interval, or if high numbers of microorganisms are foreseen, it is possible to inoculate only the necessary (decimal) dilutions needed to achieve the enumeration according to the calculations described in ISO 7218.

5 Diluents**5.1 Basic materials**

To improve the reproducibility of test results, it is recommended that either ready-made diluents or dehydrated basic components or a dehydrated complete preparation should be used. In all cases, the manufacturer's instructions shall be followed rigorously.

Chemical products shall be of recognized analytical quality and suitable for microbiological examinations.

The water used shall be distilled water or of equivalent quality (see ISO 7218 or ISO 11133).

For more detailed rules on preparation and performance testing of culture media, see ISO 11133.

5.2 Diluents for general use**5.2.1 Peptone salt solution****5.2.1.1 Composition**

Enzymatic digest of casein	1,0 g
Sodium chloride	8,5 g
Water	1 000 ml

5.2.1.2 Preparation

Dissolve the components in the water in flasks, bottles or test tubes ([6.4](#)) by heating, if necessary.

Adjust the pH if necessary so that, after sterilization, it is $7,0 \pm 0,2$ at 25 °C.