
**Cardiovascular implants and artificial
organs — Blood-gas exchangers
(oxygenators)**

*Implants cardiovasculaires et organes artificiels — Échangeurs gaz/
sang extracorporels (oxygénateurs)*

Sample Document

get full document from standards.iteh.ai



Reference number
ISO 7199:2016(E)

© ISO 2016

Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	3
4.1 Biological characteristics	3
4.1.1 Sterility and non-pyrogenicity	3
4.1.2 Biocompatibility	3
4.2 Physical characteristics	4
4.2.1 Blood pathway integrity	4
4.2.2 Heat exchanger fluid pathway integrity	4
4.2.3 Blood volumes	4
4.2.4 Connectors	4
4.3 Performance characteristics	4
4.3.1 Oxygen and carbon dioxide transfer rates	4
4.3.2 Heat exchanger performance factor	4
4.3.3 Integral arterial filtration efficiency	4
4.3.4 Integral arterial filter flow rate capacity	4
4.3.5 Integral arterial filter air handling capability	4
4.3.6 Blood cell damage	5
4.3.7 Time-dependent performance changes	5
4.3.8 Shelf life	5
5 Tests and measurements to determine compliance with this document	5
5.1 General	5
5.2 Biological characteristics	5
5.2.1 Sterility and non-pyrogenicity	5
5.2.2 Biocompatibility	5
5.3 Physical characteristics	6
5.3.1 Blood pathway integrity	6
5.3.2 Heat exchanger water pathway integrity	6
5.3.3 Blood volumes	6
5.3.4 Connectors	6
5.4 Performance characteristics	7
5.4.1 Oxygen and carbon dioxide transfer rates	7
5.4.2 Heat exchanger performance factor	8
5.4.3 Blood cell damage	8
5.4.4 Shelf life	8
5.4.5 Filtration efficiency	9
5.4.6 Integral arterial filter flow rate	9
5.4.7 Air-handling capability of integral arterial filter	9
6 Information supplied by the manufacturer	10
6.1 Information on the oxygenator	10
6.2 Information on the packaging	11
6.2.1 Unit container	11
6.2.2 Shipping container	11
6.3 Information in the accompanying documents	11
6.4 Information in the accompanying documents in a prominent form	12
7 Packaging	12
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see the following URL: <http://www.iso.org/iso/foreword.html>

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This third edition cancels and replaces the second edition (ISO 7199:2009), which has been technically revised.

It also incorporates the Amendment ISO 7199:2009/Amd.1:2012.

Introduction

This document is intended to ensure that devices designed to affect the exchange of gases in support of, or as a substitution for, the normal respiratory function of the lungs have been adequately tested for both their safety and function, and that extracorporeal device characteristics are appropriately disclosed when labelling the device.

This document therefore contains procedures to be used for evaluation of extracorporeal blood-gas exchangers (oxygenators). Type test procedures for determination of the gas transfer, blood cell damage and heat exchanger performance are described, although limits for these characteristics are not specified. Ready identification of the performance characteristics should, however, assist the user in the selection of an oxygenator that will suit the needs of the patient.

This document also includes minimum reporting requirements, which will allow the user to compare performance characteristics of oxygenators of different designs in a standard way.

This document makes reference to other International Standards in which methods for determination of characteristics common to medical devices can be found.

No provisions have been made for quantification of microbubble generation or for non-formed elements of bovine blood because there currently is no consensus regarding satisfactorily reproducible test methods.

Requirements for animal and clinical studies have not been included in this document. Such studies may be parts of a manufacturer's quality system.

This document contains only those requirements that are specific to oxygenators. Non-specific requirements are covered by references to other International Standards listed in the normative references clause. Since non-toxicity is anticipated to be the subject of a future horizontal/level 1 standard, this document does not cover non-toxicity.

get full document from standards.iteh.ai

Sample Document

get full document from standards.iteh.ai

Cardiovascular implants and artificial organs — Blood-gas exchangers (oxygenators)

1 Scope

This document specifies requirements for sterile, single-use, extracorporeal blood-gas exchangers (oxygenators) intended for supply of oxygen to, and removal of carbon dioxide from, the blood of humans.

This document also applies to heat exchangers and arterial filters that are integral parts of the oxygenator.

This document also applies to external equipment unique to the use of the oxygenator.

This document does not apply to

- implanted oxygenators,
- liquid oxygenators,
- extracorporeal circuits (blood tubing),
- separate heat exchangers,
- separate ancillary devices, and
- separate arterial line filter.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-4, *Biological evaluation of medical devices — Part 4: Selection of tests for interaction with blood*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 11135, *Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 15675, *Cardiovascular implants and artificial organs — Cardiopulmonary bypass systems — Arterial blood line filters*