
**Medical electrical equipment —
Part 2-67:
Particular requirements for basic
safety and essential performance of
oxygen-conserving equipment**

Appareils électromédicaux —

Partie 2-67: Exigences particulières pour la sécurité de base et les performances essentielles des économiseurs d'oxygène

Document Preview

[ISO 80601-2-67:2020](https://standards.iteh.ai/catalog/standards/iso/cae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020)

<https://standards.iteh.ai/catalog/standards/iso/cae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-67:2020](https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020)

<https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
Introduction.....	vi
201. 1 * Scope, object and related standards	1
201. 2 Normative references	3
201. 3 Terms and definitions.....	4
201. 4 General requirements.....	5
201. 5 General requirements for testing of <i>ME equipment</i>	7
201. 6 Classification of <i>ME equipment</i> and <i>ME systems</i>	8
201. 7 * <i>ME equipment</i> identification, marking and documents	8
201. 8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	15
201. 9 Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	15
201. 10 Protection against unwanted and excessive radiation <i>hazards</i>	16
201. 11 Protection against excessive temperatures and other <i>hazards</i>	16
201. 12 Accuracy of controls and instruments and protection against hazardous outputs.....	18
201. 13 <i>Hazardous situations</i> and fault conditions.....	21
201. 14 <i>Programmable electrical medical systems (PEMS)</i>	21
201. 15 Construction of <i>ME equipment</i>	22
201. 16 <i>ME systems</i>	22
201. 17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	22
201.101 Gas connections.....	22
201.102 Requirements for parts and <i>accessories</i>	23
201.103 Oxygen pressure regulators	25
202 Electromagnetic disturbances – Requirements and tests.....	25
202.4.3.1 * Configurations	25
206 Usability	26
Annex C (informative) Guide to marking and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>	27
Annex D (informative) Symbols on marking	32
Annex AA (informative) Particular guidance and rationale	33
Annex BB (informative) Reference to the <i>IMDRF essential principles</i> and labelling guidances.....	42
Annex CC (informative) Reference to the <i>essential principles</i>	46
Annex DD (informative) Reference to the general safety and performance requirements	49

Annex EE (informative) Terminology — Alphabetized index of defined terms	53
Bibliography	56

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-67:2020](https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020)

<https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020>

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <http://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 80601-2-67:2014), which has been technically revised.

The main changes compared to the previous edition are as follows:

- clarified the accessibility of inlet and outlet *connectors*;
- formatted to provide a unique identifier for each requirement; and
- harmonization with the 'A2 project' of the general standard.

A list of all parts in the ISO and IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Long-term oxygen therapy has been demonstrated in randomized, controlled clinical trials to prolong survival in *patients* with chronic respiratory disease and documented hypoxemia. Typical sources of therapeutic long-term oxygen therapy include gaseous oxygen from cylinders or from liquid oxygen and oxygen from an oxygen concentrator.

Most clinicians prescribe low flow oxygen therapy as continuous flow oxygen (CFO) delivery in l/min. CFO systems deliver the flow of oxygen without regard for the *patient's* breathing rate or pattern. Outside of the institutional care setting, the provision of CFO therapy is often a significant expense and can limit the mobility of a *patient* to the immediate vicinity of a stationary or fixed oxygen delivery system. To support mobility, *patients* use CFO from portable liquid or compressed oxygen systems with a limited storage capacity that can limit a *patient's* time and activities while away from a stationary oxygen supply.

Conserving equipment that delivers supplemental oxygen as a bolus conserves usage while allowing satisfactory *patient* arterial oxygen saturation (SaO_2) to be maintained during daily activities. *Conserving equipment* delivers supplemental oxygen unlike CFO in that the therapy gas flow is delivered only during the inspiratory phase of the breathing cycle, when it is most likely to reach the alveoli. During both the expiratory and pause phase of the breathing cycle, the flow of supplemental oxygen is stopped, minimizing waste. Because flow over time produces a volume, the bolus delivered by the *conserving equipment* is typically represented as a volume of gas. Therapy using *conserving equipment* versus CFO results in lower operating costs and longer ambulatory times for *patients* using the same CFO storage capacity.

Operation of *conserving equipment* from various *manufacturers* might differ in the dose delivery mechanism resulting in variations in oxygen therapy to the *patient*. The use of CFO numerical markings for dose settings on *conserving equipment* might not directly correlate with CFO settings and might lead to misinterpretation of gas delivery rates and volumes for a particular *patient*. This might result in incorrect *patient* setup and therapy delivery over all breathing rates and patterns versus CFO. Because of the differences in delivery, settings, and markings versus CFO therapy, *conserving equipment* use has requirements for *patient* titration to determine the proper setting(s) needed to provide adequate SaO_2 levels for the *patient* breathing patterns.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications and terms defined in Clause 3 of the general standard, in this particular document or as noted: italic type; and*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term.

- “clause” means one of the three numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” means that conformance with a requirement or a test is mandatory for conformance with this document;
- “should” means that conformance with a requirement or a test is recommended but is not mandatory for conformance with this document;
- “may” is used to describe a permission (e.g., permissible way to achieve conformance with a requirement or test);
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the marking and labelling requirements in this document.

Annex D contains a summary of the symbols referenced in this document.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

ITeH Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-67:2020](https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020)

<https://standards.iteh.ai/catalog/standards/iso/eae4a118-2d1b-45aa-bbf7-5e0b45b9ab0b/iso-80601-2-67-2020>