

---

---

**Medical electrical equipment —  
Part 2-85:  
Particular requirements for the basic  
safety and essential performance of  
cerebral tissue oximeter equipment**

*Appareils électromédicaux —*

*Partie 2-85: Exigences particulières pour la sécurité de base et les performances essentielles des oxymètres pour tissu cérébral*

**Document Preview**

[ISO 80601-2-85:2021](https://standards.iteh.ai/catalog/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021)

<https://standards.iteh.ai/catalog/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021>



**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[ISO 80601-2-85:2021](https://standards.iteh.ai/catalog/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021)

<https://standards.iteh.ai/catalog/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

|  |     |
|--|-----|
| Foreword.....  | vi  |
| Introduction.....  | vii |
| 201.1 Scope, object and related standards.....   | 1   |
| 201.1.1 * Scope.....   | 1   |
| 201.1.2 Object.....  | 2   |
| 201.1.3 Collateral standards.....  | 2   |
| 201.1.4 Particular standards.....  | 3   |
| 201.2 Normative references.....  | 4   |
| 201.3 Terms and definitions.....   | 5   |
| 201.4 General requirements.....  | 10  |
| 201.4.3.101 * Additional requirements for <i>essential performance</i> .....   | 11  |
| 201.4.102 Additional requirements for acceptance criteria.....   | 11  |
| 201.4.103 Additional requirements for <i>cerebral tissue oximeter equipment, parts and accessories</i> .....                               | 11  |
| 201.5 General requirements for testing of <i>ME equipment</i> .....  | 12  |
| 201.6 Classification of <i>ME equipment</i> and <i>ME systems</i> .....  | 12  |
| 201.7 <i>ME equipment</i> identification, marking and documents.....   | 12  |
| 201.7.1.101 <i>Information to be supplied by the manufacturer</i> .....  | 12  |
| 201.7.2.3 Consult <i>accompanying documents</i> .....  | 12  |
| 201.7.2.9.101 IP classification.....   | 12  |
| 201.7.2.101 Additional requirements for marking on the outside of <i>ME equipment</i> parts.....   | 13  |
| 201.7.4.3 Units of measurement.....  | 13  |
| 201.7.9.2 Instructions for use.....  | 13  |
| 201.7.9.2.1.101 Additional general requirements.....   | 13  |
| 201.7.9.2.2.101 Additional requirements for warnings and safety notices.....   | 15  |
| 201.7.9.2.9.101 Additional requirements for operating instructions.....  | 15  |
| 201.7.9.2.14.101 Additional requirements for <i>accessories, supplementary equipment, used material</i> .....                              | 15  |
| 201.7.9.3.1.101 * Additional general requirements.....   | 16  |
| 201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i> .....  | 16  |
| 201.8.3.101 Additional requirements for classification of <i>applied parts</i> .....   | 16  |
| 201.8.5.5.1.101 Defibrillation protection.....   | 16  |
| 201.8.7.4.7.101 Additional requirements for measurement of the <i>patient leakage current</i> .....  | 16  |
| 201.9 Protection against mechanical <i>hazards</i> of <i>ME equipment</i> and <i>ME systems</i> .....                                      | 17  |
| 201.10 Protection against unwanted and excessive radiation <i>hazards</i> .....  | 17  |
| 201.10.4 Lasers.....   | 17  |
| 201.11 Protection against excessive temperatures and other <i>hazards</i> .....  | 17  |
| 201.11.1.2.2 <i>Applied parts</i> not intended to supply heat to a <i>patient</i> .....  | 17  |
| 201.11.6.5.101 * Additional requirements for ingress of water or particulate matter into the <i>ME equipment</i> or <i>ME system</i> ..... | 18  |
| 201.11.6.7 <i>Sterilization</i> of <i>ME equipment</i> or <i>ME system</i> .....   | 18  |
| 201.11.8.101 Additional requirements for interruption of the power supply/ <i>supply mains</i> to <i>ME equipment</i> .....                | 18  |

|                  |   |    |
|------------------|---|----|
| 201.11.8.101.1   | <i>Technical alarm condition for power supply failure</i> .....   | 18 |
| 201.11.8.101.2   | Settings and data storage following short interruptions or automatic<br>switchover .....  | 19 |
| 201.11.8.101.3   | Operation following long interruptions .....  | 19 |
| 201.12           | Accuracy of controls and instruments and protection against hazardous outputs.....  | 19 |
| 201.12.1.101     | * <i>StO<sub>2</sub> accuracy of cerebral tissue oximeter equipment</i> .....   | 19 |
| 201.12.1.101.1   | * Specification .....   | 19 |
| 201.12.1.101.2   | * Data collection for determination of <i>StO<sub>2</sub> accuracy</i> .....  | 21 |
| 201.12.1.101.3   | * Data analysis for determination of <i>StO<sub>2</sub> accuracy</i> .....  | 22 |
| 201.12.1.101.4   | Characteristics of the study used for determination of <i>StO<sub>2</sub> accuracy</i> .....  | 23 |
| 201.12.4         | Protection against hazardous output .....   | 23 |
| 201.12.4.101     | * <i>Data update period</i> .....   | 23 |
| 201.12.4.102     | * Signal inadequacy .....   | 23 |
| 201.13           | <i>Hazardous situations and fault conditions for ME equipment</i> .....   | 24 |
| 201.13.101       | Detection of <i>probe faults and probe cable extender faults</i> .....  | 24 |
| 201.14           | <i>Programmable electrical medical systems (PEMS)</i> .....   | 24 |
| 201.15           | Construction of <i>ME equipment</i> .....   | 24 |
| 201.15.3.5.101   | * Additional requirements for rough handling.....   | 25 |
| 201.15.3.5.101.1 | * Shock and vibration (robustness).....   | 25 |
| 201.15.3.5.101.2 | * Shock and vibration for a <i>transit-operable cerebral tissue oximeter</i><br>during operation .....                                  | 26 |
| 201.15.101       | Mode of operation.....  | 27 |
| 201.16           | <i>ME systems</i> .....   | 27 |
| 201.17           | Electromagnetic compatibility of <i>ME equipment and ME systems</i> .....   | 27 |
| 201.101          | * <i>Cerebral tissue oximeter probes and probe cable extenders</i> .....  | 27 |
| 201.101.1        | General .....   | 27 |
| 201.101.2        | Labelling .....   | 28 |
| 201.102          | <i>Functional connection</i> .....  | 28 |
| 201.102.1        | General .....   | 28 |
| 201.102.2        | * Connection to an electronic health record or <i>integrated clinical environment</i> .....   | 28 |
| 201.102.3        | Connection to a <i>distributed alarm system</i> .....   | 28 |
| 202              | Electromagnetic disturbances — Requirements and tests.....  | 29 |
| 202.4.3.1        | Configurations.....   | 29 |
| 202.5.2.2.1      | Requirements applicable to all <i>ME equipment and ME systems</i> .....   | 29 |
| 202.8.1.101      | Additional general requirements .....   | 29 |
| 202.8.2          | <i>Patient physiological simulation</i> .....   | 29 |
| 206              | Usability.....  | 30 |
| 208              | General requirements, tests and guidance for alarm systems in medical electrical<br>equipment and medical electrical systems.....       | 30 |
| 208.6.1.2.101    | * Additional requirements for <i>alarm condition priority</i> .....   | 30 |
| 208.6.5.4.101    | * Additional requirements for <i>default alarm preset</i> .....   | 31 |
| 208.6.8.5.101    | Additional requirements for <i>alarm signal inactivation states, indication and</i><br>access.....                                      | 31 |
| 211              | Requirements for medical electrical equipment and medical electrical systems used<br>in the home healthcare environment .....           | 31 |
| 212              | Requirements for medical electrical equipment and medical electrical systems used<br>in the emergency medical services environment..... | 31 |

|  |            |
|--|------------|
| <b>Annex C</b> (informative) <b>Guide to marking and labelling requirements for <i>ME</i> equipment and <i>ME</i> systems</b> .....  | <b>32</b>  |
| <b>Annex D</b> (informative) <b>Symbols on marking</b> .....   | <b>36</b>  |
| <b>Annex AA</b> (informative) <b>Particular guidance and rationale</b> .....   | <b>37</b>  |
| <b>Annex BB</b> (informative) <b>Skin temperature at the <i>cerebral tissue oximeter probe</i></b> .....   | <b>48</b>  |
| <b>Annex CC</b> (informative) <b>Determination of <i>accuracy</i></b> .....  | <b>50</b>  |
| <b>Annex DD</b> (informative) <b>Characteristics of a <i>tissue haemoglobin phantom</i> for the verification of the <i>accuracy</i> of <i>cerebral tissue oximeter equipment</i></b> ..... | <b>56</b>  |
| <b>Annex EE</b> (informative) <b>Guideline for evaluating and documenting <i>StO<sub>2</sub> accuracy</i> in human subjects</b> .....  | <b>66</b>  |
| <b>Annex FF</b> (informative) <b><i>Functional testers for cerebral tissue oximeter equipment</i></b> .....  | <b>72</b>  |
| <b>Annex GG</b> (informative) <b>Concepts of <i>ME equipment</i> response time</b> .....   | <b>75</b>  |
| <b>Annex HH</b> (normative) <b>Data interface requirements</b> .....   | <b>80</b>  |
| <b>Annex II</b> (informative) <b>Comparison of methods of performance evaluation</b> .....   | <b>84</b>  |
| <b>Annex JJ</b> (informative) <b>Reference to the <i>IMDRF essential principles</i> and labelling guidances</b> .....  | <b>89</b>  |
| <b>Annex KK</b> (informative) <b>Reference to the <i>essential principles</i></b> .....  | <b>92</b>  |
| <b>Annex LL</b> (informative) <b>Reference to the general safety and performance requirements</b> .....  | <b>95</b>  |
| <b>Annex MM</b> (informative) <b>Terminology — alphabetized index of defined terms</b> .....   | <b>98</b>  |
| <b>Bibliography</b> .....  | <b>102</b> |

## Document Preview

[ISO 80601-2-85:2021](https://standards.iteh.ai/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021)

<https://standards.iteh.ai/catalog/standards/iso/0c6e9d2f-b16c-4b0e-8576-2f3be35472f2/iso-80601-2-85-2021>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee 62D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO and IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The estimation of blood oxygen saturation in the brain tissue by *cerebral tissue oximetry equipment* is increasingly used in many areas of medicine. This document covers *basic safety* and *essential performance* requirements achievable within the limits of existing technology.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committees that led to a requirement and into the *hazards* that the requirement addresses.

Annex BB is a literature review and provides recommendations relevant to determining the maximum safe temperature of the interface between a *cerebral tissue oximeter probe* and a *patient's* tissue.

Annex CC discusses both the formulae used to evaluate the *StO<sub>2</sub> accuracy* of *cerebral tissue oximeter equipment* measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on using in-vitro methods (phantoms) for *verification* of *StO<sub>2</sub> accuracy* of *cerebral tissue oximeter equipment*.

Annex EE presents a guideline for an in-vivo (human subjects) *controlled desaturation study* for the *verification* of *StO<sub>2</sub> accuracy* of *cerebral tissue oximeter equipment*.

Annex FF is a description of *functional testers* for use with *cerebral tissue oximeter equipment*.

Annex GG describes concepts of *cerebral tissue oximeter equipment* response time.

Annex HH describes data interface requirements.

Annex II is a comparison between human desaturations (in-vivo) and *tissue haemoglobin phantom* desaturations (in-vitro) for assessing *StO<sub>2</sub> accuracy*.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *Instructions, test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type; normative text of tables is also in a smaller type.

In referring to the structure of this document, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201.7 includes subclauses 201.7.1, 201.7.2) and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.