INTERNATIONAL STANDARD

ISO 80601-2-90

First edition
2021-08

Medical electrical equipment —

Part 2-90: Particular requirements for basic safety and essential performance of respiratory high-flow therapy equipment

Appareils électromédicaux —

Partie 2-90: Exigences particulières pour la sécurité de base et les performances essentielles des équipements de thérapie respiratoire à haut débit

Reference number
ISO 80601-2-90:2021(E)

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see patents.iec.ch).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment,* and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

Introduction

Respiratory high-flow therapy equipment has been used successfully for years with neonatal *patients*. In recent years there is more information about treating adults with *respiratory high-flow therapy equipment* when it is used as an intermediate therapy to improve oxygenation in adult critical care *patients*, respiratory care units and for palliative care. *High-flow therapy equipment* is also used in the treatment of chronic respiratory disease to reduce exacerbation, improve physiological outcomes and quality of life[30][43][44][47] [1]. The use of *respiratory high-flow therapy equipment* continues to increase as it is easily set up and is well tolerated by *patients*.

Since the outbreak of COVID-19 in January of 2020, its spread has been rapid and fierce. In hospitals across the world, all kinds of *respiratory high-flow therapy equipment* have been widely used. In general, there is a trend to use more non-invasive respiratory therapy. More and more new *manufacturers* of *respiratory high-flow therapy equipment* have rapidly emerged. Neither international nor national standards are available for *respiratory high-flow therapy equipment*. With the spread of the epidemic globally, the demand for this document is clear and very urgent.

The first *respiratory high-flow therapy equipment* was constructed by the connection of a *humidifier*, air/oxygen mixer/blender, flowmeter, breathing tube and cannula. Based on the improvement in technical integration in recent years, there are several technical routes for *respiratory high-flow therapy equipment* on the market. *Respiratory high-flow therapy equipment* is not fully covered by the existing standards for *humidifiers*, gas mixers for medical use, flowmeters or *ventilators*.

This document addresses the *basic safety* and *essential performance* requirements of *respiratory high-flow therapy equipment*, including *risks* related to oxygen (e.g., fires, incorrect oxygen concentration, incorrect flow delivery, etc.).

Specifically, the following *risks* and related requirements were considered in the development of this document.

— Contaminated air entering the *gas intake port* of the *respiratory high-flow therapy equipment*.

— Instability of gas supply from a *high-pressure inlet*.

— Insufficient pressure from a *high-pressure inlet*, and subsequent effects on oxygen delivered to the *patient*.

— Insufficient oxygen being delivered to the *patient*, and related *alarm condition*.

— *Usability* by *operators* wearing personal protective equipment (such as gloves and blurred visors), when setting up equipment, or viewing or changing settings.

— Instability of output delivered to *patients*, necessitating frequent *operator* adjustment.

— *Processing* of equipment, including the surface of the *enclosure* and internal *gas pathways*, particularly after use on infectious *patients*.

— Infectious exhaled gas.

— Overheating of *respiratory high-flow therapy equipment*.

In this document, the following print types are used:

— requirements and definitions: roman type;

— *test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type*;

[1] Numbers in square brackets refer to the Bibliography.

— informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term:

— “clause” means one of the five numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.);

— “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

— “shall” means that conformance with a requirement or a test is mandatory for conformance with this document;

— “should” means that conformance with a requirement or a test is recommended but is not mandatory for conformance with this document;

— “may” is used to describe permission (e.g. a permissible way to achieve conformance with a requirement or test);

— "can" is used to describe a possibility or capability;

— "must" is used express an external constraint.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.