

Redline version
compares Third edition to
Second edition



**Optics and photonics —
Environmental test methods —**

**Part 2:
Cold, heat and humidity**

*Optique et photonique — Méthodes d'essais d'environnement —
Partie 2: Froid, chaleur et humidité*

Sample Document

get full document from standards.iteh.ai



Reference number
ISO 9022-2:redline:2015(E)

© ISO 2015

IMPORTANT — PLEASE NOTE

This is a mark-up copy and uses the following colour coding:

- | | |
|---|---|
| Text example 1 | — indicates added text (in green) |
| Text example 2 | — indicates removed text (in red) |
|  | — indicates added graphic figure |
|  | — indicates removed graphic figure |
| 1.x ... | — Heading numbers containg modifications are highlighted in yellow in the Table of Contents |

DISCLAIMER

This Redline version provides you with a quick and easy way to compare the main changes between this edition of the standard and its previous edition. It doesn't capture all single changes such as punctuation but highlights the modifications providing customers with the most valuable information. Therefore it is important to note that this Redline version is not the official ISO standard and that the users must consult with the clean version of the standard, which is the official standard, for implementation purposes.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Normative reference references	1
3 General information and test conditions	1
4 Conditioning	1
4.1 General	1
4.2 Constant stress conditions	2
4.3 Cycling exposure conditions	4
5 Procedure	11
5.1 General	11
5.2 Pre-test	11
6 Environmental test code	11
7 Specification	11

Sample Document

get full document from standards.iteh.ai

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

~~International Standards are~~ The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the ~~rules given in~~ editorial rules of the ISO/IEC Directives, Part 32 (see www.iso.org/directives).

~~Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.~~

Attention is drawn to the possibility that some of the elements of this ~~part of document~~ ISO 9022 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information www.iso.org/technicalbarriers

~~International Standard~~ The committee ISO 9022-2 was prepared by Technical Committee responsible for this document is ISO/TC 172, *Optics and optical instruments* photonics, Subcommittee SC 1, *Fundamental standards*.

This ~~second~~ ^{third} edition cancels and replaces the ~~first~~ ^{second} edition (ISO 9022-2:1994/2002), Tables 2, 3 and 7 of which have been technically revised of which it constitutes a minor revision.

ISO 9022 consists of the following parts, under the general title *Optics and optical instruments* photonics — *Environmental test methods*:

- Part 1: Definitions, extent of testing
- Part 2: Cold, heat and humidity
- Part 3: Mechanical stress
- Part 4: Salt mist
- ~~Part 5: Combined cold, low air pressure~~ 6: Dust
- ~~Part 6: Dust~~
- Part 7: ~~Drip,~~ Resistance to drip or rain
- Part 8: High ^{internal} pressure, low ^{internal} pressure, immersion
- Part 9: Solar radiation and weathering
- ~~Part 10: Combined sinusoidal vibration and dry heat or cold~~

- Part 11: Mould growth
- Part 12: Contamination
- ~~— Part 13: Combined shock, bump or free fall and dry heat or cold~~
- Part 14: Dew, hoarfrost, ice
- ~~— Part 15: Combined digitally controlled broad band random vibration and dry heat or cold~~
- ~~— Part 16: Combined bounce or steady state acceleration and dry heat or cold~~
- Part 17: Combined contamination, solar radiation
- ~~— Part 18: Combined damp heat and low internal pressure~~
- ~~— Part 19: Temperature cycles combined with sinusoidal or random vibration~~
- Part 20: Humid atmosphere containing sulfur dioxide or hydrogen sulfide
- ~~Part 21: Combined low pressure and ambient temperature or dry heat~~ 22: Combined cold, dry heat or temperature change with bump or random vibration
- Part 23: Low pressure combined with cold, ambient temperature and dry and damp heat

Sample Document

get full document from standards.iteh.ai

Introduction

Optical instruments are affected during their use by a number of different environmental parameters which they are required to resist without significant reduction in performance and to remain within defined specifications.

The type and severity of these parameters depend on the conditions of use of the instrument (for example, in the laboratory or workshop) and on its geographical location. The environmental effects on optical instrument performance in the tropics and subtropics are totally different from those found when they are used in the arctic regions. Individual parameters cause a variety of different and overlapping effects on instrument performance.

The manufacturer attempts to ensure, and the user naturally expects, that instruments will resist the likely rigours of their environment throughout their life. This expectation can be assessed by exposure of the instrument to a range of simulated environmental parameters under controlled laboratory conditions. The severity of these conditions is often increased to obtain meaningful results in a relatively short period of time.

In order to allow assessment and comparison of the response of optical instruments to appropriate environmental conditions, ISO 9022 contains details of a number of laboratory tests which reliably simulate a variety of different environments. The tests are based largely on IEC standards, modified where necessary to take into account features special to optical instruments.

~~It should be noted that, as~~ As a result of continuous progress in all fields, optical instruments are no longer only precision-engineered optical products, but, depending on their range of application, also contain additional assemblies from other fields. For this reason, the principal function of the instrument ~~must~~ is to be assessed to determine which International Standard should be used for testing. If the optical function is of primary importance, then ISO 9022 is applicable, but if other functions take precedence then the appropriate International Standard in the field concerned should be applied. Cases may arise where application of both ISO 9022 and other appropriate International Standards will be necessary.

Optics and photonics — Environmental test methods —

Part 2: Cold, heat and humidity

1 Scope

This part of ISO 9022 specifies ~~methods for the testing~~ the methods relating to the environmental tests of optical instruments ~~and instruments containing optical components~~ including additional assemblies from other fields (e.g. mechanical, chemical, and electronic devices), under equivalent conditions, for their ability to resist the influence of temperature and ~~air~~/or humidity.

The purpose of the testing is to investigate to what extent the optical, ~~thermal~~climatic, mechanical, chemical, and electrical (including electrostatic) performance characteristics of the specimen are affected by temperature and/or humidity.

2 Normative ~~reference~~references

The following ~~normative document contains provisions which, through reference in this text, constitute provisions of this part of documents, in whole or in part, are normatively referenced in this document and are indispensable for its application~~ISO 9022. For dated references, ~~subsequent amendments to, or revisions of, this publication do not apply. However, parties to agreements based on this part of only ISO 9022 are encouraged to investigate the possibility of applying the most recent edition of the normative document indicated below~~The edition cited applies. For undated references, the latest edition of the ~~normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.~~referenced document (including any amendments) applies.

ISO 9022-1:1994, *Optics and ~~optical instruments~~photonics — Environmental test methods — Part 1: Definitions, extent of testing*

3 General information and test conditions

The specimen is exposed to climatic stress in conditioning chambers or cabinets providing air circulation.

The size of the specimens and their arrangement shall be such as to ensure a uniform conditioning of all specimens. Where ~~moisture~~condensation is likely to occur, the condensate shall be prevented from dripping onto the specimen.

Where heat-dissipating specimens are involved, the relevant specification shall state the number, method of installation and location of the heat sensors.

WARNING — Persons entering test chambers with dew-point temperatures of equal to or greater than 38 °C must be equipped with breathing apparatus (e.g. conditioning methods 12, 13, and 16).

4 Conditioning

4.1 General

The required exposure time shall not commence until all parts of the specimen have reached a temperature within at least 3 K of the test chamber temperature. For heat-dissipating specimens, the period of exposure or dwell time (conditioning methods 14 and 15) shall not begin, or end, until the temperature of the specimen changes not more than 1 K within 1 h at the stabilized test chamber