
**Terminal units for medical gas
pipeline systems —**

**Part 1:
Terminal units for use with
compressed medical gases and vacuum**

Prises murales pour systèmes de distribution de gaz médicaux —

Partie 1: Prises murales pour les gaz médicaux comprimés et le vide

Sample Document

get full document from standards.iteh.ai



Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 *Normative references	2
3 Terms and definitions	2
4 General requirements	7
4.1 Safety	7
4.2 *Alternative construction	7
4.3 Materials	7
5 Design requirements	8
5.1 Medical gas supply	8
5.2 Terminal units for different pressures	9
5.3 Retention of gas specificity	9
5.4 Gas-specific connection point	9
5.5 Terminal unit check valve	9
5.6 Terminal unit maintenance valve	9
5.7 Connection of terminal units to the pipeline (see also 9.2)	10
5.8 Socket	10
5.9 Compliance	10
5.10 Endurance (connection/release)	10
5.10.1 Socket	10
5.10.2 Probe	10
5.11 *Pressure drop	10
5.12 Connection force and torque	11
5.13 Disconnection force and torque	11
5.14 Mechanical strength	11
5.15 Leakage	12
5.16 Gas specificity	12
5.17 Effective connection of probes	12
5.18 Electrical requirements	12
6 Constructional requirements	12
6.1 Cleaning	12
6.2 Lubricants	12
7 Test methods	12
7.1 General	12
7.2 Test for endurance	13
7.3 Test for pressure drop	13
7.4 Test for connection force and torque	15
7.5 Test for disconnection force and torque	15
7.6 Test for mechanical strength	16
7.7 Test for leakage	16
7.8 Test for gas specificity	17
7.9 Test for effective connection of probe	17
7.10 Test for durability of markings and colour coding	17
8 Marking, colour coding and packaging	17
8.1 Marking	17
8.2 Colour coding	17
8.3 Packaging	18
9 Information to be supplied by the manufacturer	18
9.1 Technical description	18
9.2 Instructions	18

Annex A (informative) Rationale	19
Annex B (informative) Environmental aspects	20
Annex C (informative) Special national and regional conditions for electrical installations	22
Bibliography	24

Sample Document

get full document from standards.iteh.ai

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC6, *Medical gas systems*.

This third edition cancels and replaces the second edition (ISO 9170-1:2008), which has been technically revised.

This edition includes the following significant changes with respect to the previous edition:

- a) oxygen 93, detailing marking and colour coding, was introduced;
- b) figures for test conditions were clarified.

A list of all parts in the ISO 9170 series can be found on the ISO website.

Introduction

Terminal units are the points on a medical gas pipeline system where the operator makes connections and disconnections for the supply of specified medical gases to anaesthetic machines, lung ventilators or other items of medical equipment. Terminal units are also used for vacuum pipeline systems. A wrong connection can create a hazard to the patient or operator. It is important that terminal units and their components be designed, manufactured, installed and maintained in such a way as to meet the requirements specified in this document.

This document pays particular attention to

- suitability of materials,
- gas-specificity,
- cleanliness,
- testing,
- identification, and
- information supplied.

This document contains information for the installation and testing of terminal units prior to use. Testing of terminal units prior to use is critical to patient safety, and it is essential that terminal units are not used until full testing in accordance with ISO 7396-1 has been completed.

[Annex A](#) contains rationale statements for some of the requirements of this document. The clauses and subclauses marked with an asterisk (*) after their number have corresponding rationale contained in [Annex A](#), included to provide additional insight into the reasoning that led to the requirements and recommendations that have been incorporated in this document. It is considered that knowledge of the reasons for the requirements will not only facilitate the proper application of this document, but will also expedite any subsequent revisions.

[Annex B](#) contains environmental aspects that should be considered.

Terminal units for medical gas pipeline systems —

Part 1:

Terminal units for use with compressed medical gases and vacuum

1 Scope

This document is intended especially to ensure the gas-specific assembly, mechanical resistance, flow, leakage and pressure drop of terminal units and to prevent their interchange between different gases and services and applies to terminal units:

- a) intended for use in medical gas pipeline systems in accordance with ISO 7396-1;
- b) used as pressure outlets on pressure regulators in accordance with ISO 10524-1;
- c) used as pressure outlets on pressure regulators integrated with cylinder valves (VIPR) in accordance with ISO 10524-3.

This document applies to terminal units for use with the following gases for administration to patients or for medical uses (A):

- oxygen (A);
- nitrous oxide (A);
- medical air (A);
- carbon dioxide (A);
- oxygen/nitrous oxide mixture (A);
- helium/oxygen mixtures (A);
- oxygen 93 (A);
- gases and gas mixtures classified as medical device (A);
- gases delivered to medical devices or intended for medical purposes (A);
- gases and gas mixtures for medicinal use not specified above (A).

This document applies to terminal units for use with the following gases (B):

- air for driving surgical tools (B);
- nitrogen for driving surgical tools (B).

This document applies to terminal units for use with vacuum systems (C).

NOTE The requirements of this document can be used as guidelines for terminal units for other gases. These other gases will be considered for inclusion in this document when they come into general use.

This document specifies requirements for terminal units for supply and disposal of nitrogen and air for driving surgical tools.

This document specifies requirements for probes intended to be connected to the gas-specific connection point.

This document does not specify the dimensions of probes or of the gas-specific connection points.

NOTE Regional or national standards specifying dimensions of probes and gas-specific connection points are given in the Bibliography.

Other connection systems in national use may be acceptable under this document. Dimensioning for such connections will be specified by their respective national standards.

This document does not specify the requirements for terminal units for anaesthetic gas scavenging systems (AGSS), which are specified in ISO 9170-2.

2 *Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 32, *Gas cylinders for medical use — Marking for identification of content*

ISO 5359:2014, *Low-pressure hose assemblies for use with medical gases*

ISO 6506-1, *Metallic materials — Brinell hardness test — Part 1: Test method*

ISO 7396-1:2016, *Medical gas pipeline systems — Part 1: Pipeline systems for compressed medical gases and vacuum*

ISO 11114-3, *Transportable gas cylinders — Compatibility of cylinder and valve materials with gas contents — Part 3: Autogenous ignition test in oxygen atmosphere*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15001:2010, *Anaesthetic and respiratory equipment — Compatibility with oxygen*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE A diagram of a typical terminal unit and probe, with an example of terminology, is shown in [Figure 1](#).

3.1 diameter-index safety system connector DISS connector

any of a range of male and female components intended to maintain gas specificity by allocation of a set of different diameters to the mating connectors for each particular gas

3.2 gas specific

having characteristics which prevent connections between different gas services or vacuum services

3.3 gas-specific connection point

part of the socket which is the receptor for a gas-specific probe

3.4**gas-specific connector**

connector with dimensional characteristics that prevent connections between different gas services

Note 1 to entry: Examples of gas-specific connectors are *quick connectors* (3.14), screw-threaded connectors, *diameter-index safety system (DISS) connectors* (3.1), *non-interchangeable screw-threaded (NIST) connectors* (3.11) or *sleeve indexed (SIS) connectors* (3.16).

3.5**low-pressure hose assembly**

assembly consisting of a flexible hose with permanently attached gas-specific inlet and outlet connectors and designed to conduct a *medical gas* (3.7) at pressures less than 1 400 kPa and vacuum

3.6**medical device gas**

any gas or mixture of gases intended by the manufacturer to be used for human beings for the purpose of

- diagnosis, prevention, monitoring, treatment or alleviation of disease,
- investigation, replacement or modification of the anatomy or of a physiological process,
- control of conception,

and does not achieve its principal intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its function by such means

Note 1 to entry: In Europe, these gases are classified as a medical device in accordance with Directive 93/42/EC.

3.7**medical gas**

any gas or mixture of gases having properties for treating or preventing disease in human beings which may be used in or administered either with a view to restore, correct or modify physiological functions by exerting a pharmacological, immunological or metabolic action, or to make a medical diagnosis

Note 1 to entry: This is also sometimes referred to as medicinal gas.

Note 2 to entry: In Europe, this is classified as a medicinal product in accordance with Directive 2001/83/EC.

3.8**medical gas pipeline system**

complete system which comprises a supply system, a monitoring and alarm system and a distribution system with terminal units at the points where *medical gases* (3.7) or vacuum may be required

3.9**medical gas supply system**

either

- a) a medical gas pipeline system, or
- b) an installation having no permanent pipeline system but employing a *medical gas* (3.7) supply source complete with pressure regulator(s)

3.10**nominal distribution pressure**

pressure which the *medical gas pipeline system* (3.8) is intended to deliver at the terminal units

Note 1 to entry: Unless otherwise specified, pressures in this document are expressed as gauge pressures (i.e. atmospheric pressure is defined as 0 kPa gauge pressure).