

ISO/FDIS 11929-2:2025(en)

ISO/TC 85/SC 2/WG 17

Secretariat: AFNOR

Date: 2025-06-18xx

Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and applications — ~~Part 2: Advanced applications~~ application —

**Part 2:**  
**Advanced applications**

*Détermination des limites caractéristiques (seuil de décision, limite de détection et extrémités de l'intervalle élargi) pour mesurages de rayonnements ionisants — Principes fondamentaux et applications —*

*Partie 2: Applications avancées*

ISO/FDIS 11929-2

<https://standards.iteh.ai/catalog/standards/iso/e6f623bf-730e-4e7a-b574-6d97c208d4c1/iso-fdis-11929-2>

FDIS stage

**ISO/FDIS 11929-2(E:2025(en))**

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
~~Email~~**E-mail:** [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

ISO/FDIS 11929-2

<https://standards.iteh.ai/catalog/standards/iso/e6f623bf-730e-4e7a-b574-6d97c208d4c1/iso-fdis-11929-2>

## Contents—Page

Foreword .....	iv
Introduction .....	vi
1 Scope .....	1
2 Normative references .....	2
3 Terms and definitions .....	2
4 Quantities and symbols .....	6
5 Summary of procedures for evaluating and reporting uncertainty and characteristic limits .....	8
6 Evaluation of a measurement on the basis of ISO/IEC Guide 98-3:2008/Suppl 1:2008 .....	11
6.1 Introduction and decisions to be made .....	11
6.2 General aspects concerning the measurand and the model of evaluation .....	11
6.3 Establishing probability distributions for the input quantities .....	12
6.4 Propagating probability distributions .....	14
6.5 Evaluation of the primary measurement result .....	15
6.6 Standard uncertainty associated with the primary measurement result .....	15
7 PDF for an assumed true value of the measurand .....	16
8 Decision threshold, detection limit and assessments .....	16
8.1 Specifications .....	16
8.2 Decision threshold .....	17
8.3 Detection limit .....	17
8.4 Assessments .....	18
9 Limits of the coverage interval .....	18
9.1 General Aspects .....	18
9.2 The probabilistically symmetric coverage interval .....	19
9.3 The shortest coverage interval .....	19
10 The best estimate and its associated standard uncertainty .....	20
11 Documentation .....	21
Annex A (normative) Measurements with low count numbers .....	22
Annex B (informative) Explanatory notes .....	24
Bibliography .....	38

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO ~~had~~ had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, Nuclear energy, nuclear technologies, and radiological protection, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement)*.

This third edition of ISO 11929-2 replaces ISO 11929-2:2019, of which it constitutes a minor revision.

The main changes are as follows:

- ~~correction of the internal references within the text;~~
- ~~3.12 and 3.13:~~ correction of the definitions of decision threshold (3.12) and the detection limit (3.13);
- ~~Clause 7:~~ editorial correction in 7;
- ~~B.2.1:~~ correction of 0 Formula (B.3);
- ~~B.3.1:~~ correction of 0 Formula (B.17);
- ~~B.2.2:~~ correction of 0 Formula (B.27);
- ~~B.4:~~ editorial correction in B.4.

A list of all the parts in the ISO 11929 series can be found on the ISO website.

## ISO/FDIS 11929-2:2025(en)

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# iTeh Standards (<https://standards.iteh.ai>) Document Preview

## ISO/FDIS 11929-2

<https://standards.iteh.ai/catalog/standards/iso/e6f623bf-730e-4e7a-b574-6d97c208d4c1/iso-fdis-11929-2>

## Introduction

Measurement uncertainties and characteristic values, such as the decision threshold, the detection limit and limits of the coverage interval for measurements as well as the best estimate and its associated standard measurement uncertainty, are of importance in metrology in general, and for radiological protection in particular. The quantification of the uncertainty associated with a measurement result provides a basis for the trust an individual can have in a measurement result. Conformity with regulatory limits, constraints or reference values can only be demonstrated by taking into account and quantifying all sources of uncertainty. Characteristic limits provide – in the end – the basis for deciding under uncertainty.

The ISO 11929 series provides characteristic values of a non-negative measurand of ionizing radiation. It is also applicable for a wide range of measuring methods extending beyond measurements of ionizing radiation.

The limits to be provided according to the ISO 11929 series for specified probabilities of wrong decisions allow detection possibilities to be assessed for a measurand and for the physical effect quantified by this measurand as follows:

- the “decision threshold” allows a decision to be made on whether or not the physical effect quantified by the measurand is present;
- the “detection limit” indicates the smallest true quantity value of the measurand that can still be detected with the applied measurement procedure; this gives a decision on whether or not the measurement procedure satisfies the requirements and is therefore suitable for the intended measurement purpose;
- the “limits of the coverage interval” enclose, in the case of the physical effect recognized as present, a coverage interval containing the true quantity value of the measurand with a specified probability.

Hereinafter, the limits mentioned are jointly called the “characteristic limits”.

**NOTE** According to ISO/IEC Guide 99:2007 updated by JCGM 200:2012 the term “coverage interval” is used here instead of “confidence interval” in order to distinguish the wording of Bayesian terminology from that of conventional statistics.

All the characteristic values are based on Bayesian statistics and on the ISO/IEC 98-3 Guide to the Expression of Uncertainty in Measurement as well as on the ISO/IEC Guide 98-3:2008/Suppl 1:2008 and ISO/IEC Guide 98-3:2008/Suppl 2:2011. As explained in detail in ISO 11929-2, the characteristic values are mathematically defined by means of moments and quantiles of probability distributions of the possible measurand values.

Since measurement uncertainty plays an important part in the ISO 11929 series, the evaluation of measurements and the treatment of measurement uncertainties are carried out by means of the general procedures according to the ISO/IEC Guide 98-3 and to the ISO/IEC Guide 98-3:2008/Suppl 1:2008; see also References [0] to [13]. This enables the strict separation of the evaluation of the measurements, on the one hand, and the provision and calculation of the characteristic values, on the other hand. The ISO 11929 series makes use of a theory of uncertainty in measurement<sup>[14 to 16]</sup> based on Bayesian statistics (e.g. References [0] to [22]) in order to allow to take into account also those uncertainties that cannot be derived from repeated or counting measurements. The latter uncertainties cannot be handled by frequentist statistics.

Because of developments in metrology concerning measurement uncertainty laid down in the ISO/IEC Guide 98-3, ISO 11929:2010 was drawn up on the basis of the ISO/IEC Guide 98-3, but using Bayesian statistics and the Bayesian theory of measurement uncertainty. This theory provides a Bayesian foundation for the ISO/IEC Guide 98-3. Moreover, ISO 11929:2010 was based on the definitions of the characteristic values [0] the standard proposal [0] and the introducing article [0]. It unified and replaced all earlier parts of ISO 11929 and was applicable not only to a large variety of particular measurements of ionizing radiation but also, in analogy, to other measurement procedures.