



FINAL DRAFT International Standard

ISO/FDIS 11929-3

Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application —

Part 3: Applications to unfolding methods

Détermination des limites caractéristiques (seuil de décision, limite de détection et extrémités de l'intervalle élargi) pour mesurages de rayonnements ionisants — Principes fondamentaux et applications —

Partie 3: Applications aux méthodes de déconvolution

ISO/TC 85/SC 2

Secretariat: **AFNOR**

Voting begins on:

Voting terminates on:

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO/FDIS 11929-3

<https://standards.iteh.ai/catalog/standards/iso/f61c95fe-09bf-4c7e-b7f0-b53916c72788/iso-fdis-11929-3>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Quantities and symbols	6
5 Evaluation of a measurement using unfolding methods	8
5.1 General aspects	8
5.2 Models of unfolding and general uncertainty evaluation	8
5.3 Unfolding as a sub-model	10
5.4 Input quantities and their uncertainties	10
5.5 Parameters of unfolding	11
5.6 Procedure for unfolding	12
5.7 Modification for Poisson distributed count numbers for unfolding	14
5.8 Evaluation of the primary results and their associated standard uncertainties	15
5.9 Standard uncertainty as a function of an assumed true value of the measurand	16
5.10 Decision threshold, detection limit and assessments	17
5.10.1 Specifications	17
5.10.2 Decision threshold	17
5.10.3 Detection limit	18
5.10.4 Assessments	18
5.11 Coverage interval and the best estimate and its associated standard uncertainty	18
5.11.1 General aspects	18
5.11.2 The probabilistically symmetric coverage interval	19
5.11.3 The shortest coverage interval	19
5.12 Documentation	19
Annex A (informative) Correlations and covariances	21
Annex B (informative) Spectrum unfolding in nuclear spectrometric measurement	24
Bibliography	35

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, *Nuclear energy, nuclear technologies, and radiological protection*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition of ISO 11929-3 replaces ISO 11929-3:2019, of which it constitutes a minor revision.

The main changes are as follows:

- correction of the internal references within the text;
- correction of the definitions of decision threshold (3.12) and the detection limit (3.13);
- correction of [Clause 4](#) according to comments;
- correction of [Formulae \(B.2\)](#) and [\(B.4\)](#);
- correction of the 8th paragraph in [B.3.2](#);
- correction of the 1st paragraph [B.3.5](#);
- correction of the 7th paragraph [B.5.2](#).

A list of all the parts in the ISO 11929 series can be found on the ISO website.

Introduction

Measurement uncertainties and characteristic values, such as the decision threshold, the detection limit and limits of the coverage interval for measurements as well as the best estimate and its associated standard measurement uncertainty, are of importance in metrology in general and for radiological protection in particular. The quantification of the uncertainty associated with a measurement result provides a basis for the trust an individual can have in a measurement result. Conformity with regulatory limits, constraints or reference values can only be demonstrated by taking into account and quantifying all sources of uncertainty. Characteristic limits provide, at the end, the basis for deciding under uncertainty.

This standard provides characteristic values of a non-negative measurand of ionizing radiation. It is also applicable for a wide range of measuring methods extending beyond measurements of ionizing radiation.

The limits to be provided according to the ISO 11929 series for specified probabilities of wrong decisions allow detection possibilities to be assessed for a measurand and for the physical effect quantified by this measurand as follows:

- the “decision threshold” allows a decision to be made on whether or not the physical effect quantified by the measurand is present;
- the “detection limit” indicates the smallest true quantity value of the measurand that can still be detected with the applied measurement procedure; this gives a decision on whether or not the measurement procedure satisfies the requirements and is therefore suitable for the intended measurement purpose;
- the “limits of the coverage interval” enclose, in the case of the physical effect recognized as present, a coverage interval containing the true quantity value of the measurand with a specified probability.

Hereinafter, the limits mentioned are jointly called the “characteristic limits”.

NOTE According to ISO/IEC Guide 99:2007 updated by JCGM 200:2012 the term “coverage interval” is used here instead of “confidence interval” in order to distinguish the wording of Bayesian terminology from that of conventional statistics.

All the characteristic values are based on Bayesian statistics and on the ISO/IEC 98-3 Guide to the Expression of Uncertainty in Measurement as well as on the ISO/IEC Guide 98-3:2008/Suppl 1:2008 and ISO/IEC Guide 98-3:2008/Suppl 2:2011. As explained in detail in ISO 11929-2, the characteristic values are mathematically defined by means of moments and quantiles of probability distributions of the possible measurand values.

Since measurement uncertainty plays an important part in ISO 11929, the evaluation of measurements and the treatment of measurement uncertainties are carried out by means of the general procedures according to the ISO/IEC Guide 98-3 and to the ISO/IEC Guide 98-3:2008/Suppl 1:2008; see also References [9] to [13]. This enables the strict separation of the evaluation of the measurements, on the one hand, and the provision and calculation of the characteristic values, on the other hand. The ISO 11929 series makes use of a theory of uncertainty in measurement^{[14] to [16]} based on Bayesian statistics (e.g. References [17] to [22]) in order to allow to take into account also those uncertainties that cannot be derived from repeated or counting measurements. The latter uncertainties cannot be handled by frequentist statistics.

Because of developments in metrology concerning measurement uncertainty laid down in the ISO/IEC Guide 98-3, ISO 11929:2010 was drawn up on the basis of the ISO/IEC Guide 98-3, but using Bayesian statistics and the Bayesian theory of measurement uncertainty. This theory provides a Bayesian foundation for the ISO/IEC Guide 98-3. Moreover, ISO 11929:2010 was based on the definitions of the characteristic values^[9], the standard proposal^[10], and the introducing article^[11]. It unified and replaced all earlier parts of ISO 11929 and was applicable not only to a large variety of particular measurements of ionizing radiation but also, in analogy, to other measurement procedures.

Since the ISO/IEC Guide 98-3:2008/Suppl 1:2008 has been published, dealing comprehensively with a more general treatment of measurement uncertainty using the Monte Carlo method in complex measurement evaluations. This provided an incentive for writing a corresponding Monte Carlo supplement^[12] to ISO 11929:2010 and to revise ISO 11929:2010. The revised ISO 11929 is also essentially founded on Bayesian statistics and can serve as a bridge between ISO 11929:2010 and the ISO/IEC Guide 98-3:2008/

ISO/FDIS 11929-3:2025(en)

Suppl 1:2008. Moreover, more general definitions of the characteristic values (ISO 11929-2) and the Monte Carlo computation of the characteristic values make it possible to go a step beyond the present state of standardization laid down in ISO 11929:2010 since probability distributions rather than uncertainties can be propagated. It is thus more comprehensive and extending the range of applications.

The revised ISO 11929, moreover, is more explicit on the calculation of the characteristic values. It corrects also a problem in ISO 11929:2010 regarding uncertain quantities and influences, which do not behave randomly in measurements repeated several times. Reference [13] gives a survey on the basis of the revision. Furthermore, this document gives detailed advice how to calculate characteristic values in the case of multivariate measurements using unfolding methods. For such measurements, the ISO/IEC Guide 98-3:2008/Suppl 2:2011 provides the basis of the uncertainty evaluation.

Formulas are provided for the calculation of the characteristic values of an ionizing radiation measurand via the “standard measurement uncertainty” of the measurand (hereinafter the “standard uncertainty”) derived according to the ISO/IEC Guide 98-3 as well as via probability density functions (PDFs) of the measurand derived in accordance with ISO/IEC Guide 98-3:2008/Suppl 1:2008. The standard uncertainties or probability density functions take into account the uncertainties of the actual measurement as well as those of sample treatment, calibration of the measuring system and other influences. The latter uncertainties are assumed to be known from previous investigations.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

ISO/FDIS 11929-3

<https://standards.iteh.ai/catalog/standards/iso/f61c95fe-09bf-4c7e-b7f0-b53916c72788/iso-fdis-11929-3>