
Ugotavljanje karakterističnih mej (odločitveni prag, zaznavanje meje in omejitev intervala pokritja) pri meritvah ionizirnega sevanja - Osnove in uporaba - 1. del: Osnovne aplikacije (ISO/FDIS 11929-1:2025)

Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation - Fundamentals and application - Part 1: Elementary applications (ISO/FDIS 11929-1:2025)

Bestimmung der charakteristischen Grenzen (Erkennungsgrenze, Nachweisgrenze und Grenzen des Überdeckungsintervalls) bei Messungen ionisierender Strahlung - Grundlagen und Anwendungen - Teil 1: Elementare Anwendungen (ISO/FDIS 11929-1:2025)

Détermination des limites caractéristiques (seuil de décision, limite de détection et extrémités de l'intervalle élargi) pour mesurages de rayonnements ionisants - Principes fondamentaux et applications - Partie 1: Applications élémentaires (ISO/FDIS 11929-1:2025)

Ta slovenski standard je istoveten z: prEN ISO 11929-1

ICS:

17.240	Merjenje sevanja	Radiation measurements
--------	------------------	------------------------

oSIST prEN ISO 11929-1:2025

en,fr,de



FINAL DRAFT

International Standard

ISO/FDIS 11929-1

Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application —

Part 1: Elementary applications

Détermination des limites caractéristiques (seuil de décision, limite de détection et extrémités de l'intervalle élargi) pour mesurages de rayonnements ionisants — Principes fondamentaux et applications —

Partie 1: Applications élémentaires

ISO/TC 85/SC 2

Secretariat: **AFNOR**

Voting begins on:

Voting terminates on:

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

ISO/FDIS 11929-1:2025(en)

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[oSIST prEN ISO 11929-1:2025](https://standards.iteh.ai/catalog/standards/sist/dfc056a3-cce7-457a-8056-8c1b368d52b3/osist-pren-iso-11929-1-2025)

<https://standards.iteh.ai/catalog/standards/sist/dfc056a3-cce7-457a-8056-8c1b368d52b3/osist-pren-iso-11929-1-2025>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO/FDIS 11929-1:2025(en)

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Quantities and symbols	6
5 Summary of procedures for evaluating a measurement and calculating the characteristic limits	8
5.1 General aspects	8
5.2 Modelling the measurement	9
5.3 General considerations about the applicability of the ISO/IEC Guide 98-3	9
5.4 Evaluating the input quantities, standard uncertainties and covariances, and the primary result and its associated standard uncertainty	10
5.5 Evaluating the standard uncertainty as a function of an assumed true value of the measurand	10
5.6 Calculating the decision threshold and decisions to be made	11
5.7 Calculating the detection limit and assessment of the measurement procedure	12
5.8 Calculating a coverage interval for the measurand	12
5.9 Calculating the best estimate of the measurand and its associated standard uncertainty	13
5.10 Reporting the results	14
6 Fundamentals and evaluation of the measurement	16
6.1 General aspects concerning the measurand	16
6.2 Model of evaluation	16
6.2.1 General model	16
6.2.2 Model in ionizing radiation counting measurements	17
6.3 Evaluation of the primary measurement result	18
6.4 Standard uncertainty associated with the primary measurement result	18
7 Standard uncertainty as a function of an assumed value of the measurand	19
8 Decision threshold, detection limit and assessments	21
8.1 Specifications	21
8.2 Decision threshold	21
8.3 Detection limit	21
8.4 Assessments	22
9 Limits of the coverage interval	23
9.1 General aspects	23
9.2 The probabilistically symmetric coverage interval	23
9.3 The shortest coverage interval	24
10 The best estimate and its associated standard uncertainty	24
11 Documentation	25
Annex A (normative) Repeated counting measurements with random influences	27
Annex B (normative) Measurements with ratemeters	32
Annex C (informative) Calculation of the detection limit by iteration	34
Annex D (informative) Calculations with a generalized net count rate	36
Annex E (informative) Distribution function of the standardized normal distribution	39
Bibliography	44

ISO/FDIS 11929-1:2025(en)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, *Nuclear energy, nuclear technologies, and radiological protection*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11929-1:2019), of which it constitutes a minor revision.

The main changes are as follows:

- correction of the internal references within the text;
- correction of [Formulae \(3\)](#), [\(A.5\)](#) and [\(B.3\)](#);
- correction of the definitions of decision threshold [\(3.12\)](#) and the detection limit [\(3.13\)](#);
- introduction of NOTE 3 in [5.8](#), for clarity;
- complement of the NOTE in [9.1](#);
- correction of the 1st paragraph and NOTE 2 in [Clause 10](#);
- correction of the 8th paragraph in [Annex B](#).

A list of all the parts in the ISO 11929 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO/FDIS 11929-1:2025(en)

Introduction

Measurement uncertainties and characteristic values, such as the decision threshold, the detection limit and limits of the coverage interval for measurements, as well as the best estimate and its associated standard measurement uncertainty, are of importance in metrology in general and for radiological protection in particular. The quantification of the uncertainty associated with a measurement result provides a basis for the trust an individual can have in a measurement result. Conformity with regulatory limits, constraints or reference values can only be demonstrated by taking into account and quantifying all sources of uncertainty. Characteristic limits provide, at the end, the basis for deciding if uncertainties have to be taken into account.

This standard provides characteristic values of a non-negative measurand of ionizing radiation. It is also applicable for a wide range of measuring methods extending beyond measurements of ionizing radiation.

The limits to be provided according to ISO 11929 series for specified probabilities of wrong decisions allow detection possibilities to be assessed for a measurand and for the physical effect quantified by this measurand as follows:

- the “decision threshold” allows a decision to be made on whether or not the physical effect quantified by the measurand is present;
- the “detection limit” indicates the smallest true quantity value of the measurand that can still be detected with the applied measurement procedure; this gives a decision on whether or not the measurement procedure satisfies the requirements and is therefore suitable for the intended measurement purpose;
- the “limits of the coverage interval” enclose, in the case of the physical effect recognized as present, a coverage interval containing the true quantity value of the measurand with a specified probability.

Hereinafter, the limits mentioned are jointly called the “characteristic limits”.

NOTE According to ISO/IEC Guide 99:2007 updated by JCGM 200:2012 the term “coverage interval” is used here instead of “confidence interval” in order to distinguish the wording of Bayesian terminology from that of conventional statistics.

All the characteristic values are based on Bayesian statistics and on the ISO/IEC 98-3 as well as on the ISO/IEC Guide 98-3:2008/Suppl 1:2008 and ISO/IEC Guide 98-3:2008/Suppl 2:2011. As explained in detail in ISO 11929-2, the characteristic values are mathematically defined by means of moments and quantiles of probability distributions of the possible measurand values.

Since measurement uncertainty plays an important part in ISO 11929, the evaluation of measurements and the treatment of measurement uncertainties are carried out by means of the general procedures according to the ISO/IEC Guide 98-3 and to the ISO/IEC Guide 98-3:2008/Suppl 1:2008; see also References [13] to [17]. This enables the strict separation of the evaluation of the measurements, on the one hand, and the provision and calculation of the characteristic values, on the other hand. The ISO 11929 series makes use of a theory of uncertainty in measurement[18] to[20] based on Bayesian statistics (e.g. see References [21] to [26]) in order to take into account those uncertainties that cannot be derived from repeated or counting measurements. The latter uncertainties cannot be handled by frequentist statistics.

Because of developments in metrology concerning measurement uncertainty laid down in the ISO/IEC Guide 98-3, ISO 11929:2010 was drawn up on the basis of the ISO/IEC Guide 98-3, but using Bayesian statistics and the Bayesian theory of measurement uncertainty. This theory provides a Bayesian foundation for the ISO/IEC Guide 98-3. Moreover, ISO 11929:2010 was based on the definitions of the characteristic values,[13] the standard proposal,[14] and the explanatory article.[15] It unified and replaced all earlier parts of ISO 11929 and was applicable not only to a large variety of particular measurements of ionizing radiation but also, in analogy, to other measurement procedures.

In 2008 the ISO/IEC Guide 98-3:2008/Suppl 1:2008 has been published, dealing comprehensively with a more general treatment of measurement uncertainty using the Monte Carlo method in complex measurement evaluations. This provided an incentive for writing a corresponding Monte Carlo supplement[16] to ISO 11929:2010 and to revise ISO 11929:2010. The revised ISO 11929 is also essentially founded on Bayesian statistics and can serve as a bridge between ISO 11929:2010 and the ISO/IEC Guide 98-3:2008/Suppl 1:2008. Moreover, more general definitions of the characteristic values (ISO 11929-2) and the Monte