
**Očesna optika - Kontaktne leče in izdelki za vzdrževanje kontaktnih leč -
Mikrobiološke zahteve in preskusne metode za izdelke in postopki higienskega
ravnanja s kontaktnimi lečam (ISO/DIS 14729:2026)**

Ophthalmic optics - Contact lens care products - Microbiological requirements and test methods for products and regimens for hygienic management of contact lenses (ISO/DIS 14729:2026)

Augenoptik - Kontaktlinsenpflegemittel - Mikrobiologische Anforderungen und Prüfverfahren für Produkte und Systeme zum Hygienemanagement von Kontaktlinsen (ISO/DIS 14729:2026)

Optique ophtalmique - Produits d'entretien des lentilles de contact - Exigences microbiologiques et méthodes d'essai des produits et protocoles d'entretien des lentilles de contact (ISO/DIS 14729:2026)

Ta slovenski standard je istoveten z: prEN ISO 14729

ICS:

11.040.70 Oftalmološka oprema Ophthalmic equipment

oSIST prEN ISO 14729:2026

en,fr,de

Sample Document

get full document from standards.iteh.ai



DRAFT International Standard

ISO/DIS 14729

Ophthalmic optics — Contact lens care products — Microbiological requirements and test methods for products and regimens for hygienic management of contact lenses

*Optique ophtalmique — Produits d'entretien des lentilles de
contact — Exigences microbiologiques et méthodes d'essai des
produits et protocoles d'entretien des lentilles de contact*

ICS: 11.040.70

This document has not been edited by the ISO Central Secretariat.

ISO/CEN PARALLEL PROCESSING

Reference number
ISO/DIS 14729:2026(en)

ISO/TC 172/SC 7

Secretariat: **DIN**

Voting begins on:
2026-08-06

Voting terminates on:
2026-10-29

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENTS AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

© ISO 2026

Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2026

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO/DIS 14729:2026(en)

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
4.1 General	1
4.2 Stand-alone test (Inoculum challenge test)	2
4.3 Regimen test	3
5 Performance requirements	4
5.1 Stand-alone test: Primary criteria (see also Table 1)	4
5.1.1 Bacteria	4
5.1.2 Moulds and yeasts	4
5.2 Stand-alone test: Secondary criteria (see also Table 1)	4
5.3 Regimen test: Regimen criteria (see also Table 1)	4
6 Test methods	5
6.1 Materials and reagents	5
6.1.1 Test organisms	5
6.1.2 Culture media and reagents	5
6.1.3 Test equipment	6
6.1.4 Test samples	6
6.1.5 Culture maintenance	6
6.2 Preparation of microbial challenge (Inoculum)	6
6.3 Stand-alone procedure	7
6.3.1 Inoculum challenge test procedure	7
6.3.2 Controls	8
6.3.3 Test report	9
6.4 Regimen procedure	9
6.4.1 Lens inoculation	9
6.4.2 Lens treatment	10
6.4.3 Recovery of surviving challenge organisms (see Annex B for an example of a membrane filtration procedure)	10
6.4.4 Controls	11
6.4.5 Test report	11
Annex A (informative) Test organisms from other culture collections	13
Annex B (informative) Example of a membrane filtration procedure	14
Annex C (informative) Technical report: Virus testing	16
Annex D (informative) Technical report: <i>Acanthamoeba</i> testing	17
Annex E (informative) Technical report: Artificial tears (organic soil) in laboratory testing	18
Bibliography	19

ISO/DIS 14729:2026(en)**Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 14729 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

[Annex A](#) to [Annex E](#) of this International Standard are for information only.

Sample Document

get full document from standards.iteh.ai

ISO/DIS 14729:2026(en)**Introduction**

Products for contact lens disinfection by chemical means are intended to reduce microbial contamination introduced during lens wear and removal, cleaning and storage and are required to contain antimicrobial agents capable of achieving this.

It is essential that all liquid contact lens care products are sterile until opened. Dry products (tablets, granules, etc.) should be subject to control of microbial contamination and should be dissolved in a suitable diluent immediately prior to use. Multidose contact lens care products must be adequately preserved or be packaged in a container designed and labelled to minimize the risk of injury resulting from in-use contamination.

Contact lenses are normally subject to a regimen of cleaning and contact lens disinfection between periods of wear. Aqueous solutions containing cleaning and/or disinfecting agents are commonly used for this purpose. These products may be marketed as solutions or as tablets for dissolution immediately prior to use in a suitable diluent such as saline.

The past 20 years of experience in the use and regulation of contact lens disinfecting products has shown distinct disinfecting antimicrobial criteria for this class of medical devices. Ocular toxicology concerns, process convenience and product comfort on the eye, have meant an evolution of products which maintain a low incidence of contact lens associated ocular infection when used as instructed by the manufacturer. This International Standard gives these distinct contact lens disinfecting antimicrobial criteria along with annexes to explain why viruses ([Annex C](#)) and Acanthamoeba ([Annex D](#)) are not included as challenges. Organic soil is not required for evaluation of contact lens care disinfecting products but may be used; an informative annex ([Annex E](#)) is included to discuss organic soil in the context of contact lenses and contact lens care products.

Sample Document

get full document from standards.iteh.ai