
Oftalmični instrumenti - Topografi roženice (ISO/DIS 19980:2026)

Ophthalmic instruments - Corneal topographers (ISO/DIS 19980:2026)

Ophthalmische Instrumente - Hornhauttopographen (ISO/DIS 19980:2026)

Instruments ophtalmiques - Topographes cornéens (ISO/DIS 19980:2026)

Ta slovenski standard je istoveten z: prEN ISO 19980

ICS:

11.040.70 Oftalmološka oprema Ophthalmic equipment

oSIST prEN ISO 19980:2026

en,fr,de

Sample Document

get full document from standards.iteh.ai



DRAFT International Standard

ISO/DIS 19980

Ophthalmic instruments — Corneal topographers

Instruments ophtalmiques — Topographes de la cornée

ICS: 11.040.70

ISO/TC 172/SC 7

Secretariat: **DIN**

Voting begins on:
2026-08-10

Voting terminates on:
2026-11-02

Sample Document

get full document from standards.iteh.ai

This document has not been edited by the ISO Central Secretariat.

ISO/CEN PARALLEL PROCESSING

Reference number
ISO/DIS 19980:2026(en)

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENTS AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

© ISO 2026

Sample Document

get full document from standards.iteh.ai



COPYRIGHT PROTECTED DOCUMENT

© ISO 2026

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2026 – All rights reserved

ISO/DIS 19980:2026(en)

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	8
4.1 Area measured	8
4.2 Measurement sample density	9
4.3 Measurement and report of performance	9
4.4 Colour presentation of results	9
5 Test methods and test devices	9
5.1 Tests	9
5.1.1 General	9
5.1.2 Accuracy test	9
5.1.3 Repeatability test	9
5.2 Accuracy test specifications	9
5.2.1 Test surface geometry specification	9
5.2.2 Test surface requirements according to topographer technology	10
5.2.3 Verification and calibration of test surfaces	10
5.2.4 Accuracy test tolerances	11
5.3 Data collection — Test surfaces	11
5.4 Analysis of the data	12
5.4.1 General	12
5.4.2 Structure of the accuracy data set	12
5.4.3 Analysis of the paired data sets	12
5.4.4 Report of accuracy performance	13
6 Accompanying documents	13
7 Marking	13
Annex A (informative) Test surfaces for corneal topographers (CTs)	14
Annex B (normative) Standardized displays for corneal topographers (CTs)	16
Annex C (normative) Calculation of area-weighting values	19
Annex D (normative) Test methods for measuring human corneas	21
Bibliography	22

ISO/DIS 19980:2026(en)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fourth edition cancels and replaces the third edition (ISO 19980:2021), which has been technically revised. The main changes compared to the previous edition are as follows:

- Adapted [3.11](#) and [3.12](#).
- Added the requirement to test with all test surfaces for type A corneal topographers in 5.2.6.
- Replaced standard deviation with RMS as measure of accuracy in [5.2](#) and [5.4](#).
- Removed surface profile tolerance requirement for test surfaces in 5.2.6.
- Removed references to IEC 60601-1.
- Corrected [Annex B](#) status from informative to normative (editorial).
- Removed the HBS columns in [Table B.2](#).
- Editorial changes.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Ophthalmic instruments — Corneal topographers

1 Scope

This document specifies minimum requirements for instruments and systems that fall into the class of corneal topographers (CTs). It also specifies tests and procedures to verify that a system or instrument complies with this document and thus qualifies as a CT according to this document. It also specifies tests and procedures that allow the verification of capabilities of systems that are beyond the minimum requirements for CTs.

This document defines terms that are specific to the characterization of the corneal shape so that they may be standardized throughout the field of vision care.

This document is applicable to instruments, systems and methods that are intended to measure the surface shape of the cornea of the human eye.

NOTE The measurements can be of the curvature of the surface in local areas, three-dimensional topographical measurements of the surface or other more global parameters used to characterize the surface.

This document is not applicable to ophthalmic instruments classified as ophthalmometers.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

corneal apex

location on the corneal surface where the mean of the local principal curvature is greatest

Note 1 to entry: See [Figure 1](#).

3.2

corneal eccentricity

e_c

eccentricity, e , of the conic section that best fits the *corneal meridian* ([3.3](#)) of interest

Note 1 to entry: If the meridian is not specified, the corneal eccentricity is that of the flattest corneal meridian (see [Table 1](#) and [Annex A](#)).