

# SLOVENSKI STANDARD

## SIST EN ISO 10993-1:2026

01-februar-2026

Nadomešča:  
**SIST EN ISO 10993-1:2021**

---

**Biološko ovrednotenje medicinskih pripomočkov - 1. del: Zahteve in splošna načela za oceno biološke varnosti znotraj procesa obvladovanja tveganja (ISO 10993-1:2025)**

Biological evaluation of medical devices - Part 1: Requirements and general principles for the evaluation of biological safety within a risk management process (ISO 10993-1:2025)

iTeh Standards

Biologische Beurteilung von Medizinprodukten - Teil 1: Anforderungen und allgemeine Grundsätze für die Beurteilung der biologischen Sicherheit im Rahmen eines Risikomanagementsystems (ISO 10993-1:2025)

Évaluation biologique des dispositifs médicaux - Partie 1: Exigences et principes généraux pour l'évaluation de la sécurité biologique au sein d'un processus de gestion des risques (ISO 10993-1:2025)

Ta slovenski standard je istoveten z: **EN ISO 10993-1:2025**

---

**ICS:**

11.100.20

Biološko ovrednotenje  
medicinskih pripomočkov

Biological evaluation of  
medical devices

**SIST EN ISO 10993-1:2026**

**en,fr,de**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

EN ISO 10993-1

December 2025

ICS 11.100; 11.100.20

Supersedes EN ISO 10993-1:2020

English Version

Biological evaluation of medical devices - Part 1:  
Requirements and general principles for the evaluation of  
biological safety within a risk management process (ISO  
10993-1:2025)

Évaluation biologique des dispositifs médicaux - Partie  
1: Exigences et principes généraux pour l'évaluation de  
la sécurité biologique au sein d'un processus de gestion  
des risques (ISO 10993-1:2025)

Biologische Beurteilung von Medizinprodukten - Teil 1:  
Anforderungen und allgemeine Grundsätze für die  
Beurteilung der biologischen Sicherheit im Rahmen  
eines Risikomanagementsystems (ISO 10993-1:2025)

This European Standard was approved by CEN on 18 August 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
<b>European foreword.....</b>	<b>3</b>
<b>Annex ZA (informative) Relationship between this European Standard the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered .....</b>	<b>4</b>

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[SIST EN ISO 10993-1:2026](#)

<https://standards.iteh.ai/catalog/standards/sist/44effe77-e359-4409-8de4-fc818d8b342f/sist-en-iso-10993-1-2026>

## European foreword

This document (EN ISO 10993-1:2025) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2026, and conflicting national standards shall be withdrawn at the latest by June 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10993-1:2020.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

## iTeh Standards

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## Endorsement notice

The text of ISO 10993-1:2025 has been approved by CEN as EN ISO 10993-1:2025 without any modification.

## Annex ZA

(informative)

### Relationship between this European Standard the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under M/575 to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up.

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in Table ZA.1 and application of the edition of the normatively referenced standards as given in Table ZA.2 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA Regulations.

Where a definition in this harmonised standard differs from a definition of the same term set out in Regulation (EU) 2017/745, the differences shall be indicated in the Annex Z. For the purpose of using this standard in support of the requirements set out in Regulation (EU) 2017/745, the definitions set out in this Regulation prevail.

Where the European standard is an adoption of an International Standard, the scope of this document can differ from the scope of the European Regulation that it supports. As the scope of the applicable regulatory requirements differ from nation to nation and region to region the standard can only support European regulatory requirements to the extent of the scope of the European Regulation for medical devices ((EU) 2017/745).

<https://standards.iteh.ai> [SIST EN ISO 10993-1:2026](https://standards.iteh.ai)  
 NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 When a General Safety and Performance Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

NOTE 4 General Safety and Performance Requirements listed in Table ZA.1 are specific to biological safety for patients only, unless the medical device is used for personal protection of the user

**Table ZA.1 — Correspondence between this European Standard and Annex I of Regulation (EU) 2017/745 [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up**

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s)/subclause(s) of this EN	Remarks/Notes
10.1. a)	4.1, 4.3, 6.2, 6.3, 6.5, 7, 8, 10	This is partly covered by this document since the standard does not provide requirements for performance. Flammability is not covered.
10.1. b), d) and g)	4.1, 4.3, 5, 6.1, 6.2, 6.3, 6.4, 6.5, 7, 10	These are partly covered by this document since the standard does not provide requirements for performance.
10.2	4.1, 4.3, 5, 6.1, 6.2, 6.3, 6.4, 6.5, 7, 8, 10	Persons involved in the transport, storage and use of medical devices are not covered unless the user is the patient, or the medical device is used for personal protection.
10.4.1 (first paragraph)	4.1, 4.3, 5, 6.1, 6.2, 6.3, 6.4, 6.5, 7, 8, 10	
10.6	4.1, 4.3, 6.2, 6.3, 6.5, 7, 8, 10	

**Table ZA.2 — Normative references from clause 2 of this document and their corresponding European publications**

Column 1 Reference in Clause 2	Column 2 International Standard Edition	Column 3 Title	Column 4 Corresponding European Standard Edition
ISO 10993-2	ISO 10993-2:2022	Biological evaluation of medical devices — Part 2: Animal welfare requirements	EN ISO 10993-2:2022
ISO 10993-3	ISO 10993-3:2014 ISO 10993-3:2014/AMD 1:2025	Biological evaluation of medical devices — Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity	EN ISO 10993-3:2014 EN ISO 10993-4:2017/A1:2025
ISO 10993-4	ISO 10993-4:2017	Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood	EN ISO 10993-4:2017
ISO 10993-5	ISO 10993-5:2009	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity	EN ISO 10993-5:2009 EN ISO 10993-5:2009/A11:2025