

01-maj-2026**Nadomešča:****SIST EN ISO 15883-6:2015**

Čistilno-dezinfekcijske naprave - 6. del: Zahteve in preskusi čistilno-dezinfekcijskih naprav s toplotno dezinfekcijo za nenujne medicinske pripomočke in zdravstveno opremo (ISO 15883-6:2026)

Washer-disinfectors - Part 6: Requirements and tests for washer-disinfectors employing thermal disinfection for non-critical medical devices and health care equipment (ISO 15883-6:2026)

Reinigungs-Desinfektionsgeräte - Teil 6: Anforderungen und Prüfverfahren für Reinigungs-Desinfektionsgeräte mit thermischer Desinfektion für nicht kritische Medizinprodukte und Zubehör im Gesundheitswesen (ISO 15883-6:2026)

Laveurs désinfecteurs - Partie 6: Exigences et essais pour les laveurs désinfecteurs utilisant une désinfection thermique pour les dispositifs médicaux non invasifs, non critiques et pour l'équipement de soins de santé (ISO 15883-6:2026)

Ta slovenski standard je istoveten z: EN ISO 15883-6:2026**ICS:**

11.080.10 Sterilizacijska oprema Sterilizing equipment

SIST EN ISO 15883-6:2026**en,fr,de**

Sample Document

get full document from standards.iteh.ai

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 15883-6

March 2026

ICS 11.080.10

Supersedes EN ISO 15883-6:2015

English Version

**Washer-disinfectors - Part 6: Requirements and tests for
washer-disinfectors employing thermal disinfection for
non-critical medical devices and health care equipment
(ISO 15883-6:2026)**

Laveurs-désinfecteurs - Partie 6: Exigences et essais
pour les laveurs-désinfecteurs utilisant la désinfection
thermique pour les dispositifs médicaux non critiques
et les équipements de soins de santé (ISO 15883-
6:2026)

Reinigungs-Desinfektionsgeräte - Teil 6:
Anforderungen und Prüfverfahren für Reinigungs-
Desinfektionsgeräte mit thermischer Desinfektion für
nicht kritische Medizinprodukte und Zubehör im
Gesundheitswesen (ISO 15883-6:2026)

This European Standard was approved by CEN on 22 January 2026.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2026 CEN All rights of exploitation in any form and by any means reserved
worldwide for CEN national Members.

Ref. No. EN ISO 15883-6:2026 E

Contents	Page
European foreword.....	3
Annex ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered.....	4

Sample Document

get full document from standards.iteh.ai

European foreword

This document (EN ISO 15883-6:2026) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 102 "Sterilizers and associated equipment for processing of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2026, and conflicting national standards shall be withdrawn at the latest by September 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15883-6:2015.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 15883-6:2026 has been approved by CEN as EN ISO 15883-6:2026 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered

This European standard has been prepared under M/575 to provide one voluntary means of conforming to the General Safety and Performance Requirements of Regulation (EU) 2017/745 of 5 April 2017 concerning medical devices [O] L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up.

Once this standard is cited in the Official Journal of the European Union under that Regulation, compliance with the normative clauses of this standard given in [Table ZA.1](#) and application of the edition of the normatively referenced standards as given in [Table ZA.2](#) confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding General Safety and Performance Requirements of that Regulation, and associated EFTA regulations.

Where a definition in this standard differs from a definition of the same term set out in Regulation (EU) 2017/745, the differences shall be indicated in this Annex Z. For the purpose of using this standard in support of the requirements set out in Regulation (EU) 2017/745, the definitions set out in this Regulation prevail.

Where the European standard is an adoption of an International Standard, the scope of this standard can differ from the scope of the European Regulation that it supports. As the scope of the applicable regulatory requirements differ from nation to nation and region to region, the standard can only support European regulatory requirements to the extent of the scope of the European regulation for medical devices (EU) 2017/745.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Regulation (EU) 2017/745. This means that risks have to be 'reduced as far as possible', 'reduced to the lowest possible level', 'reduced as far as possible and appropriate', 'removed or reduced as far as possible', 'eliminated or reduced as far as possible', 'removed or minimized as far as possible', or 'minimized', according to the wording of the corresponding General Safety and Performance Requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with General Safety and Performance Requirements 1, 2, 3, 4, 5, 8, 9, 10, 11, 14, 16, 17, 18, 19, 20, 21 and 22 of the Regulation.

NOTE 3 When a General Safety and Performance Requirement does not appear in [Table ZA.1](#), it means that it is not addressed by this European Standard.

Table ZA.1 — Correspondence between this European Standard and Annex I of Regulation (EU) 2017/745 [OJ L 117] and to system or process requirements including those relating to quality management systems, risk management, post-market surveillance systems, clinical investigations, clinical evaluation or post-market clinical follow-up

General Safety and Performance Requirements of Regulation (EU) 2017/745	Clause(s) / sub-clause(s) of this EN	Remarks / Notes
5 (a)	4.1.3, 4.1.4	<p>Clauses 4.1.3, 4.1.4, only partly covered in respect of reducing the risks related to use error by reducing the risks related to the ergonomic features of the washer-disinfectors (WDs).</p> <p>Aspects related to the environment in which the WD is intended to be used are not covered.</p> <p>Aspects related to manufacturing are not covered.</p>
14.2 (a)	4.1.5, 5.1, 5.2	<p>All selected clauses only partly cover the requirement.</p> <p>Covered in respect of reducing the risks of injury, in connection with WD physical and ergonomic features.</p> <p>Aspects related to the WD manufacturing processes are not covered.</p>
23.4 (k)	Error! Reference source not found.a), Error! Reference source not found. a) – Error! Reference source not found. b)	Covered in respect of documentation provided for installation and safe operation of the WD.