International Standard

ISO 4823

Dentistry — Elastomeric impression and bite registration materials

Médecine bucco-dentaire — Produits pour empreintes et matériaux pour enregistrement des rapports intermaxillaires à base d'élastomères

Sixth edition
2025-06

Corrected version
2025-08

SIST Preview
(https://ecommerce.sist.si)

ISO 4823:2025
https://ecommerce.sist.si/catalog/standards/iso/265f4086-4190-436f-ad7d-40ce872d239c/iso-4823-2025

Reference number
ISO 4823:2025(en)

© ISO 2025

SIST Preview
(https://ecommerce.sist.si)

ISO 4823:2025
https://ecommerce.sist.si/catalog/standards/iso/265f4086-4190-436f-ad7d-40ce872d239c/iso-4823-2025



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword ..... **v**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **1**

4 Classification **2**

5 Requirements **2**
5.1 Component colours (not applicable for transparent materials) 2
5.2 Mixing time (hand-spatulated or hand-kneaded mixes) 3
5.3 Consistency 3
5.4 Working time 3
5.5 Detail reproduction 3
5.6 Linear dimensional change 3
5.7 Compatibility with gypsum 3
5.8 Elastic recovery 3
5.9 Strain in compression 3
5.10 Minimum time in the oral cavity for bite registration materials 3
5.11 Compression set of bite registration materials 3
5.12 Hardness of bite registration materials 3

6 Pre-test planning approaches **4**
6.1 General 4
6.2 Sampling 4
6.3 Pre-test product examinations 4
6.3.1 General 4
6.3.2 Examinations for conformity with labelling requirements 4
6.3.3 Examinations for effectiveness of the packaging 5
6.3.4 Examinations for compliance with requirements for instructions for use 5
6.4 Essential pre-test preparatory practices 5
6.4.1 Laboratory conditions 5
6.4.2 Apparatus function verification steps 5
6.4.3 Volume of materials to be mixed for each specimen 5
6.4.4 Proportioning, mixing and handling of hand-mixed materials to be tested 6
6.4.5 Timing for the specimen preparation and test procedures 6
6.4.6 Simulated oral time/temperature treatment of specimens formed in completely closed mould assemblies 6
6.5 Pass/fail determinations 6
6.6 Expression of test results 6

7 Test methods **6**
7.1 Mixing time 6
7.1.1 Apparatus 6
7.1.2 Specimen preparation and test procedure (five specimens) 6
7.1.3 Pass/fail determination and expression of results 6
7.2 Consistency 7
7.2.1 Apparatus and materials 7
7.2.2 Advance preparation steps 7
7.2.3 Specimen preparation and test procedure (five specimens) 7
7.2.4 Pass/fail determination and expression of results 8
7.3 Working time 8
7.3.1 Apparatus 8
7.3.2 Working time test 9
7.4 Detail reproduction 10
7.4.1 Apparatus and materials 10
7.4.2 Specimen preparation 10
7.4.3 Test procedure 11

SIST Preview
(https://standards.iteh.ai)
ISO 4823:2025
https://ecommerce.sist.si/catalog/standards/sist/3e3f0e8e-4190-436f-ad7d-40ce872d239c/iso-4823-2025

© ISO 2025 – All rights reserved

7.4.4 Pass/fail determination and expression of results 11
7.5 Linear dimensional change 11
7.5.1 Apparatus and materials 11
7.5.2 Test block line-length measurement procedure 11
7.5.3 Specimen preparation 12
7.5.4 Test specimen measurement 12
7.6 Compatibility with gypsum 13
7.6.1 Apparatus and materials 13
7.6.2 Specimen preparation 14
7.6.3 Test procedure 14
7.6.4 Pass/fail determination and expression of results 14
7.7 Elastic recovery 14
7.7.1 Apparatus and materials 14
7.7.2 Specimen preparation 15
7.7.3 Test procedure 15
7.7.4 Calculation of results 16
7.7.5 Pass/fail determination and expression of results 16
7.8 Strain in compression 16
7.8.1 Apparatus 16
7.8.2 Specimen preparation 16
7.8.3 Test procedure 16
7.8.4 Calculation of results 17
7.8.5 Pass/fail determination and expression of results 17
7.9 Minimum time in the oral cavity and compression set for bite registration materials 17
7.9.1 Apparatus 17
7.9.2 Specimen preparation 17
7.9.3 Test procedure 17
7.9.4 Evaluation 17
7.9.5 Pass/fail determination and expression of results 18
7.10 Hardness of bite registration materials 18
7.10.1 Apparatus 18
7.10.2 Specimen preparation 18
7.10.3 Test procedure 18
7.10.4 Evaluation 18
7.10.5 Pass/fail determination and expression of results 18

8 Requirements for labelling and instructions supplied by the manufacturer 19
8.1 Labelling requirements 19
8.2 Instructions for use 19
8.2.1 General 19
8.2.2 For all materials covered by this document 20
8.2.3 For all impression materials 20
8.2.4 Additional instructions 20

Annex A (normative) **Figures 22**

Annex B (normative) **Standardized hand mixing methods 30**

Bibliography 33

SIST Preview
(https://ecommerce.sist.si)
ISO 4823:2025
https://ecommerce.sist.si/catalog/standards/sist/265f4086-4190-436f-a474-40ac873d4230/iso-4823-2025

© ISO 2025 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthodontic materials,* in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry,* in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This sixth edition cancels and replaces the fifth edition (ISO 4823:2021), which has been technically revised. The main changes are as follows:

— packaging and instructions for use requirements have been updated;

— editorial corrections have been made.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 4823:2025 incorporates the following corrections:

— In 7.2.1.2, in the key and in NOTE 1 to Figure A.1, the value "(0,5 ± 0,2) ml" was corrected to "(0,5 ± 0,02) ml".

SIST Preview
https://ecommerce.sist.si/catalog/standards/iso/265f4086-4190-436f-ad7d-40ce872d239c/iso-4823-2025

© ISO 2025 – All rights reserved

SIST Preview
(https://ecommerce.sist.si)

ISO 4823:2025
https://ecommerce.sist.si/catalog/standards/iso/265f4086-4190-436f-ad7d-40ce872d239c/iso-4823-2025